PVP

Release 0.2.0

jonny saunders et al

Jan 27, 2022

OVERVIEW

1 Software **3**

1.1 PVP Modules . . . 4
1.1.1 System Overview . . . 4
1.1.1.1 Hardware . . . 4
1.1.1.2 Software . . . 4
1.1.2 Performance . . . 5
1.1.2.1 ISO Standards Testing . . . 6
1.1.2.2 Breath Detection . . . 9
1.1.2.3 High Pressure Detection . . . 10
1.1.3 Medical Disclaimer . . . 10
1.1.4 Funding and Support . . . 11
1.1.5 Hardware Overview . . . 11
1.1.6 Components . . . 11
1.1.6.1 Hardware Design . . . 11
1.1.6.2 Actuator Selection . . . 13
1.1.6.3 Sensor Selection . . . 13
1.1.7 Assembly . . . 14
1.1.7.1 Part 1. 3D Printed Components and Enclosure . . . 15
1.1.7.2 Part 2. Basic Hardware Assembly . . . 24
1.1.7.3 Part 3. Electronics Assembly . . . 59
1.1.7.4 Part 4. Putting it all together . . . 86
1.1.8 Electronics . . . 96
1.1.8.1 Power and I/O . . . 98
1.1.8.2 Sensor PCB . . . 98
1.1.8.3 Actuator PCB . . . 100
1.1.9 Bill of Materials . . . 101
1.1.9.1 I/O . . . 102
1.1.9.2 Airway . . . 103
1.1.9.3 Framing & Structural . . . 104
1.1.9.4 Electronics . . . 106
1.1.9.5 Specialized Tools . . . 107
1.1.9.6 Etc. . . . 107
1.1.10 CAD . . . 107
1.1.10.1 3D Printed Parts . . . 107
1.1.10.2 Enclosure . . . 108
1.1.11 Software Overview . . . 109
1.1.12 Folder Structure . . . 109
1.1.12.1 PVP Modules . . . 110
1.1.13 GUI . . . 110
1.1.13.1 Main GUI Module . . . 110

1.1.13.2 GUI Widgets . . . 118
1.1.13.3 GUI Stylesheets . . . 141
1.1.13.4 Module Overview . . . 142
1.1.13.5 Screenshot . . . 142
1.1.14 Controller . . . 143
1.1.14.1 Purpose of the Controller . . . 143
1.1.14.2 Architecture of the Controller . . . 144
1.1.15 common module . . . 152
1.1.15.1 Values . . . 152
1.1.15.2 Message . . . 158
1.1.15.3 Loggers . . . 161
1.1.15.4 Prefs . . . 164
1.1.15.5 Unit Conversion . . . 167
1.1.15.6 utils . . . 168
1.1.15.7 fashion . . . 169
1.1.16 pvp.io package . . . 169
1.1.16.1 pvp.io.hal module . . . 169
1.1.16.2 devices . . . 171
1.1.17 Alarm . . . 171
1.1.17.1 Alarm System Overview . . . 171
1.1.17.2 Alarm Modules . . . 172
1.1.17.3 Main Alarm Module . . . 188
1.1.18 coordinator module . . . 190
1.1.18.1 Submodules . . . 190
1.1.18.2 coordinator . . . 190
1.1.18.3 ipc . . . 193
1.1.18.4 process_manager . . . 194
1.1.19 Index . . . 194

2 Medical Disclaimer **195**

Python Module Index **197**

Index **199**

The global COVID-19 pandemic has highlighted the need for a low-cost, rapidly-deployable ventilator, for the current as well as future respiratory virus outbreaks. While safe and robust ventilation technology exists in the commercial sector, the small number of capable suppliers cannot meet the severe demands for ventilators during a pandemic. Moreover, the specialized, proprietary equipment developed by medical device manufacturers is expensive and inaccessible in low-resource areas.

The **People's Ventilator Project (PVP)** is an open-source, low-cost pressure-control ventilator designed for minimal reliance on specialized medical parts to better adapt to supply chain shortages. The **PVP** largely follows established design conventions, most importantly active and computer-controlled inhalation, together with passive exhalation. It supports pressure-controlled ventilation, combined with standard-features like autonomous breath detection, and the suite of FDA required alarms.

See our medRxiv preprint here!

PVP is a pressure-controlled ventilator that uses a minimal set of inexpensive, off-the-self hardware components. An inexpensive proportional valve controls inspiratory flow, and a relay valve controls expiratory flow. A gauge pressure sensor monitors airway pressure, and an inexpensive D-lite spirometer used in conjunction with a differential pressure sensor monitors expiratory flow.

PVP's components are coordinated by a Raspberry Pi 4 board, which runs the graphical user interface, administers the alarm system, monitors sensor values, and sends actuation commands to the valves. The core electrical system consists of two modular board 'hats', a sensor board and an actuator board, that stack onto the Raspberry Pi via 40-pin stackable headers. The modularity of this system enables individual boards to be revised or modified to substitute components in the case of part scarcity.

CHAPTER

ONE

SOFTWARE



PVP's software was developed to bring the philosophy of free and open-source software to medical devices. PVP is not only open from top to bottom, but we have developed it as a framework for **an adaptable, general-purpose, communally-developed ventilator.**

PVP's ventilation control system is fast, robust, and **written entirely in high-level Python** (3.7) – without the development and inspection bottlenecks of split computer/microprocessor systems that require users to read and write low-level hardware firmware.

All of PVP's components are **modularly designed**, allowing them to be reconfigured and expanded for new ventilation modes and hardware configurations.

We provide complete **API-level documentation** and an **automated testing suite** to give everyone the freedom to inspect, understand, and expand PVP's software framework.

1.1 PVP Modules

1.1.1 System Overview

The **People's Ventilator Project (PVP)** is an open-source, low-cost pressure-control ventilator designed for minimal reliance on specialized medical parts to better adapt to supply chain shortages.

1.1.1.1 Hardware

The device components were selected to enable a **minimalistic and relatively low-cost ventilator design, to avoid supply chain limitations, and to facilitate rapid and easy assembly**. Most parts in the PVP are not medical-specific devices, and those that are specialized components are readily available and standardized across ventilator platforms, such as standard respiratory circuits and HEPA filters. We provide complete assembly of the PVP, including 3D-printable components, as well as justifications for selecting all actuators and sensors, as guidance to those who cannot source an exact match to components used in the Bill of Materials.

PVP Hardware

1.1.1.2 Software



PVP's software was developed to bring the philosophy of free and open-source software to medical devices. PVP is not only open from top to bottom, but we have developed it as a framework for **an adaptable, general-purpose, communally-developed ventilator.**

PVP's ventilation control system is fast, robust, and **written entirely in high-level Python** (3.7) – without the development and inspection bottlenecks of split computer/microprocessor systems that require users to read and write low-level hardware firmware.

All of PVP's components are **modularly designed**, allowing them to be reconfigured and expanded for new ventilation modes and hardware configurations.

We provide complete **API-level documentation** and an **automated testing suite** to give everyone the freedom to inspect, understand, and expand PVP's software framework.

PVP Modules

1.1.2 Performance



Fig. 1: Representative pressure control breath cycle waveforms for airway pressure and flow out. Test settings: compliance C=20 mL/cm H2O, airway resistance R=20 cm H2O/L/s, PIP=30 cm H2O, PEEP=5 cm H2O.

The completed system was tested with a standard test lung (QuickLung, IngMar Medical, Pittsburgh, PA) that allowed testing combinations of three lung compliance settings (C=5, 20, and 50 mL cm H2O) and three airway resistance settings (R=5, 20, and 50 cm H2O/L/s). The figure above shows pressure control performance for midpoint settings: C=20 mL/cm H2O, R=20 cm H2O/L/s, PIP=30 cm H2O, PEEP=5 cm H2O. PIP is reached within a 300 ms ramp period, then holds for the PIP plateau with minimal fluctuation of airway pressure for the remainder of the inspiratory cycle (blue). One the expiratory valve opens, exhalation begins and expiratory flow is measured (orange) as the airway pressure drops to PEEP and remains there for the rest of the PEEP period.

Some manual adjustment of the pressure waveforms may be warranted depending on the patient, and such adjustment is permitted through a user flow adjustment setting. This flow adjustment setting allows the user to increase the maximum flow rate during the ramp cycle to inflate lungs with higher compliance. The flow setting can be readily changed from the GUI and the control system immediately adapts to the user's input. An example of this flow adjustment is shown in the figure above for four breath cycles. While all cycles reach PIP, the latter two have a higher mean airway pressure, which may be more desirable under certain conditions than the lower mean airway pressure of the former two.



Fig. 2: Demonstration of waveform tuning via flow adjustment. If desired, the operator can increase the flow setting through the system GUI to decrease the pressure ramp time. Test settings: compliance C=20 mL/cm H2O, airway resistance R=20 cm H2O/L/s, PIP=30 cm H2O, PEEP=5 cm H2O.

1.1.2.1 ISO Standards Testing

In order to characterize the PVP's control over a wide range of conditions, we followed FDA Emergency Use Authorization guidelines, which specify ISO 80601-2-80-2018 for a battery of pressure controlled ventilator standard tests. We tested the conditions that do not stipulate a leak, and present the results here. For each configuration the following parameters are listed: the test number (from the table below), the compliance (C, mL/cm H2O), linear resistance (R, cm H2O/L/s), respiratory frequency (f, breaths/min), peak inspiratory pressure (PIP, cm H2O), positive end-expiratory pressure (PEEP, cm H2O), and flow adjustment setting.

Table 1: Standard test battery from Table 201.105 in ISO 80601-2-80-2018 for pressure controlled ventilators

Test number	Intended delivered volume (mL)	Compliance (mL (hPa)^-1)	Linear resistance (hPa(L/s)^-1) +/- 10%	Leakage (mL/min) +/- 10%	Ventilatory frequency (breaths/min)	Inspiratory time (s)	Pressure (hPa)	PEEP (hPa)
1	500	50	5	0	20	1	10	5
2	500	50	20	0	12	1	15	10
3	500	20	5	0	20	1	25	5
4	500	20	20	0	20	1	25	10
5	500	50	5	5000	20	1	25	5
6	500	50	20	10000	12	1	25	10
7	300	20	20	0	20	1	15	5
8	300	20	50	0	12	1	25	10
9	300	10	50	0	20	1	30	5
10	300	20	20	3000	20	1	25	5
11	300	20	50	6000	12	1	25	10
12	200	10	20	0	20	1	25	10

These tests cover an array of conditions, and more difficult test cases involve a high airway pressure coupled with a



Fig. 3: Performance results of the ISO 80601-2-80-2018 pressure controlled ventilator standard tests with an intended delivered tidal volume of 500 mL. For each configuration the following parameters are listed: the test number (from table 201.105 in the ISO standard), the compliance (C, mL/cm H2O), linear resistance (R, cm H2O/L/s), respiratory frequency (f, breaths/min), peak inspiratory pressure (PIP, cm H2O), positive end-expiratory pressure (PEEP, cm H2O), and flow adjustment setting. PIP is reached in every test condition except for case 2, which is approximately 2.4 cm H2O below the set point.



Fig. 4: Performance results of the ISO 80601-2-80-2018 pressure controlled ventilator standard tests with an intended delivered tidal volume of 300 mL. For each configuration the following parameters are listed: the test number (from table 201.105 in the ISO standard), the compliance (C, mL/cm H2O), linear resistance (R, cm H2O/L/s), respiratory frequency (breaths/min), peak inspiratory pressure (PIP, cm H2O), positive end-expiratory pressure (PEEP, cm H2O), and flow adjustment setting. PIP is reached in every test condition.

low lung compliance (case nos. 8 and 9). Under these conditions, if the inspiratory flow rate during the ramp phase is too high, the high airway resistance will produce a transient spike in airway pressure which can greatly overshoot the PIP value. For this reason, the system uses a low initial flow setting and allows the clinican to increase the flow rate if necessary.



Fig. 5: Tidal volume performance for the ISO 80601-2-80-2018 pressure controlled ventilator standard tests, averaged across 30 breath cycles for each condition.

The PVP integrates expiratory flow to monitor the tidal volume, which is not directly set in pressure controlled ventilation, but is an important parameter. Of the test conditions in the ISO standard, four that we tested intended a nominal delivered tidal volume of 500 mL, three intended 300 mL, and one intended 200 mL. For most cases, the estimated tidal volume has a tight spread clustered within 20% of the intended value.

1.1.2.2 Breath Detection



Fig. 6: Spontaneous breath detection.

A patient-initiated breath after exhalation will result in a momentary drop in PEEP. PVP may optionally detect these transient decreases to trigger a new pressure-controlled breath cycle. We tested this functionality by triggering numerous breaths out of phase with the intended inspiratory cycle, using a QuickTrigger (IngMar Medical, Pittsburgh, PA) to momentarily open the test lung during PEEP and simulate this transient drop of pressure.

1.1.2.3 High Pressure Detection



Fig. 7: High pressure alarm demonstration.

Above is a demonstration of the PVP's high airway pressure alarm (HAPA). An airway blockage results in a high airway pressure (above 60 cm H2O) that the system corrects within ~500 ms. Test settings: compliance C=20 mL/cm H2O, airway resistance R=20 cm H2O/L/s, PIP=30 cm H2O, PEEP=5 cm H2O.

1.1.3 Medical Disclaimer

PVP1 is not a regulated or clinically validated medical device. We have not yet performed testing for safety or efficacy on living organisms. All material described herein should be used at your own risk and do not represent a medical recommendation. PVP1 is currently recommended only for research purposes.

This website is not connected to, endorsed by, or representative of the view of Princeton University. Neither the authors nor Princeton University assume any liability or responsibility for any consequences, damages, or loss caused or alleged to be caused directly or indirectly for any action or inaction taken based on or made in reliance on the information or material discussed herein or linked to from this website.

PVP1 is under continuous development and the information here may not be up to date, nor is any guarantee made as such. Neither the authors nor Princeton University are liable for any damage or loss related to the accuracy, completeness or timeliness of any information described or linked to from this website.

By continuing to watch or read this, you are acknowledging and accepting this disclaimer.

1.1.4 Funding and Support

Funding and lab space for this project was provided by Princeton University in direct response to the COVID-19 pandemic. Note: See Disclaimer for further details.

We wish to thank the Trevor Day School of NYC (https://www.trevor.org) for their kind loan of multiple 3D printers, which greatly contributed to the rapid prototyping efforts of the team. Additional thanks are in order to recognize the indeterminate length of the loan (...no really, thanks for not asking for them back yet).

We also wish to thank Nick and Alex Winnard of the Berkshire Laser Co. in Pittsfield, MA (https://berkshirelaserco.com/) for generously donating their laser cutting and engraving services. We are grateful for their skills and expertise in creating the gorgeous acrylic enclosure, as displayed on our prototype device, and guidance improving the enclosure documentation.

1.1.5 Hardware Overview

Schematic diagram of main mechanical components

The PVP components were selected to enable a **minimalistic and relatively low-cost ventilator design**, to avoid supply chain limitations, and to facilitate rapid and easy assembly. Most parts in the PVP are not medical-specific devices, and those that are specialized components are readily available and standardized across ventilator platforms, such as standard respiratory circuits and HEPA filters. We provide complete assembly of the PVP, including 3D-printable components, as well as justifications for selecting all actuators and sensors, as guidance to those who cannot source an exact match to components used in the Bill of Materials.

Fig. 8: PVP hardware schematic

1.1.6 Components

Fig. 9: PVP hardware schematic

1.1.6.1 Hardware Design

The following is a guided walk through the main hardware components that comprise the respiratory circuit, roughly following the flow of gas from the system inlet, to the patient, then out through the expiratory valve.

Hospital gas blender. At the inlet to the system, we assume the presence of a commercial-off-the-shelf (COTS) gas blender. These devices mix air from U.S. standard medical air and O2 as supplied at the hospital wall at a pressure of around 50 psig. The device outlet fitting may vary, but we assume a male O2 DISS fitting (NIST standard). In field hospitals, compressed air and O2 cylinders may be utilized in conjunction with a gas blender, or a low-cost Venturi-based gas blender. We additionally assume that the oxygen concentration of gas supplied by the blender can be manually adjusted. Users will be able to monitor the oxygen concentration level in real-time on the device GUI.

Fittings and 3D printed adapters. Standardized fittings were selected whenever possible to ease part sourcing in the event that engineers replicating the system need to swap out a component, possibly as the result of sourcing constraints within their local geographic area. Many fittings are American national pipe thread (NPT) standard, or conform to the respiratory circuit tubing standards (15mm I.D./22 mm O.D.). To reduce system complexity and sourcing requirements of specialized adapters, a number of connectors, brackets, and manifold are provided as 3D printable parts. All 3D printed components were print-tested on multiple 3D printers, including consumer-level devices produced by MakerBot, FlashForge, and Creality3D.

Pressure regulator. The fixed pressure regulator near the inlet of the system functions to step down the pressure supplied to the proportional valve to a safe and consistent set level of 50 psi. It is essential to preventing the over-pressurization of the system in the event of a pressure spike, eases the real-time control task, and ensures that downstream valves are operating within the acceptable range of flow conditions.

Proportional valve. The proportional valve is the first of two actuated components in the system. It enables regulation of the gas flow to the patient via the PID control framework, described in a following section. A proportional valve upstream of the respiratory circuit enables the controller to modify the inspiratory time, and does not present wear limitations like pinch-valves and other analogous flow-control devices. The normally closed configuration was selected to prevent over-pressurization of the lungs in the event of system failure.

Sensors. The system includes an oxygen sensor for monitoring oxygen concentration of the blended gas supplied to the patient, a pressure sensor located proximally to the patient mouth along the respiratory circuit, and a spirometer, consisting of a plastic housing (D-Lite, GE Healthcare) with an attached differential pressure sensor, to measure flow. Individual sensor selection will be described in more detail in a following section. The oxygen sensor read-out is used to adjust the manual gas blender and to trigger alarm states in the event of deviations from a setpoint. The proximal location of the primary pressure sensor was selected due to the choice of a pressure-based control strategy, specifically to ensure the most accurate pressure readings with respect to the patient's lungs. Flow estimates from the single expiratory flow sensor are not directly used in the pressure-based control scheme, but enable the device to trigger appropriate alarm states in order to avoid deviations from the tidal volume of gas leaving the lungs during expiration. The device does not currently monitor gas temperature and humidity due to the use of an HME rather than a heated humidification system.

Pressure relief. A critical safety component is the pressure relief valve (alternatively called the "pressure release valve", or "pressure safety valve"). The proportional valve is controlled to ensure that the pressure of the gas supplied to the patient never rises above a set maximum level. The relief valve acts as a backup safety mechanism and opens if the pressure exceeds a safe level, thereby dumping excess gas to atmosphere. Thus, the relief valve in this system is located between the proportional valve and the patient respiratory circuit. The pressure relief valve we source cracks at 1 psi (approx 70 cm H2O).

Standard respiratory circuit. The breathing circuit which connects the patient to the device is a standard respiratory circuit: the flexible, corrugated plastic tubing used in commercial ICU ventilators. Because this system assumes the use of an HME/F to maintain humidity levels of gas supplied to the patient, specialized heated tubing is not required.

Anti-suffocation check valve. A standard ventilator check valve (alternatively called a "one-way valve") is used as a secondary safety component in-line between the proportional valve and the patient respiratory circuit. The check valve is oriented such that air can be pulled into the system in the event of system failure, but that air cannot flow outward through the valve. A standard respiratory circuit check valve is used because it is a low-cost, readily sourced device with low cracking pressure and sufficiently high valve flow coefficient (Cv).

Bacterial filters. A medical-grade electrostatic filter is placed on either end of the respiratory circuit. These function as protection against contamination of device internals and surroundings by pathogens and reduces the probability of the patient developing a hospital-acquired infection. The electrostatic filter presents low resistance to flow in the airway.

HME. A Heat and Moisture Exchanger (HME) is placed proximal to the patient. This is used to passively humidify and warm air inspired by the patient. HMEs are the standard solution in the absence of a heated humidifier. While we evaluated the use of an HME/F which integrates a bacteriological/viral filter, use of an HME/F increased flow resistance and compromised pressure control.

Pressure sampling filter. Proximal airway pressure is sampled at a pressure port near the wye adapter, and measured by a pressure sensor on the sensor PCB. To protect the sensor and internals of the ventilator, an additional 0.2 micron bacterial/viral filter is placed in-line between the proximal airway sampling port and the pressure sensor. This is also a standard approach in many commercial ventilators.

Expiratory solenoid. The expiratory solenoid is the second of two actuated components in the system. When this valve is open, air bypasses the lungs, thereby enabling the lungs to de-pressurize upon expiration. When the valve is closed, the lungs may inflate or hold a fixed pressure, according to the control applied to the proportional valve. The expiratory flow control components must be selected to have a sufficiently high valve flow coefficient (Cv) to prevent obstruction upon expiration. This valve is also selected to be normally open, to enable the patient to expire in the event

of system failure.

Manual PEEP valve. The PEEP valve is a component which maintains the positive end-expiratory pressure (PEEP) of the system above atmospheric pressure to promote gas exchange to the lungs. A typical COTS PEEP valve is a spring-based relief valve which exhausts when pressure within the airway exceeds a fixed limit. This limit is manually adjusted via compression of the spring. Various low-cost alternatives to a COTS mechanical PEEP valve exist, including the use of a simple water column, in the event that PEEP valves become challenging to source. We additionally provide a 3D printable PEEP valve alternative which utilizes a thin membrane, rather than a spring, to maintain PEEP.

1.1.6.2 Actuator Selection

When planning actuator selection, it was necessary to consider the placement of the valves within the larger system. Initially, we anticipated sourcing a proportional valve to operate at very low pressures (0-50 cm H20) and sufficiently high flow (over 120 LPM) of gas within the airway. However, a low-pressure, high-flow regime proportional valve is far more expensive than a proportional valve which operates within high-pressure (~50 psi), high-flow regimes. Thus, we designed the device such that the proportional valve would admit gas within the high-pressure regime and regulate air flow to the patient from the inspiratory airway limb. Conceivably, it is possible to control the air flow to the patient with the proportional valve alone. However, we couple this actuator with a solenoid and PEEP valve to ensure robust control during PIP (peak inspiratory pressure) and PEEP hold, and to minimize the loss of O2-blended gas to the atmosphere, particularly during PIP hold.

Proportional valve sourcing. Despite designing the system such that the proportional valve could be sourced for operation within a normal inlet pressure regime (approximately 50 psi), it was necessary to search for a valve with a high enough valve flow coefficient (Cv) to admit sufficient gas to the patient. We sourced an SMC PVQ31-5G-23-01N valve with stainless steel body in the normally-closed configuration. This valve has a port size of 1/8" (Rc) and has previously been used for respiratory applications. Although the manufacturer does not supply Cv estimates, we empirically determined that this valve is able to flow sufficiently for the application.

Expiratory valve sourcing. When sourcing the expiratory solenoid, it was necessary to choose a device with a sufficiently high valve flow coefficient (Cv) which could still actuate quickly enough to enable robust control of the gas flow. A reduced Cv in this portion of the circuit would restrict the ability of the patient to exhale. Initially, a number of control valves were sourced for their rapid switching speeds and empirically tested, as Cv estimates are often not provided by valve manufacturers. Ultimately, however, we selected a process valve in lieu of a control valve to ensure the device would flow sufficiently well, and the choice of valve did not present problems when implementing the control strategy. The SMC VXZ250HGB solenoid valve in the normally-open configuration was selected. The valve in particular was sourced partially due to its large port size (3/4" NPT). If an analogous solenoid with rapid switching speed and large Cv cannot be sourced, engineers replicating our device may consider the use of pneumatically actuated valves driven from air routed from a take-off downstream of the pressure regulator.

Manual PEEP valve sourcing. The PEEP valve is one of the few medical-specific COTS components in the device. The system configuration assumes the use of any ventilator-specific PEEP valve (Teleflex, CareFusion, etc.) coupled with an adapter to the standard 22 mm respiratory circuit tubing. In anticipation of potential supply chain limitations, as noted previously, we additionally provide the CAD models of a 3D printable PEEP valve.

1.1.6.3 Sensor Selection

We selected a minimal set of sensors with analog outputs to keep the system design sufficiently adaptable. If there were a part shortage for a specific pressure sensor, for example, any readily available pressure sensor with an analog output could be substituted into the system following a simple adjustment in calibration in the controller. Our system uses three sensors: an oxygen sensor, an airway pressure sensor, and a flow sensor with availability for a fourth addition, all interfaced with the Raspberry Pi via a 4-channel ADC (Adafruit ADS1115) through an I2C connection.

Oxygen sensor. We selected an electrochemical oxygen sensor (Sensironics SS-12A) designed for the range of FiO2 used for standard ventilation and in other medical devices. The cell is self-powered, generating a small DC voltage (13-16 mV) that is linearly proportional to oxygen concentration. The output signal is amplified by an instrumentation

amplifier interfacing the sensor with the Raspberry Pi controller (see electronics). This sensor is a wear part with a lifespan of about 6 years under operation at ambient air; therefore under continuous ventilator operation with oxygen-enriched gas, it will need to be replaced more frequently. This part can be replaced with any other medical O2 sensor provided calibration is performed given that these parts are typically sold as raw sensors, with a 3-pin molex interface. Moreover, the sensor we specify is compatible with a range of medical O2 sensors, including the Analytical Industries PSR-11-917-M or the Puritan Bennett 4-072214-00, so we anticipate abundant sourcing options.

Airway pressure sensor. We selected a pressure sensor with a few key characteristics in mind: 1) the sensor had to be compatible with the 5V supply of the Raspberry Pi, 2) the sensor's input pressure range had conform to the range of pressures possible in our device (up to 70 cm H2O, the pressure relief valve's cutoff), and 3) the sensor's response time had to be sufficiently fast. We selected the amplified middle pressure sensor from Amphenol (1 PSI-D-4V), which was readily available, with a measurement range up to 70 cm H2O and an analog output voltage span of 4 V. Moreover, the decision to utilize an analog sensor is convenient for engineers replicating the design, as new analog sensors can be swapped in without extensive code and electronics modifications, as in the case of I2C devices which require modifications to hardware addresses. Other pressure sensors from this Amphenol line can be used as replacements if necessary.

Spirometer. Because flow measurement is essential for measuring tidal volume during pressure-controlled ventilation, medical flow sensor availability was extremely limited during the early stages of the 2020 COVID-19 pandemic, and supply is still an issue. For that reason, we looked for inexpensive, more easily sourced spirometers to use in our system. We used the GE D-Lite spirometer, which is a mass-produced part and has been used in hospitals for nearly 30 years. The D-Lite sensor is inserted in-line with the flow of gas on the expiratory limb, and two ports are used to measure the differential pressure drop resulting from flow through a narrow physical restriction. The third pressure-measurement port on the D-Lite is blocked by a male Luer cap, but this could be used as a backup pressure measurement port if desired. An Amphenol 5 INCH-D2-P4V-MINI was selected to measure the differential pressure across the two D-Lite takeoffs. As with the primary (absolute) pressure sensor, this sensor was selected to conform to the voltage range of the Raspberry Pi, operate within a small pressure range, and have a sufficiently fast response time (partially as a function of the analog-to-digital converter). Also, this analog sensor can be readily replaced with a similar analog sensor without substantial code/electronics modifications.

1.1.7 Assembly

PVP1 Assembly Instructions

- `PDF Version`

This guide should help you assemble the ventilator from the parts found in the ***Bill of Materials***:

We'll first show you how to assemble the hardware, then the electronics, and finally put the two together.

Current Solidworks Assembly file:

- `PVP1_Mk3.SLDASM`

Prior Versions

- `PVP1_Mk2.SLDASM`
- `PVP1_Mk1.SLDASM`

All associated parts are available in the solidworks/Parts folder in the repository.

1.1.7.1 Part 1. 3D Printed Components and Enclosure



1.1 3D Printing Adapters and Brackets.

Before we can get started with assembly, you'll need to print a few parts using a standard 3D printer. (We ran our test prints on Prusa, MakerBot, FlashForge, and Creality3D printers.)

You can download all the STL files from the CAD Page

Be sure to print airway components at as close to 100% infill as possible, and be mindful of printing orientation. An example printing setup is shown below. We do not recommend using supports or rafts unless you find them to be necessary, as they are challenging to remove. If there is a cylindrical channel, try to orient it vertically (such that the circle is traced along the build plate), which will improve circularity of the channel.



If you are using a Prusa i3, **we also provide this print setup here**.

1.2 Tapping the 3D Printed Components

Several of the 3D printed parts will need to be tapped to enable connection with other parts in the device- such as push-to-connect adapters. You will be able to tap all of these parts by hand (since the plastic cuts easily), and the ***Bill of Materials*** contains the list of taps you will need, including specialized taps such as the M16.

All required taps are shown below:



First, we will tap the Sensor Atrium component, which houses the oxygen sensor and "emergency breathing valve" check valve, and passes air from the inspiratory limb to the respiratory circuit (with a DAR filter in between).

Step 1. Use a ¼" NPT tap to thread the hole as shown below (the smaller airway hole on the flat side of the Sensor Atrium), where a push-to-connect adapter will attach.



Step 2. Use the M16 tap to thread the hole on the opposite side of the Sensor Atrium, where the oxygen sensor will attach.

Next, we will tap the adapters at either side of the expiratory solenoid.

Step 3. Use the ¾" NPT tap to thread the holes on the larger end of the two "22mm to 0.75 NPTM adapter" parts.

Last, we will tap the adapter to the pressure sampling lines.

Step 4. Finally, use the ¼"-28 tap to thread the two airway holes on the "Luer lock filter mount" part. These will hold the metal luer lock adapters to the gas sampling lines for monitoring pressure.

1.3 Cutting Enclosure Pieces.

Step 1. Laser cut, or cut out by hand, the six HPDE side panels.

Cut all panels out of the 1/16” HPDE sheets. If you wish to cut these pieces using a laser cutter, we provide DXF files in the *CAD Page*, under “Enclosure”.

Cuts can also be made by hand using a sharp pair of scissors and a razor-cutter. Pieces are 17 ” by 7 ” (45.4025 cm by 20.0025 cm), or 7 ” by 7 ” (20.0025 cm by 20.0025 cm) along the outer dimensions.

(We cut our side panels for the demo out of acrylic for ease of visibility.)



Step 2. Insert the rubber grommet(s).

The bottom panel includes a hole for a large rubber grommet. Insert this by hand.

If you wish, you may also insert a small custom grommet in the back panel.

1.1.7.2 Part 2. Basic Hardware Assembly

2.1 Assembling the bottom frame



Step 1. Cut the 80/20 ("T-slotted framing") to appropriately sized pieces: In total, you will need 4 pieces of length 17 in (45.4025 cm), and 9 pieces of length 5 in (14.9225 cm).

You can cut the aluminum 80/20 pieces by hand with a hack-saw, or using machinery such as a bandsaw. Either way, be sure to file down any rough edges afterward!

Step 2. Attach the ¾" NPT (male-male) connectors to either side of the expiratory solenoid, and then attach the two "22mm to .75 NPTM adapter" parts to those, as shown. When attaching any airway threaded parts in this assembly process, be sure to use PTFE thread sealant tape, wrapping twice around the threads, in the direction shown (away from your body if the threads are oriented towards the right).

(Note: From now on, we will not explicitly write out the need for Teflon tape; be sure to use it whenever a threaded airway component is involved!)


THREAD SEAL TAPE
HUSKY
8"
N


DC24V 10.5W
N

Step 3. Ensure that the two side ports of the pressure regulator are blocked off, with the plugs included with the part (and Teflon tape, as always). Attach a ¼" NPT push-to-connect adapter to the pressure regulator on the "outlet" side: check the bottom of the part to determine which side is "IN" for "inlet". The inlet manifold will have two ¼" NPT ports on the same side: plug one of those with the manifold plug, and attach the other end to the ¼" NPT connector. Attach the other end of this connector to the "inlet" end of the pressure regulator.

Step 4. Mounting the inlet manifold.

First, attach the 3D-printed "Inlet manifold bracket", by pushing two standard button-head screws down from the top of the piece (as shown, such that the button heads fall in the grooves), then loosely attaching a hex nut to each. Then, slide the two hex nuts into a T-slot of a short (5 ") 80/20 piece, and use an Allen key to secure the screws such that the printed piece is as far to the "right" as it can go (as shown), without the hex nuts extended past the length of 80/20.

Tip: when attaching hex nuts, make sure the side with the nub is facing down (such that it touches the inner channel of the 80/20 T-slot).

Then, insert the inlet manifold into the printed piece, as shown: if the printed part is to the right of the 80/20 length, then the pressure regulator should be oriented away from you. Insert the 2" button head hex drive screws through the remaining holes in the 3D printed piece, to secure the inlet manifold, as shown (also away from you). Finally, attach a washer (W_10_NARROW_0.406OD_316_SS) and hex nut (LN_10-32_STAINLESS_18-8) to each.


SPHA180-02-1/4
WWW.DRUMANIFOLDS.COM



Step 5. Attach the "Expiratory outlet bracket to PEEP".

Drop one standard button head hex screw down the deeper channel, and attach as before: loosely screw on a hex nut, slide it in the channel, and then use an Allen key to tighten the screw. This time, tighten the hex screw such that the printed part is in-line with the edge of the 80/20 piece, as shown. Then, insert a second button head hex screw, facing outward as shown, and loosely attach a hex nut (you will tighten this later).



Step 6. Attach corner brackets to finish this frame piece.

Pre-assemble the gusset (corner) brackets by inserting a hex screw into each hole, and then loosely attaching a hex nut to each, as shown. Tighten the gussets so that they align with the edge of the 80/20 piece, as shown, keeping the second hex nut on each piece loose. (You will use these to attach another 80/20 piece later.)



Step 7. Attach cable tie and side panel screw.

If you intend to use the cable P-clip for cable management, and to attach the HPDE side panels, attach the P-clip using a standard button head hex screw, and mount an additional hex screw for attaching the side panel later.



Progress: One piece of the bottom frame is assembled!



Step 8. Assemble the opposite side of the frame.

Assemble a second short (5 ”) piece of 80/20 as shown: this uses the 3D printed part called the “Expiratory DAR filter bracket”, which will mirror the “Expiratory outlet bracket to PEEP” part on the opposing frame leg. Also, attach three gusset (corner) brackets, as shown.

Note:

if you intend to attach the side panels, leave off the corner bracket on the same side as the 3D printed piece. If you’re not using the side panels, keep the bracket here for additional support!



Step 9. Attach side panel screws.

As before, if you plan to attach HDPE side panels later on, attach the screws for them now, as shown.



Step 10. Attach the expiratory solenoid assembly to the frame.

- First, assemble and attach two gusset (corner) brackets to their hex

  screws/nuts, then affix to a long (17 ”) leg of the framing, leaving 1” from the end of the piece on each side (that is, leaving space for a vertical piece of 80/20).

- Then, attach three hex screws/nuts to the lower rows of each

  90-degree angle bracket as shown; then attach the angle brackets by inserting a short (1mm long) button head screw through the topmost hole of each angle bracket in the opposite direction. These shorter hex screws will screw directly into the solenoid.

- Finally, slide the hex nuts attached to the angle brackets into the

  T-slot of the 80/20 piece, such that the gusset brackets and solenoid are oriented above the 80/20 piece, as shown. Loosely tighten these in place; you will adjust the position of the solenoid later.





Step 11. Attach gusset (corner) brackets to the opposing 80/20 leg.

Assemble the gusset (corner) brackets with hex screw/nuts as before, and attach them to a long (17 ”) 80/20 piece. Leave 1” of space from each of the side gussets (sufficient space for a vertical 80/20 piece), and affix the third bracket just off-center (to support a centered, vertical 80/20 piece).



Step 12. Assemble the bottom frame components.

Slide the hex nuts on the shorter 80/20 pieces you have assembled into the longer 80/20 slots, and use an Allen key to tighten the screws in place, as shown. The shorter 80/20 legs should be flush with the ends of the longer 80/20 legs when you are done, and the gusset (corner) brackets will help stabilize the frame.



Step 13. Attaching the ventilator "feet" and bottom panel.

If you are attaching the lower, perforated HPDE sheet, insert the rubber grommet. Pre-assemble the shorter (1mm) button head screws to hex nuts, and the leveling mounts (feet) to hex nuts. Use these to attach the lower HPDE panel as shown, by sliding the hex nuts along the long channels on the bottom of the frame assembly; then tighten in place.



Step 14. Attach the respiratory tube segment.

Awesome- almost done with the lower level of the ventilator! As a last step, attach the short segment of respiratory circuit between the "Expiratory outlet bracket to PEEP" and the nearest "22mm to .75NPTM adapter". You'll be able to twist this on by hand.



2.2 Assembling the frame sides



Step 1. Attach the proportional valve to its mount.

Use the two socket head screws (6mm long) to attach the proportional valve to the 3D printed "proportional valve bracket". The inlet should be to the right in the orientation shown below.



Step 2. Attach the pressure relief valve to its adapters.

Assemble the T-line push-to-connect adapter, to a female-female ¼" NPT connector, and finally to the Nylon pressure release valve.

Step 3. Attach the luer lock connectors to the luer lock filter mount.

These pieces will screw in, with the luer lock portions facing outwards.

Step 4. Assemble the sensor atrium.

Attach the ¼" NPT push-to-connect adapter, check valve, and oxygen sensor into the appropriate holes in the "Sensor Atrium Manifold", as shown. Do not push the check valve into the device too far, as it may restrict air flow within the atrium.

Step 5. Attach sensor atrium to a short 80/20, and then to the device.

Affix button head screws/nuts to the sensor atrium, with hex nuts facing towards one another, as shown. Slide the hex nuts along T-slots on either side of a short (5 ") 80/20 piece; then insert this vertically into the device as shown. Tighten into place via the hex screws on the lower gusset (corner) brackets.

Note:

If you are attaching the side panels, be sure to use a SHORT nut on the sensor atrium (the nut on the side with two, farther from the oxygen sensor), and keep the short side facing down when inserting. This will allow the sensor atrium to drop lower and align with the holes in the front panel.

Step 6. Attach the second vertical leg, along the "Expiratory DAR filter bracket".

As before, slide a short (5 ") 80/20 piece into the position shown, then tighten the hex screws.



**Step 7. Attach the luer lock filter bracket to the newly inserted 80/20 leg.\

Note:

If you're attaching the side panels, use another SHORT 80/20 nut in the uppermost spot on the bracket. We'll want this part to be as high as possible to match up with the holes on the front panel.



Step 8. Attach the third vertical 80/20 leg, then attach the proportional valve mount.

Insert a third, short (5 ”) 80/20 piece, then tighten the hex screw to hold it in place. Use another standard hex screw/nut to attach the proportional valve mount to this piece, towards the inside of the device, such that the push-to-connects roughly align vertically with the push-to-connect on the sensor atrium.



Step 9. Attach the final two vertical 80/20 pieces in the remaining corners.

Note:

If you're attaching the side panels, now is the time to insert 80/20 nuts on the vertical 80/20 pieces as well as the long 80/20 pieces on the lower frame.



Step 10. Attach the D-Lite.

Attach a blue silicone connector to the “Expiratory DAR filter bracket), on the side within the device. Then, insert the D-Lite between the silicone connector and the nearest “22mm to .75 NPTM adapter”, adjusting the position of the expiratory solenoid assembly until the D-lite is firmly connected at each end. The smaller end of the D-lite should fit within the “22mm to .75 NPTM adapter”. Then, use an Allen key to tighten the hex screws on the 90 degree angle brackets to maintain the position of the solenoid.





Step 11. Attach the pressure release valve and pneumatic tubing.

Cut the pneumatic tubing into three segments: two of length 2.5” (6.35cm), and one of length 5.5” (13.97cm). Insert the long piece between the push-to-connect adapters attached to the pressure regulator and proportional valve. Insert the two shorter tubes into either push-to-connect on the pressure relief valve assembly; then attach these between the sensor atrium and proportional valve push-to-connects, as shown. These tubing lengths should keep the proportional valve fairly centered.



2.2 Assembling the frame top



Step 1. Attach the lifting handles.

Use the lifting handle screws (SHCS_0.25-20x0.75_Gr8_ASTM_F1136), with standard hex nuts, to mount the lifting handles to two short (5 ”) pieces of 80/20, such that the handles are centered on the pieces.



Step 2. Prepare all the remaining gusset (corner) brackets.

Attach the hex screws/nuts loosely to the 13 remaining gusset (corner) brackets, so that they can be attached readily in future steps.

Step 3. Attach gusset (corner) brackets, as shown.

Step 4. Cut and punch DIN rail pieces.

Cut the DIN rail to 7 ” (20.0025cm) lengths, punching or drilling holes ½” (1.27cm) from each edge large enough to support the standard button head hex screws.



Step 5. Attach the DIN rails and assemble the rest of the top level.

- Slide two long (17 ”) 80/20 pieces onto one of the shorter pieces,

  and screw in place. Add an additional gusset (corner) bracket to each long leg, as shown, leaving 1” of space from the end (to support a vertical piece).

- Use a short (1mm) button head hex screw, a zinc washer

  (W_0.25_FLAT_THICK_GR8_YELLOW_ZINC), and a standard hex nut to attach each end of the DIN rails to the frame, as shown. Slide an additional gusset (corner) bracket between the rails on one side- this will support the central vertical channel on the device.

- Once the DIN rails are roughly in place (exact positions can be

  adjusted later), use the remaining gusset (corner) brackets and assembled short 80/20 piece, and mirror the other side of the assembly, as shown.



Step 6. Finally, slide the top frame directly into the vertical 80/20 channels of the device, and tighten all hex screws.



Congrats! The basic hardware assembly is complete!

1.1.7.3 Part 3. Electronics Assembly

3.1 Assembling the Sensor Board

[Image of all components laid out in order/piles, with labels]


4-pin, 2-pin, and 3-pin 0.1" headers
40-pin stackable RPi header
TL7660, DIP8
(Rail splitter for INA126)
10 uF, 25V capacitors
330 Ohm resistor
INA126, DIP8
(Instrumentation
amplifier
for O2 sensor)
2-layer Sensor Board PCB
Amphenol 5 INCH-D2-PV4-MINI
Differential pressure sensor
Adafruit ADS1115
4x 12-bit ADC
Amphenol 1 PSI-D-4V-MINI
Pressure sensor



You will also need:

- A soldering iron
- Solder
- Helping hands for holding parts while soldering
- Wire cutters (for clipping off long capacitor/resistor legs)

Step 1. Solder the 40-pin stackable RPi header to the Actuator PCB.

Push the pins up from the bottom of the board, as shown, and then solder into place.



Step 2. Solder the 4-pin, 3-pin, and 2-pin 0.1" headers onto the board (positions J2, J3, J4).

Break the pins off the larger header array in units of 4, 2, and 3 using pliers. Insert the short end of the pins into the holes from the top of the board, and solder from below. This will leave the long ends of the pins sticking up vertically from the board, as shown:



Step 3. Solder the 330 Ohm resistor onto the board (position R1).

Bend the resistor legs before inserting into the board from above. Ideally, solder the resistor in place such that the resistor is hovering just above the board. Then, snip the long legs off from the back of the board (cutting above the tiny "cone" formed by the solder) using wire cutters.



Step 4. Solder the two 10 µF, 25V capacitors onto the board (positions C1, C2).

Insert the capacitors from the top of the board until the legs snap into place. Be sure that the longer capacitor leg is inserted into the hole corresponding to the "+" sign. After soldering, snip the long legs off from the back of the board using wire cutters.



Step 5. Solder the TL7660 (the rail splitter for the INA126) onto the board (position U1).

Be sure to orient the small circle on the top of the part to the indicated notch drawn on the board.



Step 6. Solder the INA126 (the instrumentation amplifier for the oxygen sensor) onto the board (position U2).

As before, make sure that the notch on the part aligns with the notch drawn on the PCB.



Step 7. Solder the Amphenol 5 INCH-D2-P4V-MINI (differential pressure sensor) onto the board (position U3).

Bend all four pins evenly and orient the part such that the black ports face towards the outside of the board. The pressure lines will attach here.



Step 8. Solder the Adafruit ADS1115 (12-bit ADC) to the board.

First, solder the pins to the ADC itself, with the long ends facing down. Then insert the pins through the board and solder from the back side.


VDD
GND
SCL
SDA
ADDR
ALRT
A0
A1
A2
A3
12Bit I2C ADS1015 ADC+PGA



Step 9. Finally, solder the Amphenol 1 PSI-D-4V-MINI (airway pressure sensor) to the board.

As before, bend all four pins together, then insert and solder from the back side of the board. Be sure the ports are facing out: we will attach pressure lines here as well.



3.2 Assembling the Actuator Board

[Image of all components laid out in order/piles, with labels]





You will also need:

- A soldering iron

- Solder
- Helping hands for holding parts while soldering
- Wire cutters (for clipping off long capacitor legs)

Step 1. Solder the 40-pin stackable RPi header to the Actuator PCB.

Push the pins up from the bottom of the board, as shown, and then solder into place.





Step 2. Solder the 2-pin 0.1” header onto the board (position J5).

Insert the short end of the pins into the holes from the top of the board, and solder from below. This will leave the long ends of the pins sticking up vertically from the board, as shown:



Step 3. Solder the 100 µF, 16V capacitor onto the board (position C1).

Insert the legs of the capacitor into the slots from the top of the board. The longer leg should be inserted into the side marked "+". Once the capacitor is soldered in place, use wire cutters to clip off the long legs from the back.



Step 4. Solder the ULN2003A (Darlington array) into place (position U1).

Be sure to match the notched end of the part to the notch indicated on the board.



Step 5. Solder the two 6.8 µF, 50V capacitors onto the board (positions C2, C3).

As before, ensure that the longer capacitor legs are inserted through the side marked with the "+". Once the parts are soldered in, snip the long legs from the back of the board.



Step 6. Solder the 24-to-5 V DC-DC converter (CUI PDQ15-Q24-S5-D) onto the board, as shown.



Step 7. Solder the 3 2-pin screw terminals (5.08mm pitch) to the board, as shown.

Be sure to orient all three such that wires can be screwed into the terminals from the outside of the board (as shown).


J2+J3
J1+J2
On/off Valve
(BCM 17)
On/off Valve
(BCM 27)
J3
J2
GND+24V
U2
PDQE15-Q24-S5-D
IN 9-36VDC
OUT 5VDC/3000mA
RoHS
2030
CUI INC


Tiger Vents
Actuator Board
Rev 1.0
Prop Valve
(BCM 12)
J2+J3
On/off Valve
(BCM 17)
On/off Valve
(BCM 27)
C1
C3
U1
U2
J2
+5 V to Pi
GND+24V
PDQE15-Q24-S5-D
IN 9-36VDC
OUT 5VDC/3000mA
RoHS
2030
CUI INC

Step 8. Finally, jumper the 2-pin header.

You can use a pin jumper ("shunt") if you have one, or crush the pins together with a pair of pliers, and then solder the two pins together. We display the latter method, below:



Step 9. The actuator board is ready to go!


Tiger Vents
Actuator Board
Rev 1.0
Prop Valve
(BCM 12)
J3
J2+J3
On/off Valve
(BCM 17)
J1+J2
J2
On/off Valve
(BCM 27)
C3
GND+24V
U2
U1
ULN2003AN
C1
J6
+5 V to PI
0V
Trim
+Vo
PDQE15-Q24-S5-D
IN 9-36VDC
OUT 5VDC/3000mA
RoHS
2030
Ctrl
GND
Vin
CUI INC

3.3 Assembling the PCB-RPi stack



Step 1. Attach three gas sampling lines to the port on the two pressure sensors.

The differential pressure sensor has two ports. If you have an airway-safe glue, feel free to glue these in place (not required).



Step 2. Begin stacking the board by attaching the DIN rail mounts.

You can pre-"tap" the 3D printed DIN rail mounts using the small screws that come with the Raspberry Pi. Then, screw four of the 16mm standoffs through the Raspberry Pi from above, and into the DIN rail mounts, as shown.

Step 3. Next, add the sensor board.

Connect the sensor board via the stackable headers, then use another four 16mm standoffs to attach the board by screwing these into the original four standoffs, using pliers as needed.



Step 4. Add the actuator board.

Connect the actuator board via the stackable headers, then use the four screws that came with your Raspberry Pi to attach the board, as shown.



1.1.7.4 Part 4. Putting it all together

4.1 Wrapping it up.



(Optional step: Attach the side panels using the shorter 80/20 hex screws, as you go.)

Step 1. Attach the DAR filters.

Insert these by hand to the appropriate ports on the front of the device, as shown: these will attach to the sensor atrium and the expiratory DAR filter bracket. The DAR filters indicate which end should face the patient.



Step 2. Attach the luer lock filter.

This can be twisted on by hand to the luer lock adapter at the front of the device.



Step 3. Attach the PEEP valve.

If you're using a commercial PEEP valve, begin by attaching the blue silicone connector. Then, insert the PEEP valve into the 3D printed "22mm to commercial PEEP adapter"- then plug this assembly into the silicone connector.



Step 4. Attach the power supply.

Pull the tab on the Meanwell power supply to attach it to the DIN rail closest to the back of the device. It may be necessary to adjust the position of this DIN rail to ensure that the power supply is not touching the expiratory solenoid or inlet manifold/pressure regulator. Feel free to reposition as needed!

Note:

if you wish to attach a rear HPDE panel, we recommend inserting the power cable to the power supply through the grommet of the rear panel before insertion of the power supply into the device. Do this now! You can also attach the "Rear Panel Vent"s to the rear panel; they will snap right in.



Step 5. Attach the luer lock plug to the D-Lite.

The third port on the D-Lite (physically removed from the other two) is not needed; plug that now with a luer lock plug.



Step 6. Attach the power and ground wires to the PCB stack. Then wire the proportional valve and solenoid.

Use a small flathead screwdriver to attach the power and ground from the power supply to the PCB stack where indicated. Then do the same for the proportional valve wires and the expiratory solenoid wires: the board indicates which pair goes where!



Step 7. Attach the gas sampling lines to the D-Lite.

Use the luer lock connector on the smaller of the two ports, then connect the two gas sampling lines from the differential pressure sensor (from the same sensor on the Sensor Board).



Step 8. Attach the gas sampling line to the luer lock filter mount adapter.

The third gas sampling line runs from the second pressure sensor; attach this to the inner luer lock adapter on the luer lock filter mount.



Step 9. Plug in any required cables to connect the monitor and keyboard.

Run the cables outside of the box through the rubber grommets; wire ties can assist with cable management.

Congratulations! You're done!



1.1.8 Electronics

The PVP is coordinated by a Raspberry Pi 4 board, which runs the graphical user interface, administers the alarm system, monitors sensor values, and sends actuation commands to the valves. The core electrical system consists of two modular PCB 'hats', a sensor PCB and an actuator PCB, that stack onto the Raspberry Pi via 40-pin stackable headers. The modularity of this system enables individual boards to be revised or modified to adapt to component substitutions if required.

We outsourced our PCB fabrication to Advanced Circuits, based out of Aurora, CO (for $33 each): https://www.4pcb.com/pcb-prototype-2-4-layer-boards-specials.html If you would like to do the same, you can send them the Gerber .zip files we have provided directly.



Fig. 10: PVP block diagram for main electrical components

1.1.8.1 Power and I/O

The main power to the systems is supplied by a DIN rail-mounted 150W 24V supply, which drives the inspiratory valve (4W) and expiratory valves (13W). This voltage is converted to 5V by a switched mode PCB-mounted regulated to power the Raspberry Pi and sensors. This power is transmitted across the PCBs through the stacked headers when required.

Table 2: Power and I/O bill of materials

Part	Description
Meanwell 24 V DC Power Supply	DIN Rail Power Supplies 150W 24V 5A EN55022 Class B
Raspberry Pi	Raspberry Pi- Model B-1 (1GB RAM)
USB-C Charger/cable	To power the RPi
Micro SD Card	SanDisk Ultra 32GB MicroSDHC UHS-I Card with Adapter
Raspberry Pi Display	Matrix Orbital: TFT Displays & Accessories 7 in HDMI TFT G Series
HDMI for Display	Display cable: HDMI Cables HDMI Cbl Assbly 1M Micro to STD
Mini USB for Display	Display cable: USB Cables / IEEE 1394 Cables 3 ft Ext A-B Mini USB Cable
Screen mount thumb screws	SCREEN_MOUNT_THUMB_SCREW: Brass Raised Knurled-Head Thumb Screw, 1/4”-20 Thread Size, 1/2” Long
Cable grommet	USER_INTERFACE_CABLE_GROMMET: Buna-N Rubber Grommets, for 1-3/8” Hole Diameter and 1/16” Material Thickness, 1” ID, pack of 10
Cable P-clip	USER_INTERFACE_CABLE_P-CLIP_0.375_ID_SS: Snug-Fit Vibration-Damping Loop Clamp, 304 Stainless Steel with Silicone Rubber Cushion, 3/8” ID, pack of 10, 17/64 mounting holes
Keyboard	Adesso: Mini keyboard with trackball

1.1.8.2 Sensor PCB

The sensor board interfaces four analog output sensors with the Raspberry Pi via I2C commands to a 12-bit 4-channel ADC (Adafruit ADS1015).

1. an airway pressure sensor (Amphenol 1 PSI-D-4V-MINI)
2. a differential pressure sensor (Amphenol 5 INCH-D2-P4V-MINI) to report the expiratory flow rate through a D-Lite spirometer
3. an oxygen sensor (Sensiron SS-12A) whose 13 mV differential output signal is amplified 250-fold by an instrumentation amplifier (Texas Instruments INA126)
4. a fourth auxiliary slot for an additional analog output sensor (unused)

A set of additional header pins allows for digital output sensors (such as the Sensiron SFM3300 flow sensor) to be interfaced with the Pi directly via I2C if desired.

- `Sensor PCB - [KiCad project .zip]`



Fig. 11: Sensor PCB schematic

Table 3: Sensor PCB bill of materials

Ref	Part	Purpose
J1	40-pin stackable RPi header	Connects board to RPi
J2	4-pin 0.1” header	I2C connector if desired
J3	2-pin 0.1” header	Connects ALRT pin from ADS1115 to RPi if needed
J4	3-pin 0.1” header or 3 pin fan extension cable	Connects board to oxygen sensor
R1	330 Ohm resistor	Sets gain for INA126
C1	10 uF, 25V	Cap for TL7660
C2	10 uF, 25V	Cap for TL7660
U1	TL7660, DIP8	Rail splitter for INA126
U2	INA126, DIP8	Instrumentation amplifier for oxygen sensor output
U3	Amphenol 5 INCH-D2-P4V-MINI	Differential pressure sensor (for flow measurement)
U4	Adafruit ADS1115	4x 12-bit ADC
U5	Amphenol 1 PSI-D-4V-MINI	Airway pressure sensor
U6		Auxiliary analog output sensor slot

1.1.8.3 Actuator PCB

The purpose of the actuator board is twofold:

1. regulate the 24V power supply to 5V (CUI Inc PDQE15-Q24-S5-D DC-DC converter)
2. interface the Raspberry Pi with the inspiratory and expiratory valves through an array of solenoid drivers (ULN2003A Darlington transistor array)

- `Actuator PCB - [KiCad project .zip]`



Fig. 12: Actuator PCB schematic

Table 4: Actuator PCB bill of materials

Ref	Part	Purpose
J2	2-pin screw terminal, 5.08 mm pitch, PCB mount	Connects to 24V supply
J3	2-pin screw terminal, 5.08 mm pitch, PCB mount	Connects to on/off expiratory valve
J4	2-pin screw terminal, 5.08 mm pitch, PCB mount	Connects to inspiratory valve, driven by PWM
J5	40-pin stackable RPi header	Connects board to RPi
J6	2-pin 0.1” header	Jumper between 5V and Raspberry Pi
C1	100 uF, 16V	5V rail filter cap
C2	6.8 uF, 50V	24V rail filter cap
C3	6.8 uF, 50V	24V rail filter cap
U1	ULN2003A	Darlington BJT array to drive solenoids
U2	CUI PDQ15-Q24-S5-D	24-to-5V DC-DC converter

1.1.9 Bill of Materials

The complete bill of materials is available here: * `Bill of materials - [.csv]`

The following tables are summarized versions of the CSV file above, and illustrate the approximate unit cost (state 10/2020) of a single device, together with links to manufactures of the parts:

GRAND TOTAL: $1,753.57

1.1.9.1 I/O

Table 5: PVP bill of materials

Component	Subtotal	Example Sourcing Link	
Raspberry Pi	$53.00	https://www.mouser.com/ProductDetail/Raspberry-Pi/RPI4-MODBP-2GB-BULK?qs=sGAEpiMZZMspCjQQiuQ1fFuhMF5SnQS%252Ba2qMvWLE7K2IHh0t%252BecCsw%3D%3D	
USB-C Charger/cable	$9.88	https://www.mouser.com/ProductDetail/Seeed-Studio/106990291?qs=sGAEpiMZZMtyU1cDF2RqUIiI0Dr66jYXGV7bLfwtXvw%3D	
Micro SD Card	$8.29	https://www.amazon.com/dp/B073JWXGNT/ref=twister_B07B3MFBHY?_encoding=UTF8&th=1	
7” Display	$249.18	https://lilliputdirect.com/lilliput-779gl-capacitive-touchscreen-hdmi-monitor?search=779gl	
Screen Mount	$6.00	https://lilliputdirect.com/lilliput-monitor-accessories/mounting-brackets-stands-and-arms/lilliput-monitor-stand	
HDMI for Display	$19.59	https://www.mouser.com/ProductDetail/molex/68786-0001/?qs=ASaot9jrY6HVNv%2F8Wfc2rQ%3D%3D&countrycode=US&currencycode=USD	
Mini USB for Display	$5.49	https://www.mouser.com/ProductDetail/matrix-orbital/extmusb3ft/?qs=4ybA1OVEHcjRr1VgZWmnlw%3D%3D&countrycode=US&currencycode=USD	
Cable grommet	$10.08	https://www.mcmaster.com/9307K61/	
Cable P-clip	$7.00	https://www.mcmaster.com/3225T63/	
Keyboard	$49.99	https://m.cdw.com/product/adesso-easy-touch-mini-trackball-keyboard/1839486	
	$8.15	https://www.grainger.com/product/1FD82?gclid=Cj0KCQjw1qL6BRCmARIsADV9JtZk59QIAYAwMJvQFD1dzPK-rSwcB&gucid=N:N:FPL:Free:GGL:CSM-1946:tew63h3:20501231&ef_id=Cj0KCQjw1qL6BRCmARIsADV9JtZk59QIAYAwMJvQFD1dzPK-rSkWwVBCeo8SBGndpEHDpp2NXDWsr9kaAo-wcB:G:s&s_kwcid=AL!2966!3!264922886802!!!g!439460816581!&gucid=N:N:PS:Paid:GGL:CSM-2293:FAGP9P:20500731	
Power strip	$3.63	https://www.parts-express.com/parts-express-3-outlet-strip-with-3-ft-cord-ul-black–125-406?gclid=EAIaIQobChMImKulw4PM6wIVRymzAB1bcwgfEAQYAyABEgLhBPD_BwE	
TOTAL	$419		

1.1.9.2 Airway

Component	Subtotal	Example Sourcing Link
Oxygen DISS to 3/8” NPT Adapter	$10.00	www.mcmaster.com/2633N11/
Inlet manifold	$24.56	https://www.mcmaster.com/1023N15/
1/4” NPT Connector (male/male)	$3.89	https://www.mcmaster.com/1455N133/
Teflon tape	$2.62	https://www.mcmaster.com/6802K12/
Manifold plug 1/4” NPT	$4.50	https://www.mcmaster.com/1725K13/
Hex nuts	$3.77	https://www.mcmaster.com/91841A195/
Manifold washers	$10.16	https://www.mcmaster.com/92217A440/
Fixed pressure regulator	$47.94	https://www.mcmaster.com/1777K2
Push-to-connect: 1/4” NPT	$10.96	https://www.mcmaster.com/5225K715/
Pneumatic Tubing	$2.13	https://www.mcmaster.com/50315K71
Push-to-connect: 1/8” BSPT	$10.76	https://www.mcmaster.com/5225K305/
Insp proportional valve	$107.10	https://www.ocpneumatics.com/smc-pvq31-5g-23-01n-h-valve-proport
Push-to-connect: Inline T-adapter to 1/4” NPT	$7.58	https://www.mcmaster.com/5225K806/
1/4” NPT Connector (female/female)	$9.84	https://www.mcmaster.com/4589K58/
Pressure release valve	$19.89	https://www.mcmaster.com/4277T52
Oxygen sensor	$59.00	https://www.sensoronics.com/products/ss-12a-replaces-teledyne-r22-m
Emergency breathing (check) valve	$1.05	https://serfinitymedical.com/products/teleflex-medical-one-way-valve-t
DAR filters	$3.44	https://www.bettymills.com/dar-electrostatic-filter-large-350u5865?utm
Adult Respiratory Circuit w Humidifier Limb	$13.00	Circuit: https://orsupply.com/product/1505
Gas sampling line	$5.01	https://heymedsupply.com/gas-sampling-line-mckesson-16-gsl-each-1/
Luer lock connector	$20.50	https://www.grainger.com/product/36T564?gclid=EAIaIQobChMI59a9
Luer lock filter		https://catalog.promepla.com/promepla-component/ecb15992 / https://c
Pressure sensor (for Paw)	$32.00	https://www.mouser.com/ProductDetail/Amphenol-All-Sensors/1-PSI-
Tubing connectors	$14.00	https://www.rcmedical.com/viewProduct.cfm?productID=591
Flow sensor	$4.76	https://jakenmedical.com/supplies-accessories/disposable-d-lite-sensor-
Luer plug	$2.60	https://www.mcmaster.com/51525K311
Luer lock connector for D-Lite	$22.68	https://www.mcmaster.com/5194K32/
Differential pressure sensor for flow	$52.70	https://www.mouser.com/ProductDetail/Amphenol-All-Sensors/5-INCH
3/4” NPT Connector (male/male)	$3.06	https://www.mcmaster.com/46825K31/
Exp solenoid valve	$115.10	https://www.smcpneumatics.com/VXZ250HGB.html
Exp solenoid brackets	$22.44	https://www.mcmaster.com/3136N157/

1.1.9.3 Framing & Structural

Table 7: PVP bill of materials

Component	Subtotal	Example Sourcing Link	
80/20 (x4)	$42.28	https://www.mcmaster.com/47065T101	
80/20 gusset bracket (x18)	$117.72	https://www.mcmaster.com/47065T663	
80/20 rail (x6)	$35.04	https://www.mcmaster.com/47065T101	
80/20 button head nuts (packs of 25, x2)	$11.24	https://www.mcmaster.com/47065T905/	
Button-head hex screws, pack of 50	$6.03	https://www.mcmaster.com/92949A833/	
Socket head screw, pack of 100	$9.66	https://www.mcmaster.com/91292A111/	
Button head screws (long), pack of 10	$2.65	https://www.mcmaster.com/92949A275/	
Button head screws (short), pack of 25	$8.40	https://www.mcmaster.com/92095A239/	
HDPE sheeting	$21.28	https://www.mcmaster.com/8619K423	
Lifting handle	$28.76	https://www.mcmaster.com/5190A21/	
Screws for lifting handle	$9.69	https://www.mcmaster.com/91274A164/	
Leveling mount (4 feet)	$1.89	https://www.mcmaster.com/23015T82/	
Nuts for leveling mount	$4.41	https://www.mcmaster.com/91847A029/	
PETG for 3d printing	$30.00		
TOTAL	$329		

1.1.9.4 Electronics

Table 8: PVP bill of materials

Component	Subtotal	Example Sourcing Link	
Standoffs	$3.76	https://www.mouser.com/ProductDetail/Keystone-Electronics/24424?qs=%2Fha2pyFadujGjYnOBX3ZIG5%252BNlfnctDXUaVF01IeI1E%3D	
Meanwell DC Power Supply	$30.66	https://www.mouser.com/ProductDetail/MEAN-WELL/NDR-120-24?qs=sGAEpiMZZMvkjJSSRowFE%252BMaPtHpNQit89EjeOStv9CBsXX2JvUmCQ%3D%3D	
DIN Rail	$7.74	https://www.mcmaster.com/8961K17/	
DIN washers	$13.94	https://www.mcmaster.com/98025A029/	
Sensor board (excluding pressure sensors)			
2-layer PCB	$99.00	https://www.4pcb.com/pcb-prototype-2-4-layer-boards-specials.html	
12-bit ADC	$10.00	https://www.mouser.com/ProductDetail/Adafruit/1083?qs=GURawfaeGuCmI2li4B6pKg%3D%3D	
Voltage splitter U1	$2.00	https://www.digikey.com/product-detail/en/texas-instruments/TL7660CP/296-21857-5-ND/1629031	
Instrumentation amplifier U2	$4.00	https://www.digikey.com/product-detail/en/texas-instruments/INA126PA/INA126PA-ND/300992	
R3, sensor	$0.00		
40 pin Pi header J1	$3.00	https://www.digikey.com/product-detail/en/adafruit-industries-llc/1979/1528-1783-ND/6238003	
4x1 header J2	$5.00	https://www.digikey.com/product-detail/en/adafruit-industries-llc/4150/1528-2922-ND/10123801	
2x1 header J3	$5.00		
Oxygen sensor cable (J4)	$2.00	https://www.coolerguys.com/products/3-pin-fan-cable-extension-12-thru-72-inches?variant=17666421509	
330 Ohm resistor R1	$1.00	https://www.digikey.com/product-detail/en/stackpole-electronics-inc/CF14JT330R/CF14JT330RCT-ND/1830338	
Capacitors C1,C2	$1.00	https://www.digikey.com/product-detail/en/panasonic-electronic-components/ECA-1EM100B/P19522CT-ND/6109420	
Actuator board			
2-layer PCB	$99.00	https://www.4pcb.com/pcb-prototype-2-4-layer-boards-specials.html	
5V DC/DC Converter	$20.00	https://www.mouser.com/ProductDetail/CUI-Inc/PDQE15-Q24-S5-D?qs=%2Fha2pyFaduiL94u0Ef2I7K4U1Mc9B6IPlz0S4%252BeNn6w9pBF%2FNnZWZTmWI4mUuuL4	
Darlington array U1	$2.00	https://www.mouser.com/ProductDetail/Texas-Instruments/ULN2003AN?qs=FOlmdCx%252BAA1wYQ1G8c8hpQ%3D%3D	
40 pin Pi header J1	$3.00	https://www.digikey.com/product-detail/en/adafruit-industries-llc/1979/1528-1783-ND/6238003	
Screw terminals J2, J3, J4	$4.00	https://www.digikey.com/product-detail/en/te-connectivity-amp-connectors/282837-2/A113320-ND/2187973	
2x1 header J5	$5.00		
Capacitor C1	$1.00	https://www.digikey.com/product-detail/en/panasonic-electronic-components/EEU-FR1C101B/P15330CT-ND/3072210	
Capacitors C2,C3	$1.00	https://www.digikey.com/product-detail/en/panasonic-electronic-components/EEA-GA1H6R8/P14509-ND/2504598	
Speaker	$20.00	https://www.amazon.com/Z50-smartphone-tablet-laptop-Grey/dp/B00EZ9XLF8/ref=sr_1_1?dchild=1&hvadid=78408975043567&hvbmt=be&hvdev=c&hvqmt=e&keywords=logitech+z50&qid=1595253882&sr=8-1&tag=mh0b-20	
TOTAL	$347		

1.1.9.5 Specialized Tools

Table 9: PVP bill of materials

Component	Subtotal	Example Sourcing Link	
M16 tap	$47.00	https://www.mcmaster.com/26015A236	
1/4” NPT tap	$25.00	https://www.mcmaster.com/2525A173-2525A173	
3/4” NPT tap	$71.00	https://www.mcmaster.com/2525A176/	
1/4”-28 tap	$6.99	https://www.mcmaster.com/26955A88/	
Ingmar QuickLung Test Bellows	$1,500.00	https://www.ingmarmed.com/product/quicklung/	

1.1.9.6 Etc.

Table 10: PVP bill of materials

Component	Subtotal	Example Sourcing Link	
PETG for 3d printing	$30.00		
Commercial PEEP valves			
Adult Test Lung		https://www.ebay.com/itm/Adult-Test-Lung-LNG600P-Newport-Medical-Instruments-a-division-of-Covidien/264679192553?hash=item3da01bf7e9:g:9FEAAOSw~4hbtoaY&fbclid=IwAR35qedvGn5JODcB-F5Kurw0ZH11KOdH6UpmUypnj2lWZcv5GzIfVlEDSaw	
Air compressor for initial testing		https://www.dewalt.com/products/storage-and-gear/air-compressors/16-hp-continuous-225-psi-45-gallon-compressor/d55146	

1.1.10 CAD

1.1.10.1 3D Printed Parts

Individual 3D printed parts may be downloaded here:

- `Inlet manifold bracket - [.stl]`
- `Proportional valve bracket - [.stl]`
- `Sensor atrium manifold - [.stl]`
- `Expiratory DAR filter bracket - [.stl]`
- `22mm to 0.75 NPTM adapter (x2) - [.stl]`
- `Expiratory outlet bracket to PEEP - [.stl]`
- `Luer lock filter mount - [.stl]`
- `Raspberry Pi DIN rail mount (x2) - [.stl]`
- `Rear panel vent (x2) - [.stl]`
- `22mm to commercial PEEP adapter - [.stl]`



Fig. 13: Sample Prusa project with 3D printed components

Optional RPi DIN rail mount alternative, if your printer cannot handle the feature resolution on the original DIN rail mount:

- `ALTERNATIVE Raspberry Pi DIN rail mount (x2) - [.stl]`

Download all parts:

- `All components - [.zip]`

Printing tips: Be sure to maintain high infill for airway components: ideally, use 4 wall-layers (vertical layers), minimum 40% infill, and 4 layers on both the top and bottom. Parts should be leak-tested prior to installation, for instance, with a stopper and water bath. Most of our test prints were performed using PLA, and without any supports or rafts, since these are challenging to remove later. Supports should not be necessary provided the parts are oriented mindfully on the build plate. Also, try to keep cylindrical components oriented vertically (so that the circle is traced on the build plate); this will improve circularity of the chamber.

For PRUSA users, we provide an example project, demonstrating part orientation:

- `Sample Prusa project - [.3mf]`

1.1.10.2 Enclosure

The side, top, and bottom panels are made out of 1/16" HPDE sheeting. Laser cut, or cut by hand, two of the "SIDE_IP_PANEL", and one of each of the rest! We use a perforated HPDE for the bottom panel only; the rest are solid HPDE.

- `Bottom panel (perforated) - [.dxf]`
- `Front panel - [.dxf]`
- `Rear panel - [.dxf]`
- `Side panel (x2) - [.dxf]`
- `Top panel - [.dxf]`

Download all .DXF files:

- `All DXF files - [.zip]`

1.1.11 Software Overview

PVP is modularly designed to facilitate future adaptation to new hardware configurations and ventilation modes. APIs were designed for each of the modules to a) make them easily inspected and configured and b) make it clear to future developers how to adapt the system to their needs.

PVP runs as multiple independent processes The GUI provides an interface to control and monitor ventilation, and the controller process handles the ventilation logic and interfaces with the hardware. Inter-process communication is mediated by a coordinator module via xml-rpc. Several 'common' modules facilitate system configuration and constitute the inter-process API. We designed the API around a uni-fied, configurablevaluesmodule that allow the GUI andcontroller to be reconfigured while also ensuring system robustness and simplicity.

- The *GUI* and *Coordinator* run in the first process, receive user input, display system status, and relay `ControlSetting` s to the *Controller* .
- At launch, the *Coordinator* spawns a *Controller* that runs the logic of the ventilator based on control values from the GUI.
- The *Controller* communicates with a third pigpiod process which communicates with the ventilation hardware

PVP is configured by

- The *Values* module parameterizes the different sensor and control values displayed by the GUI and used by the controller
- The *Prefs* module creates a `prefs.json` file in `~/pvp` that defines user-specific preferences.

PVP is launched like:

```
python3 -m pvp.main
```

And launch options can be displayed with the `--help` flag.

1.1.12 Folder Structure

The repository is organized as follows:

- *pvp/assets/* contains technical information like CAD drawings, circuit diagrams.
- *pvp/data/* contains information for calibrating sensors; data and information.
- *pvp/_docs* and pvp/docs is raw, and built documentation.
- *pvp/tests* contains automated tests for all software modules.
- *pvp/sandbox* is experimental code, that is not necessary to operate pvp. Place your toy programs here.
- *pvp/pvp* is the main code. It contains individual files for all modules of PVP1. Binary files like audio/graphics are deposited with the respective module, and not collected in a central site.

1.1.12.1 PVP Modules

1.1.13 GUI

1.1.13.1 Main GUI Module

Classes:

PVP_Gui(coordinator[, set_defaults, ...])	The Main GUI window.

Functions:

launch_gui(coordinator[, set_defaults, ...])	Launch the GUI with its appropriate arguments and doing its special opening routine

class pvp.gui.main.**PVP_Gui**(*coordinator:* pvp.coordinator.coordinator.CoordinatorBase, *set_defaults: bool = False*, *update_period: float = 0.05*, *screenshot=False*)

The Main GUI window.

Creates 5 sets of widgets:

- A *Control_Panel* in the top left corner that controls basic system operation and settings
- A *Alarm_Bar* along the top that displays active alarms and allows them to be dismissed or muted
- A column of *Display* widgets (according to *values.DISPLAY_MONITOR*) on the left that display sensor values and control their alarm limits
- A column of *Plot* widgets (according to *values.PLOTS*) in the center that display waveforms of sensor readings
- A column of *Display* widgets (according to *values.DISPLAY_CONTROL*) that control ventilation settings

Continually polls the `coordinator` with *update_gui()* to receive new *SensorValues* and dispatch them to display widgets, plot widgets, and the alarm manager

Note: Only one instance can be created at a time. Uses `set_gui_instance()` to store a reference to itself. after initialization, use *get_gui_instance* to retrieve a reference.

Parameters

- **coordinator** (*CoordinatorBase*) – Used to communicate with the *ControlModuleBase* .
- **set_defaults** (*bool*) – Whether default *Value* s should be set on initialization (default `False`) – used for testing and development, values should be set manually for each patient.
- **update_period** (*float*) – The global delay between redraws of the GUI (seconds), used by *timer* .
- **screenshot** (*bool*) – Whether alarms should be manually raised to show the different alarm severities, only used for testing and development and should never be used in a live system.

monitor

Dictionary mapping *values.DISPLAY_MONITOR* keys to `widgets.Display` objects

Type dict

controls
Dictionary mapping *values.DISPLAY_CONTROL* keys to `widgets.Display` objects

Type dict

plot_box
Container for plots

Type `Plot_Box`

coordinator
Some coordinator object that we use to communicate with the controller

Type *pvp.coordinator.coordinator.CoordinatorBase*

alarm_manager
Alarm manager instance

Type *Alarm_Manager*

timer
Timer that calls *PVP_Gui.update_gui()*

Type `PySide2.QtCore.QTimer`

running
whether ventilation is currently running

Type bool

locked
whether controls have been locked

Type bool

start_time
Start time as returned by `time.time()`

Type float

update_period
The global delay between redraws of the GUI (seconds)

Type float

logger
Logger generated by *loggers.init_logger()*

Attributes:

`gui_closing`(*args, **kwargs)	`PySide2.QtCore.Signal` emitted when the GUI is closing.
`state_changed`(*args, **kwargs)	`PySide2.QtCore.Signal` emitted when the gui is started (True) or stopped (False)
`MONITOR`	Values to create *Display* widgets for in the Sensor Monitor column.
`CONTROL`	Values to create *Display* widgets for in the Control column.
`PLOTS`	Values to create *Plot* widgets for.
`monitor_width`	Relative width of the sensor monitor column
`plot_width`	Relative width of the plot column
`control_width`	Relative width of the control column

continues on next page

Table 13 – continued from previous page

total_width	computed from monitor_width+plot_width+control_width
staticMetaObject	
controls_set	Check if all controls are set
update_period	The global delay between redraws of the GUI (seconds)

Methods:

update_gui([vals])	**param vals** Default None, but SensorValues can be passed manually -- usually for debugging
init_ui()	0. Create the UI components for the ventilator screen
init_ui_status_bar()	1. Create the widgets.Control_Panel and widgets.Alarm_Bar
init_ui_monitor()	2. Create the left "sensor monitor" column of widgets.Display widgets
init_ui_plots()	3. Create the *Plot_Container*
init_ui_controls()	4. Create the "controls" column of widgets.Display widgets
init_ui_signals()	5. Connect Qt signals and slots between widgets
set_value(new_value[, value_name])	Set a control value using a value and its name.
set_control(control_object)	Set a control in the alarm manager, coordinator, and gui
handle_alarm(alarm)	Receive an *Alarm* from the *Alarm_Manager*
limits_updated(control)	Receive updated alarm limits from the *Alarm_Manager*
start()	Click the start_button
toggle_start(state)	Start or stop ventilation.
closeEvent(event)	Emit *gui_closing* and close!
toggle_lock(state)	Toggle the lock state of the controls
update_state(state_type, key, val)	Update the GUI state and save it to disk with Vent_Gui.save_state()
save_state()	Try to save GUI state to prefs['VENT_DIR"] + prefs['GUI_STATE_FN']
load_state(state)	Load GUI state and reconstitute

continues on next page

Table 14 – continued from previous page

toggle_cycle_widget(button)	Set which breath cycle control is automatically calculated
set_pressure_units(units)	Select whether pressure units are displayed as "cmH2O" or "hPa"
set_breath_detection(breath_detection)	Connected to breath_detection_button - toggles autonomous breath detection in the controller
get_breath_detection()	Get the current state of breath detection from the controller
_set_cycle_control(value_name, new_value)	Compute the computed breath cycle control.
init_controls()	on startup, set controls in coordinator to ensure init state is synchronized
_screenshot()	Raise each of the alarm severities to check if they work and to take a screenshot

gui_closing(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
PySide2.QtCore.Signal emitted when the GUI is closing.

state_changed(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
PySide2.QtCore.Signal emitted when the gui is started (True) or stopped (False)

MONITOR = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.VTE: 9>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])
Values to create *Display* widgets for in the Sensor Monitor column. See *values.DISPLAY_MONITOR*

CONTROL = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.IE_RATIO: 7>, <pvp.common.values.Value object>), (<ValueName.PIP_TIME: 2>, <pvp.common.values.Value object>)])
Values to create *Display* widgets for in the Control column. See *values.CONTROL*

PLOTS = OrderedDict([(<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])
Values to create *Plot* widgets for. See *values.PLOTS*

monitor_width = 3
Relative width of the sensor monitor column

plot_width = 4
Relative width of the plot column

control_width = 3
Relative width of the control column

total_width = 10
computed from monitor_width+plot_width+control_width

update_gui(*vals: Optional[*pvp.common.message.SensorValues*] = None*)

Parameters **vals** (SensorValue) – Default None, but SensorValues can be passed manually – usually for debugging

init_ui()

0. Create the UI components for the ventilator screen

Call, in order:

- *PVP_Gui.init_ui_status_bar()*
- *PVP_Gui.init_ui_monitor()*
- *PVP_Gui.init_ui_plots()*
- *PVP_Gui.init_ui_controls()*
- *PVP_Gui.init_ui_signals()*

Create and set sizes of major layouts

init_ui_status_bar()

1. Create the widgets.Control_Panel and widgets.Alarm_Bar

and add them to the main layout

init_ui_monitor()

2. Create the left "sensor monitor" column of widgets.Display widgets

And add the logo to the bottom left corner if there's room

init_ui_plots()

3. Create the *Plot_Container*

init_ui_controls()

4. Create the "controls" column of widgets.Display widgets

init_ui_signals()

5. Connect Qt signals and slots between widgets

- Connect controls and sensor monitors to *PVP_Gui.set_value()*
- Connect control panel buttons to their respective methods

set_value(*new_value*, *value_name=None*)

Set a control value using a value and its name.

Constructs a *message.ControlSetting* object to give to *PVP_Gui.set_control()*

Note: This method is primarily intended as a means of responding to signals from other widgets, Other cases should use *set_control()*

Parameters

- **new_value** (*float*) – A new value for some control setting
- **value_name** (*values.ValueName*) – THe ValueName for the control setting. If `None`, assumed to be coming from a *Display* widget that can identify itself with its `objectName`

set_control(*control_object:* pvp.common.message.ControlSetting)
Set a control in the alarm manager, coordinator, and gui

Also update our state with `update_state()`

Parameters **control_object** (*message.ControlSetting*) – A control setting to give to `CoordinatorBase.set_control`

handle_alarm(*alarm:* pvp.alarm.alarm.Alarm)
Receive an *Alarm* from the *Alarm_Manager*

Alarms are both raised and cleared with this method – there is no separate "clear_alarm" method because an alarm of *AlarmSeverity* of `OFF` is cleared.

Give the alarm to the *Alarm_Bar* and update the alarm *Display.alarm_state* of all widgets listed as `Alarm.cause`

Parameters **alarm** (*Alarm*) – The alarm to raise (or clear)

limits_updated(*control:* pvp.common.message.ControlSetting)
Receive updated alarm limits from the *Alarm_Manager*

When a value is set that has an *Alarm_Rule* that *Alarm_Rule.depends* on it, the alarm thresholds will be updated and handled here.

Eg. the high-pressure alarm is set to be 15% above PIP. When PIP is changed, this method will receive a *message.ControlSetting* that tells us that alarm threshold has changed.

Update the *Display* and *Plot* widgets.

If we are setting a new HAPA limit, that is also sent to the controller as it needs to respond as quickly as possible to high-pressure events.

Parameters **control** (*message.ControlSetting*) – A ControlSetting with its `max_value` or

:param `min_value` set:

start()
Click the `start_button`

toggle_start(*state: bool*)
Start or stop ventilation.

Typically called by the `PVP_Gui.control_panel.start_button`.

Raises a dialogue to confirm ventilation start or stop

Starts or stops the controller via the coordinator

If starting, locks controls.

Parameters **state** (*bool*) – If True, start ventilation. If False, stop ventilation.

closeEvent(*event*)
Emit *gui_closing* and close!

Kill the coordinator with *CoordinatorBase.kill()*

toggle_lock(*state*)
Toggle the lock state of the controls

Typically called by `PVP_Gui.control_panel.lock_button`

Parameters `state` –

Returns:

`update_state`(*state_type: str*, *key: str*, *val: Union[str, float, int]*)
Update the GUI state and save it to disk with `Vent_Gui.save_state()`

Currently, just saves the state of control settings.

Parameters

- **state_type** (*str*) – What type of state to save, one of `('controls')`
- **key** (*str*) – Which of that type is being saved (eg. if 'control', 'PIP')
- **val** (*str, float, int*) – What is that item being set to?

Returns:

`save_state`()
Try to save GUI state to `prefs['VENT_DIR"] + prefs['GUI_STATE_FN']`

`load_state`(*state: Union[str, dict]*)
Load GUI state and reconstitute

currently, just *PVP_Gui.set_value()* for all previously saved values

Parameters `state` (*str, dict*) – either a pathname to a state file or an already-loaded state dictionary

`staticMetaObject = <PySide2.QtCore.QMetaObject object>`

`toggle_cycle_widget`(*button*)
Set which breath cycle control is automatically calculated

The timing of a breath cycle can be parameterized with Respiration Rate, Inspiration Time, and Inspiratory/Expiratory ratio, but if two of these parameters are set the third is already known.

This method changes which value has its *Display* widget hidden and is automatically calculated

Parameters `button` (`PySide2.QtWidgets.QAbstractButton`, *values.ValueName*) – The Qt Button that invoked the method or else a ValueName

`set_pressure_units`(*units*)
Select whether pressure units are displayed as "cmH2O" or "hPa"

calls *Display.set_units()* on controls and plots that display pressure

Parameters `units` (*str*) – one of "cmH2O" or "hPa"

`set_breath_detection`(*breath_detection: bool*)
Connected to `breath_detection_button` - toggles autonomous breath detection in the controller

Parameters `breath_detection` (*bool*) – Whether the controller detects autonomous breaths and resets the breath cycle accordingly

`get_breath_detection`() → bool
Get the current state of breath detection from the controller

Returns if True, automatic breath detection is enabled

Return type bool

_set_cycle_control(*value_name: str*, *new_value: float*)
Compute the computed breath cycle control.

We only actually have BPM and INSPt as controls, so if we're using I:E ratio we have to compute one or the other.

Computes the value and calls *set_control()* with the appropriate values:

```
# ie = inspt/expt
# inspt = ie*expt
# expt = inspt/ie
#
# cycle_time = inspt + expt
# cycle_time = inspt + inspt/ie
# cycle_time = inspt * (1+1/ie)
# inspt = cycle_time / (1+1/ie)
```

property controls_set
Check if all controls are set

Note: Note that even when RR or INSPt are autocalculated, they are still set in their control objects, so this check is the same regardless of what is set to autocalculate

---

property update_period
The global delay between redraws of the GUI (seconds)

init_controls()
on startup, set controls in coordinator to ensure init state is synchronized

_screenshot()
Raise each of the alarm severities to check if they work and to take a screenshot

Warning: should never be used except for testing and development!

pvp.gui.main.**launch_gui**(*coordinator*, *set_defaults=False*, *screenshot=False*) → Tuple[PySide2.QtWidgets.QApplication, *pvp.gui.main.PVP_Gui*]
Launch the GUI with its appropriate arguments and doing its special opening routine

To launch the gui, one must:

- Create a `PySide2.QtWidgets.QApplication`
- Set the app style using `gui.styles.DARK_THEME`
- Set the app palette with *gui.styles.set_dark_palette()*
- Call the gui's `show` method

Parameters

- **coordinator** (*coordinator.CoordinatorBase*) – Coordinator used to communicate between GUI and controller
- **set_defaults** (*bool*) – whether default control parameters should be set on startup – only to be used for development or testing
- **screenshot** (*bool*) – whether alarms should be raised to take a screenshot, should never be used on a live system.

Returns The PySide2.QtWidgets.QApplication and *PVP_Gui*

Return type (tuple)

1.1.13.2 GUI Widgets

Control Panel

The Control Panel starts and stops ventilation and controls runtime options

Classes:

Control_Panel()	The control panel starts and stops ventilation and controls runtime settings
Start_Button(*args, **kwargs)	Button to start and stop Ventilation, created by *Control_Panel*
Lock_Button(*args, **kwargs)	Button to lock and unlock controls
HeartBeat([update_interval, timeout_dur])	Track state of connection with Controller
StopWatch([update_interval])	Simple widget to display ventilation time!

class pvp.gui.widgets.control_panel.**Control_Panel**

The control panel starts and stops ventilation and controls runtime settings

It creates:

- Start/stop button
- **Status indicator - a clock that increments with heartbeats,** or some other visual indicator that things are alright
- Version indicator
- Buttons to select options like cycle autoset and automatic breath detection

Args:

start_button

Button to start and stop ventilation

Type *Start_Button*

lock_button

Button used to lock controls

Type *Lock_Button*

heartbeat

Widget to keep track of communication with controller

Type *HeartBeat*

runtime

Widget used to display time since start of ventilation

Type *StopWatch*

Attributes:

pressure_units_changed(*args, **kwargs)	Signal emitted when pressure units have been changed.
cycle_autoset_changed(*args, **kwargs)	Signal emitted when a different breath cycle control value is set to be autocalculated
staticMetaObject	

Methods:

init_ui()	Initialize all graphical elements and buttons!
_pressure_units_changed(button)	Emit the str of the current pressure units

pressure_units_changed(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
Signal emitted when pressure units have been changed.

Contains str of current pressure units

cycle_autoset_changed(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
Signal emitted when a different breath cycle control value is set to be autocalculated

init_ui()
Initialize all graphical elements and buttons!

_pressure_units_changed(*button*)
Emit the str of the current pressure units

Parameters **button** (PySide2.QtWidgets.QPushButton) – Button that was clicked

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.control_panel.**Start_Button**(**args*, ***kwargs*)
Button to start and stop Ventilation, created by *Control_Panel*

pixmaps
Dictionary containing pixmaps used to draw start/stop state

Type dict

Attributes:

states	Possible states of Start_Button
staticMetaObject	

Methods:

load_pixmaps()	Load pixmaps to *Start_Button.pixmaps*
set_state(state)	Set state of button

states = ['OFF', 'ON', 'ALARM']
Possible states of Start_Button

load_pixmaps()
Load pixmaps to *Start_Button.pixmaps*

set_state(*state*)
Set state of button

Should only be called by other objects (as there are checks to whether it's ok to start/stop that we shouldn't be aware of)

Parameters **state** (*str*) – one of `('OFF', 'ON', 'ALARM')`

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.control_panel.**Lock_Button**(**args*, ***kwargs*)

Button to lock and unlock controls

Created by *Control_Panel*

pixmaps

Dictionary containing pixmaps used to draw locked/unlocked state

Type dict

Attributes:

`states`	Possible states of Lock Button
`staticMetaObject`	

Methods:

`load_pixmaps`()	Load pixmaps used to display lock state to *Lock_Button.pixmaps*
`set_state`(state)	Set lock state of button

states = ['DISABLED', 'UNLOCKED', 'LOCKED']

Possible states of Lock Button

load_pixmaps()

Load pixmaps used to display lock state to *Lock_Button.pixmaps*

set_state(*state*)

Set lock state of button

Should only be called by other objects (as there are checks to whether it's ok to start/stop that we shouldn't be aware of)

Parameters **state** (*str*) – `('OFF', 'ON', 'ALARM')`

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.control_panel.**HeartBeat**(*update_interval: int = 100*, *timeout_dur: int = 5000*)

Track state of connection with Controller

Check when we last had contact with controller every *HeartBeat.update_interval* ms, if longer than `HeartBeat.timeout_dur` then emit a timeout signal

Parameters

- **update_interval** (*int*) – How often to do the heartbeat, in ms
- **timeout** (*int*) – how long to wait before hearing from control process, in ms

_state

whether the system is running or not

Type bool

_last_heartbeat
Timestamp of last contact with controller

Type float

start_time
Time that ventilation was started

Type float

timer
Timer that checks for last contact

Type PySide2.QtCore.QTimer

update_interval
How often to do the heartbeat, in ms

Type int

timeout
how long to wait before hearing from control process, in ms

Type int

Attributes:

timeout(*args, **kwargs)	Signal that a timeout has occurred -- too long between contact with controller.
heartbeat(*args, **kwargs)	Signal that requests to affirm contact with controller if no message has been received in timeout duration
staticMetaObject	

Methods:

init_ui()	Initialize labels and status indicator
set_state(state)	Set running state
set_indicator([state])	Set visual indicator
start_timer([update_interval])	Start *HeartBeat.timer* to check for contact with controller
stop_timer()	Stop timer and clear text
beatheart(heartbeat_time)	Slot that receives timestamps of last contact with controller
_heartbeat()	Called every (update_interval) milliseconds to set the check the status of the heartbeat.

timeout(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
Signal that a timeout has occurred – too long between contact with controller.

heartbeat(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
Signal that requests to affirm contact with controller if no message has been received in timeout duration

init_ui()
Initialize labels and status indicator

set_state(*state*)
Set running state

if just starting reset *HeartBeat._last_heartbeat*

Parameters **state** (*bool*) – Whether we are starting (True) or stopping (False)

set_indicator(*state=None*)
Set visual indicator

Parameters **state** (*'ALARM', 'OFF', 'NORMAL'*) – Current state of connection with controller

start_timer(*update_interval=None*)
Start *HeartBeat.timer* to check for contact with controller

Parameters **update_interval** (*int*) – How often (in ms) the timer should be updated. if None, use `self.update_interval`

stop_timer()
Stop timer and clear text

beatheart(*heartbeat_time*)
Slot that receives timestamps of last contact with controller

Parameters **heartbeat_time** (*float*) – timestamp of last contact with controller

_heartbeat()
Called every (update_interval) milliseconds to set the check the status of the heartbeat.

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.control_panel.**StopWatch**(*update_interval: float = 100, *args, **kwargs*)
Simple widget to display ventilation time!

Parameters

- **update_interval** (*float*) – update clock every n seconds
- ***args** – passed to `PySide2.QtWidgets.QLabel`
- ****kwargs** – passed to `PySide2.QtWidgets.QLabel`

Methods:

`__init__`([update_interval])	Simple widget to display ventilation time!
`init_ui`()	
`start_timer`([update_interval])	**param update_interval** How often (in ms) the timer should be updated.
`stop_timer`()	Stop timer and reset label
`_update_time`()	

Attributes:

`staticMetaObject`

__init__(*update_interval: float = 100, *args, **kwargs*)
Simple widget to display ventilation time!

Parameters

- **update_interval** (*float*) – update clock every n seconds
- ***args** – passed to PySide2.QtWidgets.QLabel
- ****kwargs** – passed to PySide2.QtWidgets.QLabel

staticMetaObject = <PySide2.QtCore.QMetaObject object>

init_ui()

start_timer(*update_interval=None*)

Parameters **update_interval** (*float*) – How often (in ms) the timer should be updated.

stop_timer()
Stop timer and reset label

_update_time()

Alarm Bar

The *Alarm_Bar* displays *Alarm* status with *Alarm_Card* widgets and plays alarm sounds with the *Alarm_Sound_Player*

Classes:

Alarm_Bar()	Holds and manages a collection of *Alarm_Card* s and communicates requests for dismissal to the *Alarm_Manager*
Alarm_Card(alarm)	Representation of an alarm raised by *Alarm_Manager* in GUI.
Alarm_Sound_Player([increment_delay])	Plays alarm sounds to reflect current alarm severity and active duration with PySide2.QtMultimedia.QSoundEffect objects

class pvp.gui.widgets.alarm_bar.**Alarm_Bar**
Holds and manages a collection of *Alarm_Card* s and communicates requests for dismissal to the *Alarm_Manager*

The alarm bar also manages the *Alarm_Sound_Player*

alarms
A list of active alarms

Type *List*[*Alarm*]

alarm_cards
A list of active alarm cards

Type *List*[*Alarm_Card*]

sound_player
Class that plays alarm sounds!

Type *Alarm_Sound_Player*

icons
Dictionary of pixmaps with icons for different alarm levels

Type dict

Methods:

make_icons()	Create pixmaps from standard icons to display for different alarm types
init_ui()	Initialize the UI
add_alarm(alarm)	Add an alarm created by the *Alarm_Manager* to the bar.
clear_alarm([alarm, alarm_type])	Remove an alarm card, update appearance and sound player to reflect current max severity
update_icon()	Call *set_icon()* with highest severity in *Alarm_Bar.alarms*
set_icon([state])	Change the icon and bar appearance to reflect the alarm severity

Attributes:

alarm_level	Current maximum *AlarmSeverity*
staticMetaObject	

make_icons()
Create pixmaps from standard icons to display for different alarm types

Store in *Alarm_Bar.icons*

init_ui()
Initialize the UI

- Create layout
- Set icon
- Create mute button

add_alarm(*alarm:* pvp.alarm.alarm.Alarm)
Add an alarm created by the *Alarm_Manager* to the bar.

If an alarm alreaady exists with that same *AlarmType* , *Alarm_Bar.clear_alarm()*

Insert new alarm in order the prioritizes alarm severity with highest severity on right

Set alarm sound and begin playing if not already.

Parameters **alarm** (*Alarm*) – Alarm to be added

clear_alarm(*alarm: Optional[*pvp.alarm.alarm.Alarm*] = None*, *alarm_type: Optional[*pvp.alarm.AlarmType*] = None*)
Remove an alarm card, update appearance and sound player to reflect current max severity

Must pass one of either `alarm` or `alarm_type`

Parameters

- **alarm** (*Alarm*) – Alarm to be cleared
- **alarm_type** (*AlarmType*) – Alarm type to be cleared

update_icon()
Call *set_icon()* with highest severity in *Alarm_Bar.alarms*

set_icon(*state: Optional[*pvp.alarm.AlarmSeverity*] = None*)
Change the icon and bar appearance to reflect the alarm severity

Parameters **state** (*AlarmSeverity*) – Alarm Severity to display, if None change to default display

property alarm_level
Current maximum *AlarmSeverity*

Returns *AlarmSeverity*

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.alarm_bar.**Alarm_Card**(*alarm:* pvp.alarm.alarm.Alarm)
Representation of an alarm raised by *Alarm_Manager* in GUI.

If allowed by alarm (by `latch` setting) , allows user to dismiss/silence alarm.

Otherwise request to dismiss is logged by *Alarm_Manager* and the card is dismissed when the conditions that generated the alarm are no longer met.

Parameters **alarm** (*Alarm*) – Alarm to represent

alarm
The represented alarm

Type *Alarm*

severity
The severity of the represented alarm

Type *AlarmSeverity*

close_button
Button that requests an alarm be dismissed

Type PySide2.QtWidgets.QPushButton

Methods:

init_ui()	Initialize graphical elements
_dismiss()	Gets the *Alarm_Manager* instance and calls *Alarm_Manager.dismiss_alarm()*

Attributes:

staticMetaObject

init_ui()
Initialize graphical elements

- Create labels
- Set stylesheets
- Create and connect dismiss button

Returns:

_dismiss()
Gets the *Alarm_Manager* instance and calls *Alarm_Manager.dismiss_alarm()*

Also change appearance of close button to reflect requested dismissal

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.alarm_bar.**Alarm_Sound_Player**(*increment_delay: int = 10000*, **args*, ***kwargs*)

Plays alarm sounds to reflect current alarm severity and active duration with `PySide2.QtMultimedia.QSoundEffect` objects

Alarm sounds indicate severity with the number and pitch of tones in a repeating tone cluster (eg. low severity sounds have a single repeating tone, but high-severity alarms have three repeating tones)

They indicate active duration by incrementally removing a low-pass filter and making tones have a sharper attack and decay.

When an alarm of any severity is started the `<severity_0.wav` file begins playing, and a timer is started to call *Alarm_Sound_Player.increment_level()*

Parameters

- **increment_delay** (*int*) – Delay between calling *Alarm_Sound_Player.increment_level()*
- ****kwargs** (**args,*) – passed to `PySide2.QtWidgets.QWidget`

idx

Dictionary of dictionaries allowing sounds to be accessed like `self.idx[AlarmSeverity][level]`

Type dict

files

list of sound file paths

Type list

increment_delay

Time between calling `Alarm_Sound_Player.increment_level`()` in ms

Type int

playing

Whether or not a sound is playing

Type bool

_increment_timer

Timer that increments alarm sound level

Type `PySide2.QtCore.QTimer`

_changing_track

used to ensure single sound changing call happens at a time.

Type `threading.Lock`

Attributes:

severity_map	mapping between string representations of severities used by filenames and *AlarmSeverity*
staticMetaObject	

Methods:

init_audio()	Load audio files in pvp/external/audio and add to *Alarm_Sound_Player.idx*
play()	Start sound playback
stop()	Stop sound playback
set_sound([severity, level])	Set sound to be played
increment_level()	If current level is below the maximum level, increment with *Alarm_Sound_Player.set_sound()* Returns:
set_mute(mute)	Set mute state

severity_map = {'high': AlarmSeverity.HIGH, 'low': AlarmSeverity.LOW, 'medium': AlarmSeverity.MEDIUM}
mapping between string representations of severities used by filenames and *AlarmSeverity*

init_audio()
Load audio files in pvp/external/audio and add to *Alarm_Sound_Player.idx*

play()
Start sound playback

Play sound listed as Alarm_Sound_Player._current_sound

Note: *Alarm_Sound_Player.set_sound()* must be called first.

stop()
Stop sound playback

set_sound(*severity: Optional[*pvp.alarm.AlarmSeverity*] = None*, *level: Optional[int] = None*)
Set sound to be played

At least an *AlarmSeverity* must be provided.

Parameters

- **severity** (*AlarmSeverity*) – Severity of alarm sound to play
- **level** (*int*) – level (corresponding to active duration) of sound to play

increment_level()
If current level is below the maximum level, increment with *Alarm_Sound_Player.set_sound()* Returns:

staticMetaObject = <PySide2.QtCore.QMetaObject object>

set_mute(*mute: bool*)
Set mute state

Parameters **mute** (*bool*) – if True, mute. if False, unmute.

Display

Unified monitor & control widget

Displays sensor values, and can optionally control system settings.

The *PVP_Gui* instantiates display widgets according to the contents of *values.DISPLAY_CONTROL* and *values.DISPLAY_MONITOR*

Classes:

Display(value[, update_period, enum_name, ...])	Unified widget for display of sensor values and control of ventilation parameters
Limits_Plot([style])	Widget to display current value in a bar graph along with alarm limits

class pvp.gui.widgets.display.**Display**(*value:* pvp.common.values.Value, *update_period: float = 0.5*, *enum_name: Optional[*pvp.common.values.ValueName*] = None*, *button_orientation: str = 'left'*, *control_type: Union[None, str] = None*, *style: str = 'dark'*, **args*, ***kwargs*)

Unified widget for display of sensor values and control of ventilation parameters

Displayed values are updated according to `Dispaly.timed_update()`

Parameters

- **value** (*Value*) – Value object to represent
- **update_period** (*float*) – Amount of time between updates of the textual display of values
- **enum_name** (*ValueName*) – Value name of object to represent
- **button_orientation** (*'left', 'right'*) – whether the controls are drawn on the `'left'` or `'right'`
- **control_type** (*None, 'slider', 'record'*) – type of control - either `None` (no control), `slider` (a slider can be opened to set a value), or `record` where recent sensor values are averaged and used to set the control value. Both types of control allow values to be input from the keyboard by clicking on the editable label
- **style** (*'light', 'dark'*) – whether the widget is `'dark'` (light text on dark background) or `'light'` (dark text on light background
- ****kwargs** (**args,*) – passed on to `PySide2.QtWidgets.QWidget`

`self.`**name**
Unpacked from `value`

`self.`**units**
Unpacked from `value`

`self.`**abs_range**
Unpacked from `value`

`self.`**safe_range**
Unpacked from `value`

`self.`**alarm_range**
initialized from `value`, but updated by alarm manager

`self.`**decimals**
Unpacked from `value`

`self.update_period`
Amount of time between updates of the textual display of values

Type float

`self.enum_name`
Value name of object to represent

Type *ValueName*

`self.orientation`
whether the controls are drawn on the `'left'` or `'right'`

Type 'left', 'right'

`self.control`
type of control - either `None` (no control), `slider` (a slider can be opened to set a value), or `record` where recent sensor values are averaged and used to set the control value.

Type None, 'slider', 'record'

`self._style`
whether the widget is `'dark'` (light text on dark background) or `'light'` (dark text on light background)

Type 'light', 'dark'

`self.set_value`
current set value of controlled value, if any

Type float

`self.sensor_value`
current value of displayed sensor value, if any.

Type float

Attributes:

value_changed(*args, **kwargs)	Signal emitted when controlled value of display object has changed.
is_set	Check if value has been set for this control.
alarm_state	Current visual display of alarm severity
staticMetaObject	

Methods:

init_ui()	UI is initialized in several stages
init_ui_labels()	
init_ui_toggle_button()	
init_ui_limits()	Create widgets to display sensed value alongside set value
init_ui_slider()	
init_ui_record()	
init_ui_layout()	Basically two methods.

continues on next page

Table 35 – continued from previous page

init_ui_signals()	
toggle_control(state)	Toggle the appearance of the slider, if a slider
toggle_record(state)	Toggle the recording state, if a recording control
_value_changed(new_value)	"outward-directed" method to emit new changed control value when changed by this widget
update_set_value(new_value)	Update to reflect new control value set from elsewhere (inwardly directed setter)
update_sensor_value(new_value)	Receive new sensor value and update display widgets
update_limits(control)	Update the alarm range and the GUI elements corresponding to it
redraw()	Redraw all graphical elements to ensure internal model matches view
timed_update()	Refresh textual sensor values only periodically to prevent them from being totally unreadable from being changed too fast.
set_units(units)	Set pressure units to display as cmH2O or hPa.
set_locked(locked)	Set locked status of control

value_changed(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>
Signal emitted when controlled value of display object has changed.

Contains new value (float)

init_ui()
UI is initialized in several stages

- 0: this method, get stylesheets based on `self._style` and call remaining initialization methods
- 1: *Display.init_ui_labels()* - create generic labels shared by all display objects
- 2: *Display.init_ui_toggle_button()* - create the toggle or record button used by controls
- 3: *Display.init_ui_limits()* - create a plot that displays the sensor value graphically relative to the alarm limits
- 4: *Display.init_ui_slider()* or `Display.ini_ui_record()` - depending on what type of control this is
- 5: *Display.init_ui_layout()* since the results of the previous steps varies, do all layout at the end depending on orientation
- 6: *Display.init_ui_signals()* connect slots and signals

init_ui_labels()

init_ui_toggle_button()

init_ui_limits()
Create widgets to display sensed value alongside set value

init_ui_slider()

init_ui_record()

init_ui_layout()
Basically two methods… lay objects out depending on whether we're oriented with our button to the left or right

init_ui_signals()

toggle_control(*state*)
Toggle the appearance of the slider, if a slider

Parameters **state** (*bool*) – Whether to show or hide the slider

toggle_record(*state*)
Toggle the recording state, if a recording control

Parameters **state** (*bool*) – Whether recording should be started or stopped. when started, start storing new sensor values in a list. when stopped, average thgem and emit new value.

_value_changed(*new_value: float*)
"outward-directed" method to emit new changed control value when changed by this widget

Pop a confirmation dialog if values are set outside the safe range.

Parameters

- **new_value** (*float*) – new value!
- **emit** (*bool*) – whether to emit the *value_changed* signal (default True) – in the case that our value is being changed by someone other than us

update_set_value(*new_value: float*)
Update to reflect new control value set from elsewhere (inwardly directed setter)

Parameters **new_value** (*float*) – new value to set!

update_sensor_value(*new_value: float*)
Receive new sensor value and update display widgets

Parameters **new_value** (*float*) – new sensor value!

update_limits(*control:* pvp.common.message.ControlSetting)
Update the alarm range and the GUI elements corresponding to it

Parameters **control** (*ControlSetting*) – control setting with min_value or max_value

redraw()
Redraw all graphical elements to ensure internal model matches view

Typically used when changing units

timed_update()
Refresh textual sensor values only periodically to prevent them from being totally unreadable from being changed too fast.

set_units(*units: str*)
Set pressure units to display as cmH2O or hPa.

Uses functions from *pvp.common.unit_conversion* such that

- `self._convert_in` converts internal, canonical units to displayed units (eg. `cmH2O` is used by all other modules, so we convert it to `hPa`
- `self._convert_out` converts displayed units to send to other parts of the system

---

Note: currently unit conversion is only supported for Pressure.

Parameters **units** (*'cmH2O', 'hPa'*) – new units to display

set_locked(*locked: bool*)
Set locked status of control

Parameters **locked** (*bool*) – If True, disable all controlling widgets, if False, re-enable.

property is_set
Check if value has been set for this control.

Used to check if all settings have been set preflight by *PVP_Gui*

Returns whether we have an `Display.set_value`

Return type bool

property alarm_state: *pvp.alarm.AlarmSeverity*
Current visual display of alarm severity

Change sensor value color to reflect the alarm state of that set parameter –

eg. if we have a HAPA alarm, set the PIP control to display as red.

Returns *AlarmSeverity*

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.display.**Limits_Plot**(*style: str = 'light', *args, **kwargs*)
Widget to display current value in a bar graph along with alarm limits

Parameters **style** (*'light', 'dark'*) – Whether we are being displayed in a light or dark styled *Display* widget

set_value
Set value of control, displayed as horizontal black line always set at center of bar

Type float

sensor_value
Sensor value to compare against control, displayed as bar

Type float

When initializing PlotWidget, *parent* and *background* are passed to `GraphicsWidget.__init__()` and all others are passed to `PlotItem.__init__()`.

Attributes:

staticMetaObject

Methods:

init_ui()	Create bar chart and horizontal lines to reflect
update_value([min, max, sensor_value, set_value])	Move the lines! Pass any of the represented values.
update_yrange()	Set yrange to ensure that the set value is always in the center of the plot and that all the values are in range.

staticMetaObject = <PySide2.QtCore.QMetaObject object>

init_ui()
Create bar chart and horizontal lines to reflect

- Sensor Value
- Set Value

- High alarm limit
- Low alarm limit

update_value(*min: Optional[float] = None*, *max: Optional[float] = None*, *sensor_value: Optional[float] = None*, *set_value: Optional[float] = None*)

Move the lines! Pass any of the represented values.

Also updates yrange to ensure that the control value is always centered in the plot

Parameters

- **min** (*float*) – new alarm minimum
- **max** (*float*) – new alarm _maximum
- **sensor_value** (*float*) – new value for the sensor bar plot
- **set_value** (*float*) – new value for the set value line

update_yrange()

Set yrange to ensure that the set value is always in the center of the plot and that all the values are in range.

Plot

Widgets to plot waveforms of sensor values

The *PVP_Gui* creates a *Plot_Container* that allows the user to select

- which plots (of those in `values.PLOT`) are displayed
- the timescale (x range) of the displayed waveforms

Plots display alarm limits as red horizontal bars

Data:

PLOT_TIMER	A `QTimer` that updates :class:`.TimedPlotCurveItem`s
PLOT_FREQ	Update frequency of *Plot* s in Hz

Classes:

Plot(name[, buffer_size, plot_duration, ...])	Waveform plot of single sensor value.
Plot_Container(plot_descriptors, *args, **kwargs)	Container for multiple :class;`.Plot` objects

pvp.gui.widgets.plot.**PLOT_TIMER = None**

A `QTimer` that updates :class:`.TimedPlotCurveItem`s

pvp.gui.widgets.plot.**PLOT_FREQ = 5**

Update frequency of *Plot* s in Hz

class pvp.gui.widgets.plot.**Plot**(*name: str*, *buffer_size: int = 4092*, *plot_duration: float = 10*, *plot_limits: Optional[tuple] = None*, *color=None*, *units=''*, ***kwargs*)

Waveform plot of single sensor value.

Plots values continuously, wrapping at zero rather than shifting x axis.

Parameters

- **name** (*str*) – String version of *ValueName* used to set title
- **buffer_size** (*int*) – number of samples to store

- **plot_duration** (*float*) – default x-axis range
- **plot_limits** (*tuple*) – tuple of (ValueName)s for which to make pairs of min and max range lines
- **color** (*None, str*) – color of lines
- **units** (*str*) – Unit label to display along title
- ****kwargs** –

timestamps
deque of timestamps

Type collections.deque

history
deque of sensor values

Type collections.deque

cycles
deque of breath cycles

Type collections.deque

When initializing PlotWidget, *parent* and *background* are passed to `GraphicsWidget.__init__()` and all others are passed to `PlotItem.__init__()`.

Attributes:

limits_changed(*args, **kwargs)	
staticMetaObject	

Methods:

set_duration(dur)	Set duration, or span of x axis.
update_value(new_value)	Update with new sensor value
set_safe_limits(limits)	Set the position of the max and min lines for a given value
reset_start_time()	Reset start time -- return the scrolling time bar to position 0
set_units(units)	Set displayed units

limits_changed(**args*, ***kwargs*) **= <PySide2.QtCore.Signal object>**

set_duration(*dur: float*)
Set duration, or span of x axis.

Parameters **dur** (*float*) – span of x axis (in seconds)

update_value(*new_value: tuple*)
Update with new sensor value

Parameters **new_value** (*tuple*) – (timestamp from time.time(), breath_cycle, value)

set_safe_limits(*limits:* pvp.common.message.ControlSetting)
Set the position of the max and min lines for a given value

Parameters **limits** (*ControlSetting*) – Controlsetting that has either a `min_value` or `max_value` defined

reset_start_time()
Reset start time – return the scrolling time bar to position 0

set_units(*units*)
Set displayed units

Currently only implemented for Pressure, display either in cmH2O or hPa

Parameters **units** (*'cmH2O', 'hPa'*) – unit to display pressure as

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.plot.**Plot_Container**(*plot_descriptors: Dict[*pvp.common.values.ValueName, pvp.common.values.Value*]*, **args*, ***kwargs*)
Container for multiple :class;`.Plot` objects

Allows user to show/hide different plots and adjust x-span (time zoom)

Note: Currently, the only unfortunately hardcoded parameter in the whole GUI is the instruction to initially hide FIO2, there should be an additional parameter in `Value` that says whether a plot should initialize as hidden or not

---

Todo: Currently, colors are set to alternate between orange and light blue on initialization, but they don't update when plots are shown/hidden, so the alternating can be lost and colors can look random depending on what's selcted.

Parameters **plot_descriptors** (*Dict[ValueName, Value]*) – dict of *Value* items to plot

plots
Dict mapping *ValueName* s to *Plot* s

Type dict

slider
slider used to set x span

Type `PySide2.QtWidgets.QSlider`

Methods:

init_ui()	

update_value(vals)	Try to update all plots who have new sensorvalues
toggle_plot(state)	Toggle the visibility of a plot.
set_safe_limits(control)	Try to set horizontal alarm limits on all relevant plots
set_duration(duration)	Set the current duration (span of the x axis) of all plots
reset_start_time()	Call *Plot.reset_start_time()* on all plots
set_units(units)	Call *Plot.set_units()* for all contained plots
set_plot_mode()	

Attributes:

staticMetaObject

init_ui()

update_value(*vals:* pvp.common.message.SensorValues)
Try to update all plots who have new sensorvalues

Parameters **vals** (*SensorValues*) – Sensor Values to update plots with

toggle_plot(*state: bool*)
Toggle the visibility of a plot.

get the name of the plot from the `sender`'s `objectName`

Parameters **state** (*bool*) – Whether the plot should be visible (True) or not (False)

set_safe_limits(*control:* pvp.common.message.ControlSetting)
Try to set horizontal alarm limits on all relevant plots

Parameters **control** (*ControlSetting*) – with either `min_value` or `max_value` set

Returns:

set_duration(*duration: float*)
Set the current duration (span of the x axis) of all plots

Also make sure that the text box and slider reflect this duration

Parameters **duration** (*float*) – new duration to set

Returns:

reset_start_time()
Call *Plot.reset_start_time()* on all plots

set_units(*units: str*)
Call *Plot.set_units()* for all contained plots

staticMetaObject = <PySide2.QtCore.QMetaObject object>

set_plot_mode()

Todo: switch between longitudinal timeseries and overlaid by breath cycle!!!

Components

Very basic components used by other widgets.

These are relatively sparsely documented because their operation is mostly self-explanatory

Classes:

DoubleSlider([decimals])	Slider capable of representing floats
KeyPressHandler	Custom key press handler https://gist.github.com/mfessenden/baa2b87b8addb0b60e54a11c1da48046
EditableLabel([parent])	Editable label https://gist.github.com/mfessenden/baa2b87b8addb0b60e54a11c1da48046

continues on next page

Table 44 – continued from previous page

QHLine([parent, color])	with respct to https://stackoverflow.com/a/51057516
QVLine([parent, color])	
OnOffButton([state_labels, toggled])	Simple extension of toggle button with styling for clearer 'ON' vs 'OFF'

class pvp.gui.widgets.components.**DoubleSlider**(*decimals=1*, **args*, ***kargs*)

Slider capable of representing floats

Ripped off from and https://stackoverflow.com/a/50300848 ,

Thank you!!!

Attributes:

doubleValueChanged(*args, **kwargs)
staticMetaObject

Methods:

setDecimals(decimals)
emitDoubleValueChanged()
value()
setMinimum(value)
setMaximum(value)
minimum()
_minimum()
maximum()
_maximum()
setSingleStep(value)
singleStep()
_singleStep()
setValue(value)

doubleValueChanged(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>

setDecimals(*decimals*)

emitDoubleValueChanged()

value()

setMinimum(*value*)

setMaximum(*value*)

minimum()

_minimum()

maximum()

_maximum()

setSingleStep(*value*)

singleStep()

_singleStep()

setValue(*value*)

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.components.**KeyPressHandler**

Custom key press handler https://gist.github.com/mfessenden/baa2b87b8addb0b60e54a11c1da48046

Attributes:

escapePressed(*args, **kwargs)

returnPressed(*args, **kwargs)

staticMetaObject

Methods:

eventFilter(obj, event)

escapePressed(**args*, ***kwargs*) **= <PySide2.QtCore.Signal object>**

returnPressed(**args*, ***kwargs*) **= <PySide2.QtCore.Signal object>**

eventFilter(*obj*, *event*)

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.components.**EditableLabel**(*parent=None*, ***kwargs*)

Editable label https://gist.github.com/mfessenden/baa2b87b8addb0b60e54a11c1da48046

Attributes:

textChanged(*args, **kwargs)

staticMetaObject

Methods:

create_signals()	

text()	Standard QLabel text getter
setText(text)	Standard QLabel text setter
labelPressedEvent(event)	Set editable if the left mouse button is clicked
setLabelEditableAction()	Action to make the widget editable
setEditable(editable)	

labelUpdatedAction()	Indicates the widget text has been updated
returnPressedAction()	Return/enter event handler
escapePressedAction()	Escape event handler

textChanged(**args*, ***kwargs*) = <PySide2.QtCore.Signal object>

create_signals()

text()
Standard QLabel text getter

setText(*text*)
Standard QLabel text setter

labelPressedEvent(*event*)
Set editable if the left mouse button is clicked

setLabelEditableAction()
Action to make the widget editable

setEditable(*editable: bool*)

labelUpdatedAction()
Indicates the widget text has been updated

returnPressedAction()
Return/enter event handler

escapePressedAction()
Escape event handler

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.components.**QHLine**(*parent=None*, *color='#FFFFFF'*)
with respct to https://stackoverflow.com/a/51057516

Methods:

setColor(color)	

Attributes:

staticMetaObject	

setColor(*color*)

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.components.**QVLine**(*parent=None*, *color='#FFFFFF'*)

Methods:

setColor(color)	

Attributes:

staticMetaObject	

setColor(*color*)

staticMetaObject = <PySide2.QtCore.QMetaObject object>

class pvp.gui.widgets.components.**OnOffButton**(*state_labels: Tuple[str, str] = ('ON', 'OFF')*, *toggled: bool = False*, **args*, ***kwargs*)

Simple extension of toggle button with styling for clearer 'ON' vs 'OFF'

Parameters

- **state_labels** (*tuple*) – tuple of strings to set when toggled and untoggled
- **toggled** (*bool*) – initialize the button as toggled
- ***args** – passed to QPushButton
- ****kwargs** – passed to QPushButton

Methods:

__init__([state_labels, toggled])	**param state_labels** tuple of strings to set when toggled and untoggled
set_state(state)	

Attributes:

staticMetaObject	

__init__(*state_labels: Tuple[str, str] = ('ON', 'OFF')*, *toggled: bool = False*, **args*, ***kwargs*)

Parameters

- **state_labels** (*tuple*) – tuple of strings to set when toggled and untoggled
- **toggled** (*bool*) – initialize the button as toggled
- ***args** – passed to QPushButton
- ****kwargs** – passed to QPushButton

set_state(*state: bool*)

staticMetaObject = <PySide2.QtCore.QMetaObject object>

Dialog

Function to display a dialog to the user and receive feedback!

Functions:

pop_dialog(message[, sub_message, modality, ...])	Creates a dialog box to display a message.

pvp.gui.widgets.dialog.**pop_dialog**(*message: str*, *sub_message: typing.Optional[str] = None*, *modality: <class 'PySide2.QtCore.Qt.WindowModality'> = PySide2.QtCore.Qt.WindowModality.NonModal*, *buttons: <class 'PySide2.QtWidgets.QMessageBox.StandardButton'> = PySide2.QtWidgets.QMessageBox.StandardButton.Ok*, *default_button: <class 'PySide2.QtWidgets.QMessageBox.StandardButton'> = PySide2.QtWidgets.QMessageBox.StandardButton.Ok*)

Creates a dialog box to display a message.

Note: This function does *not* call *.exec_* on the dialog so that it can be managed by the caller.

Parameters

- **message** (*str*) – Message to be displayed
- **sub_message** (*str*) – Smaller message displayed below main message (InformativeText)
- **modality** (*QtCore.Qt.WindowModality*) – Modality of dialog box - QtCore.Qt.NonModal (default) is unblocking, QtCore.Qt.WindowModal is blocking
- **buttons** (*QtWidgets.QMessageBox.StandardButton*) – Buttons for the window, can be | ed together
- **default_button** (*QtWidgets.QMessageBox.StandardButton*) – one of `buttons` , the highlighted button

Returns QtWidgets.QMessageBox

1.1.13.3 GUI Stylesheets

Data:

MONITOR_UPDATE_INTERVAL	inter-update interval (seconds) for `Monitor`

Functions:

set_dark_palette(app)	Apply Dark Theme to the Qt application instance.

pvp.gui.styles.**MONITOR_UPDATE_INTERVAL = 0.5**

inter-update interval (seconds) for `Monitor`

Type (float)

pvp.gui.styles.**set_dark_palette**(*app*)

Apply Dark Theme to the Qt application instance.

borrowed from https://github.com/gmarull/qtmodern/blob/master/qtmodern/styles.py

Args: app (QApplication): QApplication instance.

The GUI is written using PySide2 and consists of one main *PVP_Gui* object that instantiates a series of *GUI Widgets*. The GUI is responsible for setting ventilation control parameters and sending them to the `controller` (see *set_control()*), as well as receiving and displaying sensor values (*get_sensors()*).

The GUI also feeds the *Alarm_Manager* *SensorValues* objects so that it can compute alarm state. The *Alarm_Manager* reciprocally updates the GUI with *Alarm* s (*PVP_Gui.handle_alarm()*) and Alarm limits (*PVP_Gui.limits_updated()*).

The main **polling loop** of the GUI is *PVP_Gui.update_gui()* which queries the controller for new *SensorValues* and distributes them to all listening widgets (see method documentation for more details). The rest of the GUI is event driven, usually with Qt Signals and Slots.

The GUI is **configured** by the *values* module, in particular it creates

- *Display* widgets in the left "sensor monitor" box from all *Value* s in *DISPLAY_MONITOR* ,
- *Display* widgets in the right "control" box from all *Value* s in *DISPLAY_CONTROL* , and
- *Plot* widgets in the center plot box from all *Value* s in PLOT

The GUI is not intended to be launched alone, as it needs an active `coordinator` to communicate with the controller process and a few prelaunch preparations (*launch_gui()*). PVP should be started like:

```
python3 -m pvp.main
```

1.1.13.4 Module Overview

1.1.13.5 Screenshot



1.1.14 Controller

1.1.14.1 Purpose of the Controller



Shown above is a typical respiratory waveform (without averaging) as produced with PVP1. Blue is the recorded pressure, orange is the flow out of the system. Note that airflow (and also oxygen concentration) are only measured during expiration, so that the main control-loop during inspiration runs as fast as possible, and is not slowed down by communication delays. Pressure is recorded continuously. Empirically, the Raspberry pi allowed for the primary control loop to run at speeds of ~5ms per loop, which was considerably faster than all hardware delays (i.e. the time it takes for a mechanical, physical valve to open or close; see main manuscript).

The purpose of the controller is to produce a breath waveform, as the one shown above. More specifically, it's job is to reach a certain target-pressure (PIP), and to hold that pressure for a certain amount of time. These numbers are provided by the user via thee UI. Exhalation is passive, and PEEP pressure is mechanically controlled with a spring-valve.

Conceptually, the controller is written as a hybrid system of state and PID control. During inspiration, it actively controls pressure with a simple PID controller. That means that during inspiration, it measures the deviation of the pressure-is-vale from the pressure-target-value, and depending an that distance (and its integral and derivative), it adjusts the opening of the inspiratory valve. Expiration was then instantiated by closing the inspiratory, and opening the expiratory valve to passively release PIP pressure as fast as possible. After reaching PEEP, the controller opens the inspiratory valve slightly to sustain a small flow during PEEP, using the aforementioned manually operated PEEP-valve. We found, empirically, that a combination of proportional and integral term worked best across different physical lung settings.

The controller was also built to allow the user to adjust flow through the system. This is done by a linear correction of the proportional-term. With this adjustment, the user can manipulate the rise-time of the pressure waveform.

In addition to this core function, the controller module continuously monitors for autonomous breaths, high airway pressure, and general system status. Autonomous breathing was detected by transient pressure drops below PEEP. A detected breath triggered a new breath cycle. High airway pressure is defined as exceeding a certain pressure for a certain time (as to not be triggered by a cough). This event triggered an alarm, and an immediate release of air to drop to a safe pressure and not to exceed PIP. Both of these functionalities are fast, and respond, at the latest, within few hundred milliseconds. The controller also assesses whether numerical values and sensor readings are reasonable, and changing over time. If this is not the case, it raises an technical alarm. All alarms are collected and maintained by an intelligent alarm manager, that provides the UI with the alarms to display in order of their importance.

The final functionality of the control module is the estimation of VTE (VTE stands for exhaled tidal volume), which is thee volume of air that made it in- and out of the lung. We estimate this number by integrating the expiratory flow during expiration, and subtracting the baseline flow used to sustain PEEP (details in the accompanying manuscript):

1.1.14.2 Architecture of the Controller

In terms of software components, the Controller consists of one main `controller` class, that is instantiated in its own thread. This object receives sensor-data from HAL, and computes control parameters, to change the mechanical position of valves. The Controller also receives ventilation control parameters (see *set_control()*). All exchanged variables are mutex'd.

The Controller also feeds the `Logger` a continuous stream of *SensorValues* objects so as to store high-temporal resolution data, including the control signals.

The main **control loop** is `pvp.controller._start_mainloop()` which queries the Hardware for new variables, and performs a PID update using *.pvp.controller._PID_update()*.

The Controller is **configured** by the *values* module,

The Controller can be launched alone, but was not intended to be launched alone. The alarm functionality requires the UI.

Classes:

ControlModuleBase([save_logs, flush_every])	Abstract controller class for simulation/hardware.
ControlModuleDevice([save_logs, ...])	Uses ControlModuleBase to control the hardware.
Balloon_Simulator(peep_valve)	Physics simulator for inflating a balloon with an attached PEEP valve.
ControlModuleSimulator([save_logs, ...])	Controlling Simulation.

Functions:

get_control_module([sim_mode, simulator_dt])	Generates control module.

class pvp.controller.control_module.**ControlModuleBase**(*save_logs: bool = False*, *flush_every: int = 10*)

Bases: `object`

Abstract controller class for simulation/hardware.

1. General notes: All internal variables fall in three classes, denoted by the beginning of the variable:

- *COPY_varname*: These are copies (for safe threading purposes) that are regularly sync'ed with internal variables.
- *__varname*: These are variables only used in the ControlModuleBase-Class
- *_varname*: These are variables used in derived classes.

2. Set and get values. Internal variables should only to be accessed though the **set_** and **get_** functions. These functions act on COPIES of internal variables (__ and _), that are sync'd every few iterations. How often this is done is adjusted by the variable *self._NUMBER_CONTROLL_LOOPS_UNTIL_UPDATE*. To avoid multiple threads manipulating the same variables at the same time, every manipulation of *COPY_* is surrounded by a thread lock.

Public Methods:

- *get_sensors()*: Returns a copy of the current sensor values.
- *get_alarms()*: Returns a List of all alarms, active and logged
- *get_control(ControlSetting)*: Sets a controll-setting. Is updated at latest within self._NUMBER_CONTROLL_LOOPS_UNTIL_UPDATE

- *get_past_waveforms()*: Returns a List of waveforms of pressure and volume during at the last N breath cycles, N<self. _RINGBUFFER_SIZE, AND clears this archive.
- *start()*: Starts the main-loop of the controller
- *stop()*: Stops the main-loop of the controller
- *set_control()*: Set the control
- *is_running()*: Returns a bool whether the main-thread is running
- *get_heartbeat()*: Returns a heartbeat, more specifically, the continuously increasing iteration-number of the main control loop.

Initializes the ControlModuleBase class.

Parameters

- **save_logs** (*bool, optional*) – Should sensor data and controls should be saved with the *DataLogger*? Defaults to False.
- **flush_every** (*int, optional*) – Flush and rotate logs every n breath cycles. Defaults to 10.

Raises **alert** – [description]

Methods:

__init__([save_logs, flush_every])	Initializes the ControlModuleBase class.
_initialize_set_to_COPY()	Makes a copy of internal variables.
_sensor_to_COPY()	
_controls_from_COPY()	
get_sensors()	A method callable from the outside to get a copy of sensorValues
get_alarms()	A method callable from the outside to get a copy of the alarms, that the controller checks: High airway pressure, and technical alarms.
set_control(control_setting)	A method callable from the outside to set alarms.
get_control(control_setting_name)	A method callable from the outside to get current control settings.
set_breath_detection(breath_detection)	
get_breath_detection()	Return current state of autonomous breath detection
_get_control_signal_in()	Produces the INSPIRATORY control-signal that has been calculated in *__calculate_control_signal_in(dt)*
_get_control_signal_out()	Produces the EXPIRATORY control-signal for the different states, i.e. open/close.
_control_reset()	Resets the internal controller cycle to zero, i.e. restarts the breath cycle.

continues on next page

Table 62 – continued from previous page

_PID_update(dt)	This instantiates the PID control algorithms. During the breathing cycle, it goes through the four states: 1) Rise to PIP, speed is controlled by flow (variable: *__SET_PIP_GAIN*) 2) Sustain PIP pressure 3) Quick fall to PEEP 4) Sustaint PEEP pressure Once the cycle is complete, it checks the cycle for any alarms, and starts a new one. A record of pressure/volume waveforms is kept and saved.
get_past_waveforms()	Public method to return a list of past waveforms from *__cycle_waveform_archive*. Note: After calling this function, archive is emptied! The format is - Returns a list of [Nx3] waveforms, of [time, pressure, volume] - Most recent entry is waveform_list[-1].
_start_mainloop()	Prototype method to start main PID loop.
start()	Method to start *_start_mainloop* as a thread.
stop()	Method to stop the main loop thread, and close the logfile.
is_running()	Public Method to assess whether the main loop thread is running.
get_heartbeat()	Returns an independent heart-beat of the controller, i.e. the internal loop counter incremented in *_start_mainloop*.

__init__(*save_logs: bool = False*, *flush_every: int = 10*)
Initializes the ControlModuleBase class.

Parameters

- **save_logs** (*bool, optional*) – Should sensor data and controls should be saved with the *DataLogger*? Defaults to False.
- **flush_every** (*int, optional*) – Flush and rotate logs every n breath cycles. Defaults to 10.

Raises **alert** – [description]

_initialize_set_to_COPY()
Makes a copy of internal variables. This is used to facilitate threading

_sensor_to_COPY()

_controls_from_COPY()

__comptest(*phase*, *ls*, *selector*)
Helper function to identify the index the first occurence of a number in *list* exceeding *threshold*, and returns phase[idx]

Parameters

- **phase** (*array*) – a list of numbers
- **list** (*array*) – array of bools with same length as phase
- **selector** (*string*) – 'first' or 'last' whichever is wanted

Returns phase[idx] where *idx* is first, or last point where numbers in list exceed threshold

Return type float

__analyze_last_waveform()

This goes through the last waveform, and updates the internal variables: VTE, PEEP, PIP, PIP_TIME, I_PHASE, FIRST_PEEP and BPM.

`get_sensors()` → *pvp.common.message.SensorValues*
A method callable from the outside to get a copy of sensorValues

Returns A set of current sensorvalues, handeled by the controller.

Return type *SensorValues*

`get_alarms()` → Union[None, Tuple[*pvp.alarm.alarm.Alarm*]]
A method callable from the outside to get a copy of the alarms, that the controller checks: High airway pressure, and technical alarms.

Returns A tuple of alarms

Return type *Union*[None, *Tuple*[*Alarm*]]

`set_control`(*control_setting:* pvp.common.message.ControlSetting)
A method callable from the outside to set alarms. This updates the entries of COPY with new control values.

Parameters **control_setting** (ControlSetting) – [description]

`get_control`(*control_setting_name:* pvp.common.values.ValueName) →
pvp.common.message.ControlSetting
A method callable from the outside to get current control settings. This returns values of COPY to the outside world.

Parameters **control_setting_name** (ValueName) – The specific control asked for

Returns ControlSettings-Object that contains relevant data

Return type *ControlSetting*

`set_breath_detection`(*breath_detection: bool*)

`get_breath_detection()` → bool
Return current state of autonomous breath detection

Returns bool

`__get_PID_error`(*ytarget*, *yis*, *dt*, *RC*)
Calculates the three terms for PID control. Also takes a timestep "dt" on which the integral-term is smoothed.

Parameters

- **ytarget** (*float*) – target value of pressure
- **yis** (*float*) – current value of pressure
- **dt** (*float*) – timestep
- **RC** (*float*) – time constant for calculation of integral term.

`__calculate_control_signal_in`(*dt*)

Calculates the PID control signal by:

- Combining the the three gain parameters.
- And smoothing the control signal with a moving window of three frames (~10ms)

Parameters **dt** (*float*) – timestep

`_get_control_signal_in()`
Produces the INSPIRATORY control-signal that has been calculated in *__calculate_control_signal_in(dt)*

Returns the numerical control signal for the inspiratory prop valve

Return type float

`_get_control_signal_out()`
Produces the EXPIRATORY control-signal for the different states, i.e. open/close

Returns numerical control signal for expiratory side: open (1) close (0)

Return type float

`_control_reset()`
Resets the internal controller cycle to zero, i.e. restarts the breath cycle. Used for autonomous breath detection.

`__test_for_alarms()`

Implements tests that are to be executed in the main control loop:

- Test for HAPA
- Test for Technical Alert, making sure sensor values are plausible
- Test for Technical Alert, make sure continuous in contact

Currently: Alarms are time.time() of first occurance.

`__start_new_breathcycle()`

Some housekeeping. This has to be executed when the next breath cycles starts:

- starts new breathcycle
- initializes newe __cycle_waveform
- analyzes last breath waveform for PIP, PEEP etc. with *__analyze_last_waveform()*
- flushes the logfile

`_PID_update(dt)`
This instantiates the PID control algorithms. During the breathing cycle, it goes through the four states:

1) Rise to PIP, speed is controlled by flow (variable: *__SET_PIP_GAIN*)
2) Sustain PIP pressure
3) Quick fall to PEEP
4) Sustaint PEEP pressure

Once the cycle is complete, it checks the cycle for any alarms, and starts a new one. A record of pressure/volume waveforms is kept and saved

Parameters `dt` (*float*) – timesstep since last update

`__save_values()`
Helper function to reorganize key parameters in the main PID control loop, into a *SensorValues* object, that can be stored in the logfile, using a method from the DataLogger.

`get_past_waveforms()`
Public method to return a list of past waveforms from *__cycle_waveform_archive*. Note: After calling this function, archive is emptied! The format is

- Returns a list of [Nx3] waveforms, of [time, pressure, volume]
- Most recent entry is waveform_list[-1]

Returns [Nx3] waveforms, of [time, pressure, volume]

Return type list

_start_mainloop()
Prototype method to start main PID loop. Will depend on simulation or device, specified below.

start()
Method to start *_start_mainloop* as a thread.

stop()
Method to stop the main loop thread, and close the logfile.

is_running()
Public Method to assess whether the main loop thread is running.

Returns Return true if and only if the main thread of controller is running.

Return type bool

get_heartbeat()
Returns an independent heart-beat of the controller, i.e. the internal loop counter incremented in *_start_mainloop*.

Returns exact value of *self._loop_counter*

Return type int

class pvp.controller.control_module.**ControlModuleDevice**(*save_logs=True*, *flush_every=10*, *config_file=None*)

Bases: *pvp.controller.control_module.ControlModuleBase*

Uses ControlModuleBase to control the hardware.

Initializes the ControlModule for the physical system. Inherits methods from ControlModuleBase

Parameters

- **save_logs** (*bool, optional*) – Should logs be kept? Defaults to True.
- **flush_every** (*int, optional*) – How often are log-files to be flushed, in units of main-loop-itertions? Defaults to 10.
- **config_file** (*str, optional*) – Path to device config file, e.g. 'pvp/io/config/dinky-devices.ini'. Defaults to None.

Methods:

__init__([save_logs, flush_every, config_file])	Initializes the ControlModule for the physical system.
_sensor_to_COPY()	Copies the current measurements to`COPY_sensor_values`, so that it can be queried from the outside.
_set_HAL(valve_open_in, valve_open_out)	Set Controls with HAL, decorated with a timeout.
_get_HAL()	Get sensor values from HAL, decorated with timeout.
set_valves_standby()	This returns valves back to normal setting (in: closed, out: open)
_start_mainloop()	This is the main loop.

__init__(*save_logs=True*, *flush_every=10*, *config_file=None*)
Initializes the ControlModule for the physical system. Inherits methods from ControlModuleBase

Parameters

- **save_logs** (*bool, optional*) – Should logs be kept? Defaults to True.
- **flush_every** (*int, optional*) – How often are log-files to be flushed, in units of main-loop-itertions? Defaults to 10.
- **config_file** (*str, optional*) – Path to device config file, e.g. 'pvp/io/config/dinky-devices.ini'. Defaults to None.

__get_hal(***kwargs*)

_sensor_to_COPY()
Copies the current measurements to`COPY_sensor_values`, so that it can be queried from the outside.

_set_HAL(*valve_open_in*, *valve_open_out*)
Set Controls with HAL, decorated with a timeout.

As hardware communication is the speed bottleneck. this code is slightly optimized in so far as only changes are sent to hardware.

Parameters

- **valve_open_in** (*float*) – setting of the inspiratory valve; should be in range [0,100]
- **valve_open_out** (*float*) – setting of the expiratory valve; should be 1/0 (open and close)

_get_HAL()
Get sensor values from HAL, decorated with timeout. As hardware communication is the speed bottleneck. this code is slightly optimized in so far as some sensors are queried only in certain phases of the breatch cycle. This is done to run the primary PID loop as fast as possible:

- pressure is always queried
- Flow is queried only outside of inspiration
- In addition, oxygen is only read every 5 seconds.

set_valves_standby()
This returns valves back to normal setting (in: closed, out: open)

_start_mainloop()
This is the main loop. This method should be run as a thread (see the *start()* method in *ControlModuleBase*)

class pvp.controller.control_module.**Balloon_Simulator**(*peep_valve*)
Bases: object

Physics simulator for inflating a balloon with an attached PEEP valve. For math, see https://en.wikipedia.org/wiki/Two-balloon_experiment

Methods:

get_pressure()	
set_flow_in(Qin, dt)	
set_flow_out(Qout, dt)	

continues on next page

Table 64 – continued from previous page

update(dt)	
OUupdate(variable, dt, mu, sigma, tau)	This is a simple function to produce an OU process on *variable*.
_reset()	Resets Balloon to default settings.

get_pressure()

set_flow_in(*Qin*, *dt*)

set_flow_out(*Qout*, *dt*)

update(*dt*)

OUupdate(*variable*, *dt*, *mu*, *sigma*, *tau*)

This is a simple function to produce an OU process on *variable*. It is used as model for fluctuations in measurement variables.

Parameters

- **variable** (*float*) – value of a variable at previous time step
- **dt** (*float*) – timestep
- **mu** (*float*) – mean
- **sigma** (*float*) – noise amplitude
- **tau** (*float*) – time scale

Returns value of "variable" at next time step

Return type float

_reset()

Resets Balloon to default settings.

class pvp.controller.control_module.**ControlModuleSimulator**(*save_logs: bool = False*, *simulator_dt=None*, *peep_valve_setting=5*)

Bases: *pvp.controller.control_module.ControlModuleBase*

Controlling Simulation.

Initializes the ControlModuleBase with the simple simulation (for testing/dev).

Parameters

- **save_logs** (*bool, optional*) – should logs be saved? (Useful for testing)
- **simulator_dt** (*float, optional*) – timestep between updates. Defaults to None.
- **peep_valve_setting** (*int, optional*) – Simulates action of a PEEP valve. Pressure cannot fall below. Defaults to 5.

Methods:

__init__([save_logs, simulator_dt, ...])	Initializes the ControlModuleBase with the simple simulation (for testing/dev).
_sensor_to_COPY()	Make the sensor value object from current (simulated) measurements
_start_mainloop()	This is the main loop.

__init__(*save_logs: bool = False*, *simulator_dt=None*, *peep_valve_setting=5*)
Initializes the ControlModuleBase with the simple simulation (for testing/dev).

Parameters

- **save_logs** (*bool, optional*) – should logs be saved? (Useful for testing)
- **simulator_dt** (*float, optional*) – timestep between updates. Defaults to None.
- **peep_valve_setting** (*int, optional*) – Simulates action of a PEEP valve. Pressure cannot fall below. Defaults to 5.

__SimulatedPropValve(*x*)
This simulates the action of a proportional valve. Flow-current-curve eye-balled from generic prop vale with logistic activation.

Parameters **x** (*float*) – A control variable [like pulse-width-duty cycle or mA]

Returns flow through the valve

Return type float

__SimulatedSolenoid(*x*)
This simulates the action of a two-state Solenoid valve.

Parameters **x** (*float*) – If x==0: valve closed; x>0: flow set to "1"

Returns current flow

Return type float

_sensor_to_COPY()
Make the sensor value object from current (simulated) measurements

_start_mainloop()
This is the main loop. This method should be run as a thread (see the *start()* method in *ControlModuleBase*)

pvp.controller.control_module.**get_control_module**(*sim_mode=False*, *simulator_dt=None*)
Generates control module.

Parameters

- **sim_mode** (*bool, optional*) – if `true`: returns simulation, else returns hardware. Defaults to False.
- **simulator_dt** (*float, optional*) – a timescale for thee simulation. Defaults to None.

Returns Either configured for simulation, or physical device.

Return type ControlModule-Object

1.1.15 common module

1.1.15.1 Values

Parameterization of variables and values used in PVP.

Value objects define the existence and behavior of values, including creating *Display* and *Plot* widgets in the GUI, and the contents of *SensorValues* and *ControlSetting* s used between the GUI and controller.

Classes:

ValueName(value)	Canonical names of all values used in PVP.
Value(name, units, abs_range, safe_range, ...)	Class to parameterize how a value is used in PVP.

Data:

VALUES	Declaration of all values used by PVP
SENSOR	Sensor values
CONTROL	Values to control but not monitor.
DISPLAY_MONITOR	Those sensor values that should also have a widget created in the GUI
DISPLAY_CONTROL	Control values that should also have a widget created in the GUI
PLOTS	Values that can be plotted

class pvp.common.values.**ValueName**(*value*)

Bases: enum.Enum

Canonical names of all values used in PVP.

Attributes:

PIP
PIP_TIME
PEEP
PEEP_TIME
BREATHS_PER_MINUTE
INSPIRATION_TIME_SEC
IE_RATIO
FIO2
VTE
PRESSURE
FLOWOUT

PIP = 1

PIP_TIME = 2

PEEP = 3

PEEP_TIME = 4

BREATHS_PER_MINUTE = 5

INSPIRATION_TIME_SEC = 6

IE_RATIO = 7

FIO2 = 8

VTE = 9

PRESSURE = 10

FLOWOUT = 11

class pvp.common.values.**Value**(*name: str*, *units: str*, *abs_range: tuple*, *safe_range: tuple*, *decimals: int*, *control: bool*, *sensor: bool*, *display: bool*, *plot: bool = False*, *plot_limits: typing.Union[None, typing.Tuple[pvp.common.values.ValueName]] = None*, *control_type: typing.Union[None, str] = None*, *group: typing.Union[None, dict] = None*, *default: (<class 'int'>, <class 'float'>) = None*)

Bases: object

Class to parameterize how a value is used in PVP.

Sets whether a value is a sensor value, a control value, whether it should be plotted, and other details for the rest of the system to determine how to use it.

Values should only be declared in this file so that they are kept consistent with *ValueName* and to not leak stray values anywhere else in the program.

Parameters

- **name** (*str*) – Human-readable name of the value
- **units** (*str*) – Human-readable description of units
- **abs_range** (*tuple*) – tuple of ints or floats setting the logical limit of the value, eg. a percent between 0 and 100, (0, 100)
- **safe_range** (*tuple*) – tuple of ints or floats setting the safe ranges of the value,

  note:

  ```
  this is not the same thing as the user-set alarm values,
  though the user-set alarm values are initialized as ``safe_range``.
  ```

- **decimals** (*int*) – the number of decimals of precision used when displaying the value
- **control** (*bool*) – Whether or not the value is used to control ventilation
- **sensor** (*bool*) – Whether or not the value is a measured sensor value
- **display** (*bool*) – whether the value should be created as a gui.widgets.Display widget.
- **plot** (*bool*) – whether or not the value is plottable in the center plot window
- **plot_limits** (*None, tuple(ValueName)*) – If plottable, and the plotted value has some alarm limits for another value, plot those limits as horizontal lines in the plot. eg. the PIP alarm range limits should be plotted on the Pressure plot
- **control_type** (*None, "slider", "record"*) – If a control sets whether the control should use a slider or be set by recording recent sensor values.
- **group** (*None, str*) – Unused currently, but to be used to create subgroups of control & display widgets
- **default** (*None, int, float*) – Default value, if any. (Not automatically set in the GUI.)

Methods:

__init__(name, units, abs_range, safe_range, ...)	**param name** Human-readable name of the value
to_dict()	Gather up all attributes and return as a dict!

Attributes:

name	Human readable name of value
abs_range	tuple of ints or floats setting the logical limit of the value, eg.
safe_range	tuple of ints or floats setting the safe ranges of the value,
decimals	The number of decimals of precision used when displaying the value
default	Default value, if any.
control	Whether or not the value is used to control ventilation
sensor	Whether or not the value is a measured sensor value
display	Whether the value should be created as a `gui.widgets.Display` widget.
control_type	If a control sets whether the control should use a slider or be set by recording recent sensor values.
group	Unused currently, but to be used to create subgroups of control & display widgets
plot	whether or not the value is plottable in the center plot window
plot_limits	If plottable, and the plotted value has some alarm limits for another value, plot those limits as horizontal lines in the plot.

__init__(*name: str*, *units: str*, *abs_range: tuple*, *safe_range: tuple*, *decimals: int*, *control: bool*, *sensor: bool*, *display: bool*, *plot: bool = False*, *plot_limits: typing.Union[None, typing.Tuple[pvp.common.values.ValueName]] = None*, *control_type: typing.Union[None, str] = None*, *group: typing.Union[None, dict] = None*, *default: (<class 'int'>, <class 'float'>) = None*)

Parameters

- **name** (*str*) – Human-readable name of the value
- **units** (*str*) – Human-readable description of units
- **abs_range** (*tuple*) – tuple of ints or floats setting the logical limit of the value, eg. a percent between 0 and 100, (0, 100)
- **safe_range** (*tuple*) – tuple of ints or floats setting the safe ranges of the value,

  note:

  ```
  this is not the same thing as the user-set alarm values,
  though the user-set alarm values are initialized as ``safe_range``.
  ```

- **decimals** (*int*) – the number of decimals of precision used when displaying the value
- **control** (*bool*) – Whether or not the value is used to control ventilation
- **sensor** (*bool*) – Whether or not the value is a measured sensor value

- **display** (*bool*) – whether the value should be created as a `gui.widgets.Display` widget.
- **plot** (*bool*) – whether or not the value is plottable in the center plot window
- **plot_limits** (*None, tuple(ValueName)*) – If plottable, and the plotted value has some alarm limits for another value, plot those limits as horizontal lines in the plot. eg. the PIP alarm range limits should be plotted on the Pressure plot
- **control_type** (*None, "slider", "record"*) – If a control sets whether the control should use a slider or be set by recording recent sensor values.
- **group** (*None, str*) – Unused currently, but to be used to create subgroups of control & display widgets
- **default** (*None, int, float*) – Default value, if any. (Not automatically set in the GUI.)

property name: *str*
Human readable name of value

Returns str

property abs_range: *tuple*
tuple of ints or floats setting the logical limit of the value, eg. a percent between 0 and 100, (0, 100)

Returns tuple

property safe_range: *tuple*
tuple of ints or floats setting the safe ranges of the value,

note:

```
this is not the same thing as the user-set alarm values,
though the user-set alarm values are initialized as ``safe_range``.
```

Returns tuple

property decimals: *int*
The number of decimals of precision used when displaying the value

Returns int

property default
Default value, if any. (Not automatically set in the GUI.)

property control: *bool*
Whether or not the value is used to control ventilation

Returns bool

property sensor: *bool*
Whether or not the value is a measured sensor value

Returns bool

property display
Whether the value should be created as a `gui.widgets.Display` widget.

Returns bool

property control_type
If a control sets whether the control should use a slider or be set by recording recent sensor values.

Returns None, “slider”, “record”

property group
Unused currently, but to be used to create subgroups of control & display widgets

Returns None, str

property plot
whether or not the value is plottable in the center plot window

Returns bool

property plot_limits
If plottable, and the plotted value has some alarm limits for another value, plot those limits as horizontal lines in the plot. eg. the PIP alarm range limits should be plotted on the Pressure plot

Returns None, typing.Tuple[ValueName]

to_dict() → dict
Gather up all attributes and return as a dict!

Returns dict

pvp.common.values.**VALUES = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.IE_RATIO: 7>, <pvp.common.values.Value object>), (<ValueName.PIP_TIME: 2>, <pvp.common.values.Value object>), (<ValueName.PEEP_TIME: 4>, <pvp.common.values.Value object>), (<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.VTE: 9>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])**
Declaration of all values used by PVP

pvp.common.values.**SENSOR = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.VTE: 9>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])**
Sensor values

Automatically generated as all *Value* objects in *VALUES* where `sensor == True`

pvp.common.values.**CONTROL = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.IE_RATIO: 7>, <pvp.common.values.Value object>), (<ValueName.PIP_TIME: 2>, <pvp.common.values.Value object>), (<ValueName.PEEP_TIME: 4>, <pvp.common.values.Value object>)])**
Values to control but not monitor.

Automatically generated as all *Value* objects in *VALUES* where `control == True`

pvp.common.values.**DISPLAY_MONITOR = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.VTE: 9>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])**

Those sensor values that should also have a widget created in the GUI

Automatically generated as all *Value* objects in *VALUES* where `sensor == True` and `display == True`

pvp.common.values.**DISPLAY_CONTROL = OrderedDict([(<ValueName.PIP: 1>, <pvp.common.values.Value object>), (<ValueName.PEEP: 3>, <pvp.common.values.Value object>), (<ValueName.BREATHS_PER_MINUTE: 5>, <pvp.common.values.Value object>), (<ValueName.INSPIRATION_TIME_SEC: 6>, <pvp.common.values.Value object>), (<ValueName.IE_RATIO: 7>, <pvp.common.values.Value object>), (<ValueName.PIP_TIME: 2>, <pvp.common.values.Value object>)])**

Control values that should also have a widget created in the GUI

Automatically generated as all *Value* objects in *VALUES* where `control == True` and `display == True`

pvp.common.values.**PLOTS = OrderedDict([(<ValueName.PRESSURE: 10>, <pvp.common.values.Value object>), (<ValueName.FLOWOUT: 11>, <pvp.common.values.Value object>), (<ValueName.FIO2: 8>, <pvp.common.values.Value object>)])**

Values that can be plotted

Automatically generated as all *Value* objects in *VALUES* where `plot == True`

1.1.15.2 Message

Message objects that define the API between modules in the system.

- *SensorValues* are used to communicate sensor readings between the controller, GUI, and alarm manager
- *ControlSetting* is used to set ventilation controls from the GUI to the controller.

Classes:

SensorValues([timestamp, loop_counter, ...])	Structured class for communicating sensor readings throughout PVP.
ControlSetting(name[, value, min_value, ...])	Message containing ventilation control parameters.
ControlValues(control_signal_in, ...)	Class to save control values, analogous to SensorValues.
DerivedValues(timestamp, breath_count, ...)	Class to save derived values, analogous to SensorValues.

class pvp.common.message.**SensorValues**(*timestamp=None*, *loop_counter=None*, *breath_count=None*, *vals=typing.Union[NoneType, typing.Dict[ForwardRef('ValueName'), float]]*, ***kwargs*)

Bases: `object`

Structured class for communicating sensor readings throughout PVP.

Should be instantiated with each of the *SensorValues.additional_values*, and values for all *ValueName* s in *values.SENSOR* by passing them in the `vals` kwarg.

An `AssertionError` if an incomplete set of values is given.

Values can be accessed either via attribute name (`SensorValues.PIP`) or like a dictionary (`SensorValues['PIP']`)

Parameters

- **timestamp** (*float*) – from time.time(). must be passed explicitly or as an entry in `vals`
- **loop_counter** (*int*) – number of control_module loops. must be passed explicitly or as an entry in `vals`
- **breath_count** (*int*) – number of breaths taken. must be passed explicitly or as an entry in `vals`
- **vals** (*None, dict*) – Dict of `{ValueName: float}` that contains current sensor readings. Can also be equivalently given as `kwargs` . if None, assumed values are being passed as kwargs, but an exception will be raised if they aren't.
- ****kwargs** – sensor readings, must be in `pvp.values.SENSOR.keys`

Attributes:

additional_values	Additional attributes that are not *ValueName* s that are expected in each SensorValues message

Methods:

__init__([timestamp, loop_counter, ...])	**param timestamp** from time.time(). must be passed explicitly or as an entry in `vals`
to_dict()	Return a dictionary of all sensor values and additional values

additional_values = ('timestamp', 'loop_counter', 'breath_count')
Additional attributes that are not *ValueName* s that are expected in each SensorValues message

__init__(*timestamp=None*, *loop_counter=None*, *breath_count=None*, *vals=typing.Union[NoneType, typing.Dict[ForwardRef('ValueName'), float]]*, ***kwargs*)

Parameters

- **timestamp** (*float*) – from time.time(). must be passed explicitly or as an entry in `vals`
- **loop_counter** (*int*) – number of control_module loops. must be passed explicitly or as an entry in `vals`
- **breath_count** (*int*) – number of breaths taken. must be passed explicitly or as an entry in `vals`
- **vals** (*None, dict*) – Dict of `{ValueName: float}` that contains current sensor readings. Can also be equivalently given as `kwargs` . if None, assumed values are being passed as kwargs, but an exception will be raised if they aren't.
- ****kwargs** – sensor readings, must be in `pvp.values.SENSOR.keys`

to_dict() → dict
Return a dictionary of all sensor values and additional values

Returns dict

class pvp.common.message.**ControlSetting**(*name:* pvp.common.values.ValueName, *value:* *float* = *None*, *min_value:* *float* = *None*, *max_value:* *float* = *None*, *timestamp:* *float* = *None*, *range_severity:* AlarmSeverity = *None*)

Bases: `object`

Message containing ventilation control parameters.

At least **one of** `value`, `min_value`, or `max_value` must be given (unlike *SensorValues* which requires all fields to be present) – eg. in the case where one is setting alarm thresholds without changing the actual set value

When a parameter has multiple alarm limits for different alarm severities, the severity should be passed to `range_severity`

Parameters

- **name** (*ValueName*) – Name of value being set
- **value** (*float*) – Value to set control
- **min_value** (*float*) – Value to set control minimum (typically used for alarm thresholds)
- **max_value** (*float*) – Value to set control maximum (typically used for alarm thresholds)
- **timestamp** (*float*) – `time.time()` control message was generated
- **range_severity** (*AlarmSeverity*) – Some control settings have multiple limits for different alarm severities, this attr, when present, specifies which is being set.

Methods:

__init__(name[, value, min_value, ...])	Message containing ventilation control parameters.

__init__(*name:* pvp.common.values.ValueName, *value:* *float* = *None*, *min_value:* *float* = *None*, *max_value:* *float* = *None*, *timestamp:* *float* = *None*, *range_severity:* AlarmSeverity = *None*)

Message containing ventilation control parameters.

At least **one of** `value`, `min_value`, or `max_value` must be given (unlike *SensorValues* which requires all fields to be present) – eg. in the case where one is setting alarm thresholds without changing the actual set value

When a parameter has multiple alarm limits for different alarm severities, the severity should be passed to `range_severity`

Parameters

- **name** (*ValueName*) – Name of value being set
- **value** (*float*) – Value to set control
- **min_value** (*float*) – Value to set control minimum (typically used for alarm thresholds)
- **max_value** (*float*) – Value to set control maximum (typically used for alarm thresholds)
- **timestamp** (*float*) – `time.time()` control message was generated
- **range_severity** (*AlarmSeverity*) – Some control settings have multiple limits for different alarm severities, this attr, when present, specifies which is being set.

class pvp.common.message.**ControlValues**(*control_signal_in*, *control_signal_out*)

Bases: `object`

Class to save control values, analogous to SensorValues.

Used by the controller to save waveform data in *DataLogger.store_waveform_data()* and `ControlModuleBase.__save_values`()

Key difference: SensorValues come exclusively from the sensors, ControlValues contains controller variables, i.e. control signals and controlled signals (the flows). :param control_signal_in: :param control_signal_out:

class pvp.common.message.**DerivedValues**(*timestamp*, *breath_count*, *I_phase_duration*, *pip_time*, *peep_time*, *pip*, *pip_plateau*, *peep*, *vte*)

Bases: object

Class to save derived values, analogous to SensorValues.

Used by controller to store derived values (like PIP from Pressure) in *DataLogger.store_derived_data()* and in ControlModuleBase.__analyze_last_waveform`()

Key difference: SensorValues come exclusively from the sensors, DerivedValues contain estimates of I_PHASE_DURATION, PIP_TIME, PEEP_time, PIP, PIP_PLATEAU, PEEP, and VTE. :param timestamp: :param breath_count: :param I_phase_duration: :param pip_time: :param peep_time: :param pip: :param pip_plateau: :param peep: :param vte:

1.1.15.3 Loggers

Logging functionality

There are two types of loggers:

- *loggers.init_logger()* creates a standard logging.Logger -based logging system for debugging and recording system events, and a
- *loggers.DataLogger* - a tables - based class to store continuously measured sensor values.

Data:

_LOGGERS	list of strings, which loggers have been created already.

Functions:

init_logger(module_name[, log_level, ...])	Initialize a logger for logging events.
update_logger_sizes()	Adjust each logger's maxBytes attribute so that the total across all loggers is prefs.LOGGING_MAX_BYTES

Classes:

DataLogger([compression_level])	Class for logging numerical respiration data and control settings. Creates a hdf5 file with this general structure: / root |--- waveforms (group) | |--- time | pressure_data | flow_out | control_signal_in | control_signal_out | FiO2 | Cycle No. | |--- controls (group) | |--- (time, controllsignal) | |--- derived_quantities (group) | |--- (time, Cycle No, I_PHASE_DURATION, PIP_TIME, PEEP_time, PIP, PIP_PLATEAU, PEEP, VTE) | |--- program_information (group) | |--- (version & githash).

pvp.common.loggers.**_LOGGERS = ['pvp.common.prefs', 'pvp.alarm.alarm_manager']**

list of strings, which loggers have been created already.

pvp.common.loggers.**init_logger**(*module_name: str*, *log_level: Optional[int] = None*, *file_handler: bool = True*) → logging.Logger

Initialize a logger for logging events.

To keep logs sensible, you should usually initialize the logger with the name of the module that's using it, eg:

```
logger = init_logger(__name__)
```

If a logger has already been initialized (ie. its name is in *loggers._LOGGERS*, return that.

otherwise configure and return the logger such that

- its LOGLEVEL is set to `prefs.LOGLEVEL`
- It formats logging messages with logger name, time, and logging level
- if a file handler is specified (default), create a `logging.RotatingFileHandler` according to params set in `prefs`

Parameters

- **module_name** (*str*) – module name used to generate filename and name logger
- **log_level** (*int*) – one of **:var:`logging.DEBUG`**, **:var:`logging.INFO`**, **:var:`logging.WARNING`**, or **:var:`logging.ERROR`**
- **file_handler** (*bool, str*) – if `True`, (default), log in <logdir>/module_name.log . if `False`, don't log to disk.

Returns Logger 4 u 2 use

Return type `logging.Logger`

`pvp.common.loggers.`**update_logger_sizes**()

Adjust each logger's `maxBytes` attribute so that the total across all loggers is `prefs.LOGGING_MAX_BYTES`

class `pvp.common.loggers.`**DataLogger**(*compression_level: int = 9*)

Bases: `object`

Class for logging numerical respiration data and control settings. Creates a hdf5 file with this general structure:

/ root |— waveforms (group) | |— time | pressure_data | flow_out | control_signal_in | control_signal_out | FiO2 | Cycle No. | |— controls (group) | |— (time, controllsignal) | |— derived_quantities (group) | |— (time, Cycle No, I_PHASE_DURATION, PIP_TIME, PEEP_time, PIP, PIP_PLATEAU, PEEP, VTE) | |— program_information (group) | |— (version & githash)

Public Methods: close_logfile(): Flushes, and closes the logfile. store_waveform_data(SensorValues): Takes data from SensorValues, but DOES NOT FLUSH store_controls(): Store controls in the same file? TODO: Discuss flush_logfile(): Flush the data into the file

Initialized the coontinuous numerical logger class.

Parameters **compression_level** (*int, optional*) – Compression level of the hdf5 file. Defaults to 9.

Methods:

__init__([compression_level])	Initialized the coontinuous numerical logger class.
_open_logfile()	Opens the hdf5 file and generates the file structure.
close_logfile()	Flushes & closes the open hdf file.
store_program_data()	Appends program metadata to the logfile: githash and version
store_waveform_data(sensor_values, ...)	Appends a datapoint to the file for continuous logging of streaming data.

continues on next page

Table 78 – continued from previous page

store_control_command(control_setting)	Appends a datapoint to the event-table, derived from ControlSettings
store_derived_data(derived_values)	Appends a datapoint to the event-table, derived the continuous data (PIP, PEEP etc.)
flush_logfile()	This flushes the datapoints to the file.
check_files()	make sure that the file's are not getting too large.
rotation_newfile()	This rotates through filenames, similar to a ring-buffer, to make sure that the program does not run of of space/
load_file([filename])	This loads a hdf5 file, and returns data to the user as a dictionary with two keys: waveform_data and control_data
log2mat([filename])	Translates the compressed hdf5 into a matlab file containing a matlab struct. This struct has the same structure as the hdf5 file, but is not compressed. Use for any file: dl = DataLogger() dl.log2mat(filename) The file is saved at the same path as *.mat* file, where the content is represented as matlab-structs.
log2csv([filename])	Translates the compressed hdf5 into three csv files containing:

__init__(*compression_level: int = 9*)
Initialized the coontinuous numerical logger class.

Parameters **compression_level** (*int, optional*) – Compression level of the hdf5 file. Defaults to 9.

_open_logfile()
Opens the hdf5 file and generates the file structure.

close_logfile()
Flushes & closes the open hdf file.

store_program_data()
Appends program metadata to the logfile: githash and version

store_waveform_data(*sensor_values: SensorValues*, *control_values: ControlValues*)
Appends a datapoint to the file for continuous logging of streaming data. NOTE: Not flushed yet.

Parameters

- **sensor_values** (SensorValues) – SensorValues to be stored in the file.
- **control_values** (ControlValues) – ControlValues to be stored in the file

store_control_command(*control_setting: ControlSetting*)
Appends a datapoint to the event-table, derived from ControlSettings

Parameters **control_setting** (ControlSetting) – ControlSettings object, the content of which should be stored

store_derived_data(*derived_values: DerivedValues*)
Appends a datapoint to the event-table, derived the continuous data (PIP, PEEP etc.)

Parameters **derived_values** (DerivedValues) – DerivedValues object, the content of which should be stored

flush_logfile()
This flushes the datapoints to the file. To be executed every other second, e.g. at the end of breath cycle.

check_files()
make sure that the file's are not getting too large.

rotation_newfile()
This rotates through filenames, similar to a ringbuffer, to make sure that the program does not run of of space/

load_file(*filename=None*)
This loads a hdf5 file, and returns data to the user as a dictionary with two keys: waveform_data and control_data

Parameters **filename** (*str, optional*) – Path to a hdf5-file. If none is given, uses currently open file. Defaults to None.

Returns Containing the data arranged as ` {"waveform_data": waveform_data, "control_data": control_data, "derived_data": derived_data, "program_information": program_data}`

Return type dictionary

log2mat(*filename=None*)
Translates the compressed hdf5 into a matlab file containing a matlab struct. This struct has the same structure as the hdf5 file, but is not compressed. Use for any file:

dl = DataLogger() dl.log2mat(filename)

The file is saved at the same path as *.mat* file, where the content is represented as matlab-structs.

Parameters **filename** (*str, optional*) – Path to a hdf5-file. If none is given, uses currently open file. Defaults to None.

log2csv(*filename=None*)

Translates the compressed hdf5 into three csv files containing:

- waveform_data (measurement once per cycle)
- derived_quantities (PEEP, PIP etc.)
- control_commands (control commands sent to the controller)

This approximates the structure contained in the hdf5 file. Use for any file:

dl = DataLogger() dl.log2csv(filename)

Parameters **filename** (*str, optional*) – Path to a hdf5-file. If none is given, uses currently open file. Defaults to None.

1.1.15.4 Prefs

Prefs set configurable parameters used throughout PVP.

See *prefs._DEFAULTS* for description of all available parameters

Prefs are stored in a .json file, by default located at `~/pvp/prefs.json` . Prefs can be manually changed by editing this file (when the system is not running, when the system is running use *prefs.set_pref()*).

When any module in pvp is first imported, the *prefs.init()* function is called that

- Makes any directories listed in *prefs._DIRECTORIES*
- Declares all prefs as their default values from *prefs._DEFAULTS* to ensure they are always defined
- Loads the existing `prefs.json` file and updates values from their defaults

Prefs can be gotten and set from anywhere in the system with *prefs.get_pref()* and *prefs.set_pref()* . Prefs are stored in a `multiprocessing.Manager` dictionary which makes these methods both thread- and process-safe. Whenever a pref is set, the `prefs.json` file is updated to reflect the new value, so preferences are durable between runtimes.

Additional `prefs` should be added by adding an entry in the *prefs._DEFAULTS* dict rather than hardcoding them elsewhere in the program.

Data:

_DIRECTORIES	Directories to ensure are created and added to prefs.
_DEFAULTS	Declare all available parameters and set default values.

Functions:

set_pref(key, val)	Sets a pref in the manager and, if *prefs.LOADED* is True, calls *prefs.save_prefs()*
get_pref([key])	Get global configuration value
load_prefs(prefs_fn)	Load prefs from a .json prefs file, combining (and overwriting) any existing prefs, and then saves.
save_prefs([prefs_fn])	Dumps loaded prefs to `PREFS_FN`.
make_dirs()	ensures _DIRECTORIES are created and added to prefs.
init()	Initialize prefs.

pvp.common.prefs.**_PREF_MANAGER = <multiprocessing.managers.SyncManager object>**
The `multiprocessing.Manager` that stores prefs during system operation

pvp.common.prefs.**_PREFS = <DictProxy object, typeid 'dict'>**
The dict created by *prefs._PREF_MANAGER* to store prefs.

pvp.common.prefs.**_LOGGER = <Logger pvp.common.prefs (WARNING)>**
A `logging.Logger` to log pref init and setting events

pvp.common.prefs.**_LOCK = <Lock(owner=None)>**
Locks access to *prefs_fn*

Type `mp.Lock`

pvp.common.prefs.**_DIRECTORIES = {'DATA_DIR': '/home/docs/pvp/logs', 'LOG_DIR': '/home/docs/pvp/logs', 'VENT_DIR': '/home/docs/pvp'}**
Directories to ensure are created and added to prefs.

- `VENT_DIR`: ~/pvp - base directory for user storage
- `LOG_DIR`: ~/pvp/logs - for storage of event and alarm logs
- `DATA_DIR`: ~/pvp/data - for storage of waveform data

pvp.common.prefs.**LOADED = <Synchronized wrapper for c_bool(True)>**
flag to indicate whether prefs have been loaded (and thus *set_pref()* should write to disk).

uses a `multiprocessing.Value` to be thread and process safe.

Type bool

pvp.common.prefs._DEFAULTS = {'BREATH_DETECTION': True, 'BREATH_PRESSURE_DROP': 4, 'CONTROLLER_LOOPS_UNTIL_UPDATE': 1, 'CONTROLLER_LOOP_UPDATE_TIME': 0.0, 'CONTROLLER_LOOP_UPDATE_TIME_SIMULATOR': 0.005, 'CONTROLLER_MAX_FLOW': 10, 'CONTROLLER_MAX_PRESSURE': 100, 'CONTROLLER_MAX_STUCK_SENSOR': 5, 'CONTROLLER_RINGBUFFER_SIZE': 100, 'COUGH_DURATION': 0.1, 'ENABLE_DIALOGS': True, 'ENABLE_WARNINGS': True, 'GUI_STATE_FN': 'gui_state.json', 'GUI_UPDATE_TIME': 0.05, 'HEARTBEAT_TIMEOUT': 0.02, 'LOGGING_MAX_BYTES': 2147483648, 'LOGGING_MAX_FILES': 5, 'LOGLEVEL': 'WARNING', 'OXYGEN_READ_FREQUENCY': 2, 'PREFS_FN': None, 'TIMEOUT': 0.05, 'TIME_FIRST_START': None}

Declare all available parameters and set default values. If no default, set as None.

- PREFS_FN - absolute path to the prefs file
- TIME_FIRST_START - time when the program has been started for the first time
- VENT_DIR: ~/pvp - base directory for user storage
- LOG_DIR: ~/pvp/logs - for storage of event and alarm logs
- DATA_DIR: ~/pvp/data - for storage of waveform data
- LOGGING_MAX_BYTES : the **total** storage space for all loggers – each logger gets LOGGING_MAX_BYTES/len(loggers) space (2GB by default)
- LOGGING_MAX_FILES : number of files to split each logger's logs across (default: 5)
- LOGLEVEL: One of ('DEBUG', 'INFO', 'WARNING', 'EXCEPTION') that sets the minimum log level that is printed and written to disk
- TIMEOUT: timeout used for timeout decorators on time-sensitive operations (in seconds, default 0.05)
- HEARTBEAT_TIMEOUT: Time between heartbeats between GUI and controller after which contact is assumed to be lost (in seconds, default 0.02)
- GUI_STATE_FN: Filename of gui control state file, relative to VENT_DIR (default: gui_state.json)
- GUI_UPDATE_TIME: Time between calls of *PVP_Gui.update_gui()* (in seconds, default: 0.05)
- ENABLE_DIALOGS: Enable all GUI dialogs – set as False when testing on virtual frame buffer that doesn't support them (default: True and should stay that way)
- ENABLE_WARNINGS: Enable user warnings and value change confirmations (default: True)
- CONTROLLER_MAX_FLOW: Maximum flow, above which the controller considers a sensor error (default: 10)
- CONTROLLER_MAX_PRESSURE: Maximum pressure, above which the controller considers a sensor error (default: 100)
- CONTROLLER_MAX_STUCK_SENSOR: Max amount of time (in s) before considering a sensor stuck (default: 0.2)
- CONTROLLER_LOOP_UPDATE_TIME: Amount of time to sleep in between controller update times when using *ControlModuleDevice* (default: 0.0)
- CONTROLLER_LOOP_UPDATE_TIME_SIMULATOR: Amount of time to sleep in between controller updates when using *ControlModuleSimulator* (default: 0.005)
- CONTROLLER_LOOPS_UNTIL_UPDATE: Number of controller loops in between updating its externally-available COPY attributes retrieved by ControlModuleBase.get_sensor() et al
- CONTROLLER_RINGBUFFER_SIZE: Maximum number of breath cycle records to be kept in memory (default: 100)
- COUGH_DURATION: Amount of time the high-pressure alarm limit can be exceeded and considered a cough (in seconds, default: 0.1)

- BREATH_PRESSURE_DROP: Amount pressure can drop below set PEEP before being considered an autonomous breath when in breath detection mode
- BREATH_DETECTION: Whether the controller should detect autonomous breaths in order to reset ventilation cycles (default: True)

pvp.common.prefs.**set_pref**(*key: str*, *val*)
Sets a pref in the manager and, if *prefs.LOADED* is True, calls *prefs.save_prefs()*

Parameters

- **key** (*str*) – Name of pref key
- **val** – Value to set

pvp.common.prefs.**get_pref**(*key: Optional[str] = None*)
Get global configuration value

Parameters **key** (*str, None*) – get configuration value with specific key . if None , return all config values.

pvp.common.prefs.**load_prefs**(*prefs_fn: str*)
Load prefs from a .json prefs file, combining (and overwriting) any existing prefs, and then saves.

Called on pvp import by *prefs.init()*

Also initializes *prefs._LOGGER*

Note: once this function is called, *set_pref()* will update the prefs file on disk. So if *load_prefs()* is called again at any point it should not change prefs.

Parameters **prefs_fn** (*str*) – path of prefs.json

pvp.common.prefs.**save_prefs**(*prefs_fn: Optional[str] = None*)
Dumps loaded prefs to PREFS_FN.

Parameters **prefs_fn** (*str*) – Location to dump prefs. if None, use existing PREFS_FN

pvp.common.prefs.**make_dirs**()
ensures _DIRECTORIES are created and added to prefs.

pvp.common.prefs.**init**()
Initialize prefs. Called in pvp.__init__.py to ensure prefs are initialized before anything else.

1.1.15.5 Unit Conversion

Functions that convert between units

Each function should accept a single float as an argument and return a single float

Used by the GUI to display values in different units. Widgets use these as

- _convert_in functions to convert units from the base unit to the displayed unit and
- _convert_out functions to convert units from the displayed unit to the base unit.

Note: Unit conversions are cosmetic – values are always kept as the base unit internally (ie. cmH2O for pressure) and all that is changed is the displayed value in the GUI.

Functions:

cmH2O_to_hPa(pressure)	Convert cmH2O to hPa
hPa_to_cmH2O(pressure)	Convert hPa to cmH2O
rounded_string(value[, decimals])	Create a rounded string of a number that doesnt have trailing .0 when decimals = 0

pvp.common.unit_conversion.**cmH2O_to_hPa**(*pressure: float*) → float
Convert cmH2O to hPa

Parameters **pressure** (*float*) – Pressure in cmH2O

Returns Pressure in hPa (pressure / 1.0197162129779)

Return type float

pvp.common.unit_conversion.**hPa_to_cmH2O**(*pressure: float*) → float
Convert hPa to cmH2O

Parameters **pressure** (*float*) – Pressure in hPa

Returns Pressure in cmH2O (pressure * 1.0197162129779)

Return type float

pvp.common.unit_conversion.**rounded_string**(*value: float*, *decimals: int = 0*) → str
Create a rounded string of a number that doesnt have trailing .0 when decimals = 0

Parameters

- **value** (*float*) – Value to stringify
- **decimals** (*int*) – Number of decimal places to round to

Returns Clean rounded string version of number

Return type str

1.1.15.6 utils

Exceptions:

TimeoutException

Functions:

time_limit(seconds)

timeout(func)	Defines a decorator for a 50ms timeout.
get_version()	Returns PVP version, and if available githash, as a string.

exception pvp.common.utils.**TimeoutException**
Bases: Exception

pvp.common.utils.**time_limit**(*seconds*)

pvp.common.utils.**timeout**(*func*)
Defines a decorator for a 50ms timeout. Usage/Test:

@timeout def foo(sleeptime):

time.sleep(sleeptime)

print("hello")

pvp.common.utils.**get_version**()
Returns PVP version, and if available githash, as a string.

1.1.15.7 fashion

Decorators for dangerous functions

Functions:

pigpio_command(func)	

pvp.common.fashion.**pigpio_command**(*func*)

1.1.16 pvp.io package

1.1.16.1 pvp.io.hal module

Module for interacting with physical and/or simulated devices installed on the ventilator.

Classes:

Hal([config_file])	Hardware Abstraction Layer for ventilator hardware.

class pvp.io.hal.**Hal**(*config_file='pvp/io/config/devices.ini'*)
Bases: object

Hardware Abstraction Layer for ventilator hardware. Defines a common API for interacting with the sensors & actuators on the ventilator. The types of devices installed on the ventilator (real or simulated) are specified in a configuration file.

Initializes HAL from config_file. For each section in config_file, imports the class <type> from module <module>, and sets attribute self.<section> = <type>(**opts), where opts is a dict containing all of the options in <section> that are not <type> or <section>. For example, upon encountering the following entry in config_file.ini:

[adc] type = ADS1115 module = devices i2c_address = 0x48 i2c_bus = 1

The Hal will:

1) **Import pvp.io.devices.ADS1115 (or ADS1015) as a local variable:** class_ = getattr(import_module('.devices', 'pvp.io'), 'ADS1115')
2) **Instantiate an ADS1115 object with the arguments defined in config_file and set it as an attribute:** self._adc = class_(pig=self.-pig,address=0x48,i2c_bus=1)

Note: RawConfigParser.optionxform() is overloaded here s.t. options are case sensitive (they are by default case insensitive). This is necessary due to the kwarg MUX which is so named for consistency with the config registry documentation in the ADS1115 datasheet. For example, A P4vMini pressure_sensor on pin A0 (MUX=0) of the ADC is passed arguments like:

analog_sensor = AnalogSensor(pig=self._pig, adc=self._adc, MUX=0, offset_voltage=0.25, output_span = 4.0, conversion_factor=2.54*20

)

Note: ast.literal_eval(opt) interprets integers, 0xFF, (a, b) etc. correctly. It does not interpret strings correctly, nor does it know 'adc' -> self._adc; therefore, these special cases are explicitly handled.

Parameters **config_file** (*str*) – Path to the configuration file containing the definitions of specific components on the ventilator machine. (e.g., config_file = "pvp/io/config/devices.ini")

Methods:

__init__([config_file])	Initializes HAL from config_file.

Attributes:

pressure	Returns the pressure from the primary pressure sensor.
oxygen	Returns the oxygen concentration from the primary oxygen sensor.
aux_pressure	Returns the pressure from the auxiliary pressure sensor, if so equipped.
flow_in	The measured flow rate inspiratory side.
flow_ex	The measured flow rate expiratory side.
setpoint_in	The currently requested flow for the inspiratory proportional control valve as a proportion of maximum.
setpoint_ex	The currently requested flow on the expiratory side as a proportion of the maximum.

__init__(*config_file='pvp/io/config/devices.ini'*)

Initializes HAL from config_file. For each section in config_file, imports the class <type> from module <module>, and sets attribute self.<section> = <type>(**opts), where opts is a dict containing all of the options in <section> that are not <type> or <section>. For example, upon encountering the following entry in config_file.ini:

[adc] type = ADS1115 module = devices i2c_address = 0x48 i2c_bus = 1

The Hal will:

1) **Import pvp.io.devices.ADS1115 (or ADS1015) as a local variable: class_** = getattr(import_module('.devices', 'pvp.io'), 'ADS1115')

2) **Instantiate an ADS1115 object with the arguments defined in config_file and set it as an attribute:** self._adc = class_(pig=self.-pig,address=0x48,i2c_bus=1)

Note: RawConfigParser.optionxform() is overloaded here s.t. options are case sensitive (they are by default case insensitive). This is necessary due to the kwarg MUX which is so named for consistency with the config registry documentation in the ADS1115 datasheet. For example, A P4vMini pressure_sensor on pin A0 (MUX=0) of the ADC is passed arguments like:

analog_sensor = AnalogSensor(pig=self._pig, adc=self._adc, MUX=0, offset_voltage=0.25, output_span = 4.0, conversion_factor=2.54*20

)

Note: ast.literal_eval(opt) interprets integers, 0xFF, (a, b) etc. correctly. It does not interpret strings correctly, nor does it know 'adc' -> self._adc; therefore, these special cases are explicitly handled.

Parameters **config_file** (*str*) – Path to the configuration file containing the definitions of specific components on the ventilator machine. (e.g., config_file = "pvp/io/config/devices.ini")

property pressure: float
Returns the pressure from the primary pressure sensor.

property oxygen: float
Returns the oxygen concentration from the primary oxygen sensor.

property aux_pressure: float
Returns the pressure from the auxiliary pressure sensor, if so equipped. If a secondary pressure sensor is not defined, raises a RuntimeWarning.

property flow_in: float
The measured flow rate inspiratory side.

property flow_ex: float
The measured flow rate expiratory side.

property setpoint_in: float
The currently requested flow for the inspiratory proportional control valve as a proportion of maximum.

property setpoint_ex: float
The currently requested flow on the expiratory side as a proportion of the maximum.

1.1.16.2 devices

A module for ventilator hardware device drivers

1.1.17 Alarm

1.1.17.1 Alarm System Overview

- Alarms are represented as *Alarm* objects, which are created and managed by the *Alarm_Manager*.
- A collection of *Alarm_Rule* s define the *Condition* s for raising *Alarm* s of different *AlarmSeverity* .
- The alarm manager is continuously fed *SensorValues* objects during *PVP_Gui.update_gui()*, which it uses to *check()* each alarm rule.
- The alarm manager emits *Alarm* objects to the *PVP_Gui.handle_alarm()* method.
- The alarm manager also updates alarm thresholds set as `Condition.depends` to *PVP_Gui.limits_updated()* when control parameters are set (eg. updates the `HIGH_PRESSURE` alarm to be triggered 15% above some set `PIP`).

1.1.17.2 Alarm Modules

Alarm Manager

The alarm manager is responsible for checking the *Alarm_Rule* s and maintaining the *Alarm* s active in the system.

Only one instance of the *Alarm_Manager* can be created at once, and if it is instantiated again, the existing object will be returned.

Classes:

Alarm_Manager()	The Alarm Manager

class pvp.alarm.alarm_manager.**Alarm_Manager**

The Alarm Manager

The alarm manager receives *SensorValues* from the GUI via *Alarm_Manager.update()* and emits *Alarm* s to methods given by *Alarm_Manager.add_callback()* . When alarm limits are updated (ie. the *Alarm_Rule* has *depends*), it emits them to methods registered with *Alarm_Manager.add_dependency_callback()* .

On initialization, the alarm manager calls *Alarm_Manager.load_rules()* , which loads all rules defined in *alarm.ALARM_RULES* .

active_alarms

{*AlarmType*: *Alarm*}

Type dict

logged_alarms

A list of deactivated alarms.

Type list

dependencies

A dictionary mapping *ValueName* s to the alarm threshold dependencies they update

Type dict

pending_clears

[*AlarmType*] list of alarms that have been requested to be cleared

Type list

callbacks

list of callables that accept *Alarm* s when they are raised/altered.

Type list

cleared_alarms

of *AlarmType* s, alarms that have been cleared but have not dropped back into the 'off' range to enable re-raising

Type list

snoozed_alarms

of *AlarmType* s : times, alarms that should not be raised because they have been silenced for a period of time

Type dict

callbacks

list of callables to send *Alarm* objects to

Type list

depends_callbacks
When we *update_dependencies()*, we send back a *ControlSetting* with the new min/max

Type list

rules
A dict mapping *AlarmType* to *Alarm_Rule* .

Type dict

If an Alarm_Manager already exists, when initing just return that one

Attributes:

_instance
active_alarms
logged_alarms
dependencies
pending_clears
cleared_alarms
snoozed_alarms
callbacks
depends_callbacks
rules
logger

Methods:

load_rules()	Copy alarms from *alarm.ALARM_RULES* and call *Alarm_Manager.load_rule()* for each
load_rule(alarm_rule)	Add the Alarm Rule to *Alarm_Manager.rules* and register any dependencies they have with *Alarm_Manager.register_dependency()*
update(sensor_values)	Call *Alarm_Manager.check_rule()* for all rules in *Alarm_Manager.rules*
check_rule(rule, sensor_values)	*check()* the alarm rule, handle logic of raising, emitting, or lowering an alarm.
emit_alarm(alarm_type, severity)	Emit alarm (by calling all callbacks with it).
deactivate_alarm(alarm)	Mark an alarm's internal active flags and remove from *active_alarms*
dismiss_alarm(alarm_type[, duration])	GUI or other object requests an alarm to be dismissed & deactivated

continues on next page

Table 90 – continued from previous page

get_alarm_severity(alarm_type)	Get the severity of an Alarm
register_alarm(alarm)	Be given an already created alarm and emit to callbacks.
register_dependency(condition, dependency, ...)	Add dependency in a Condition object to be updated when values are changed
update_dependencies(control_setting)	Update Condition objects that update their value according to some control parameter
add_callback(callback)	Assert we're being given a callable and add it to our list of callbacks.
add_dependency_callback(callback)	Assert we're being given a callable and add it to our list of dependency_callbacks
clear_all_alarms()	call *Alarm_Manager.deactivate_alarm()* for all active alarms.
reset()	Reset all conditions, callbacks, and other stateful attributes and clear alarms

_instance = None

active_alarms: Dict[*pvp.alarm.AlarmType*, *pvp.alarm.alarm.Alarm*] = {}

logged_alarms: List[*pvp.alarm.alarm.Alarm*] = []

dependencies = {}

pending_clears = []

cleared_alarms = []

snoozed_alarms = {}

callbacks = []

depends_callbacks = []

rules = {}

logger = <Logger pvp.alarm.alarm_manager (WARNING)>

load_rules()
Copy alarms from *alarm.ALARM_RULES* and call *Alarm_Manager.load_rule()* for each

load_rule(*alarm_rule:* pvp.alarm.rule.Alarm_Rule)
Add the Alarm Rule to *Alarm_Manager.rules* and register any dependencies they have with *Alarm_Manager.register_dependency()*

Parameters **alarm_rule** (*Alarm_Rule*) – Alarm rule to be loaded

update(*sensor_values:* pvp.common.message.SensorValues)
Call *Alarm_Manager.check_rule()* for all rules in *Alarm_Manager.rules*

Parameters **sensor_values** (*SensorValues*) – New sensor values from the GUI

check_rule(*rule:* pvp.alarm.rule.Alarm_Rule, *sensor_values:* pvp.common.message.SensorValues)
check() the alarm rule, handle logic of raising, emitting, or lowering an alarm.

When alarms are dismissed, an *alarm.Alarm* is emitted with `AlarmSeverity.OFF` .

- If the alarm severity has increased, emit a new alarm.
- If the alarm severity has decreased and the alarm is not latched, emit a new alarm
- If the alarm severity has decreased and the alarm is latched, check if the alarm has been manually dismissed, if it has emit a new alarm.

- If a latched alarm has been manually dismissed previously and the alarm condition is now no longer met, dismiss the alarm.

Parameters

- **rule** (*Alarm_Rule*) – Alarm rule to check
- **sensor_values** (*SensorValues*) – sent by the GUI to check against alarm rule

emit_alarm(*alarm_type:* pvp.alarm.AlarmType, *severity:* pvp.alarm.AlarmSeverity)
Emit alarm (by calling all callbacks with it).

Note: This method emits *and* clears alarms – a cleared alarm is emitted with `AlarmSeverity.OFF`

Parameters

- **alarm_type** (*AlarmType*) –
- **severity** (*AlarmSeverity*) –

deactivate_alarm(*alarm: (<enum 'AlarmType'>, <class 'pvp.alarm.alarm.Alarm'>)*)
Mark an alarm's internal active flags and remove from *active_alarms*

Typically called internally when an alarm is being replaced by one of the same type but a different severity.

Note: This does *not* alert listeners that an alarm has been cleared, for that emit an alarm with AlarmSeverity.OFF

Parameters **alarm** (*AlarmType* , *Alarm*) – Alarm to deactivate

dismiss_alarm(*alarm_type:* pvp.alarm.AlarmType, *duration:* *Optional[float]* = *None*)
GUI or other object requests an alarm to be dismissed & deactivated

GUI will wait until it receives an *emit_alarm* of severity == OFF to remove alarm widgets. If the alarm is not latched

If the alarm is latched, alarm_manager will not decrement alarm severity or emit *OFF* until a) the condition returns to *OFF*, and b) the user dismisses the alarm

Parameters

- **alarm_type** (*AlarmType*) – Alarm to dismiss
- **duration** (*float*) – seconds - amount of time to wait before alarm can be re-raised If a duration is provided, the alarm will not be able to be re-raised

get_alarm_severity(*alarm_type:* pvp.alarm.AlarmType)
Get the severity of an Alarm

Parameters **alarm_type** (*AlarmType*) – Alarm type to check

Returns *AlarmSeverity*

register_alarm(*alarm:* pvp.alarm.alarm.Alarm)
Be given an already created alarm and emit to callbacks.

Mostly used during testing for programmatically created alarms. Creating alarms outside of the Alarm_Manager is generally discouraged.

Parameters **alarm** (*Alarm*) –

register_dependency(*condition:* pvp.alarm.condition.Condition, *dependency: dict*, *severity:* pvp.alarm.AlarmSeverity)

Add dependency in a Condition object to be updated when values are changed

Parameters

- **condition** (*dict*) – Condition as defined in an *Alarm_Rule*
- **dependency** (*dict*) – either a (ValueName, attribute_name) or optionally also + transformation callable
- **severity** (*AlarmSeverity*) – severity of dependency

update_dependencies(*control_setting:* pvp.common.message.ControlSetting)

Update Condition objects that update their value according to some control parameter

Call any `transform` functions on the attribute of the control setting specified in the depencency.

Emit another *ControlSetting* describing the new max or min or the value.

Parameters **control_setting** (*ControlSetting*) – Control setting that was changed

add_callback(*callback: Callable*)

Assert we're being given a callable and add it to our list of callbacks.

Parameters **callback** (*Callable*) – Callback that accepts a single argument of an *Alarm*

add_dependency_callback(*callback: Callable*)

Assert we're being given a callable and add it to our list of dependency_callbacks

Parameters **callback** (*Callable*) – Callback that accepts a *ControlSetting*

Returns:

clear_all_alarms()

call *Alarm_Manager.deactivate_alarm()* for all active alarms.

reset()

Reset all conditions, callbacks, and other stateful attributes and clear alarms

Alarm Objects

Alarm objects represent the state and severity of active alarms, but are otherwise intentionally quite featureless.

They are created and maintained by the *Alarm_Manager* and sent to any listeners registered in `Alarm_Manager.callbacks` .

Classes:

Alarm(alarm_type, severity[, start_time, ...])	Representation of alarm status and parameters

class pvp.alarm.alarm.**Alarm**(*alarm_type:* pvp.alarm.AlarmType, *severity:* pvp.alarm.AlarmSeverity, *start_time: Optional[float] = None*, *latch: bool = True*, *cause: Optional[list] = None*, *value=None*, *message=None*)

Representation of alarm status and parameters

Parameterized by a `Alarm_Rule` and managed by `Alarm_Manager`

Parameters

- **alarm_type** (*AlarmType*) – Type of alarm

- **severity** (*AlarmSeverity*) – Severity of alarm
- **start_time** (*float*) – Timestamp of alarm start, (as generated by `time.time()`
- **cause** (`ValueName`) – The *ValueName* that caused the alarm to be fired
- **value** (*int, float*) – optional - numerical value that generated the alarm
- **message** (*str*) – optional - override default text generated by `AlarmManager`

id
unique alarm ID

Type int

end_time
If None, alarm has not ended. otherwise timestamp

Type None, float

active
Whether or not the alarm is currently active

Type bool

Attributes:

id_counter	used to generate unique IDs for each alarm
severity	Alarm Severity, property without setter to prevent change after instantiation
alarm_type	Alarm Type, property without setter to prevent change after instantiation

Methods:

__init__(alarm_type, severity[, start_time, ...])	**param alarm_type** Type of alarm
deactivate()	If active, register an end time and set as `active == False` Returns:

id_counter = count(0)
used to generate unique IDs for each alarm

Type `itertools.count`

__init__(*alarm_type:* pvp.alarm.AlarmType, *severity:* pvp.alarm.AlarmSeverity, *start_time: Optional[float] = None*, *latch: bool = True*, *cause: Optional[list] = None*, *value=None*, *message=None*)

Parameters

- **alarm_type** (*AlarmType*) – Type of alarm
- **severity** (*AlarmSeverity*) – Severity of alarm
- **start_time** (*float*) – Timestamp of alarm start, (as generated by `time.time()`
- **cause** (`ValueName`) – The *ValueName* that caused the alarm to be fired
- **value** (*int, float*) – optional - numerical value that generated the alarm
- **message** (*str*) – optional - override default text generated by `AlarmManager`

id
unique alarm ID
Type int

end_time
If None, alarm has not ended. otherwise timestamp
Type None, float

active
Whether or not the alarm is currently active
Type bool

property severity: *pvp.alarm.AlarmSeverity*
Alarm Severity, property without setter to prevent change after instantiation

Returns *AlarmSeverity*

property alarm_type: *pvp.alarm.AlarmType*
Alarm Type, property without setter to prevent change after instantiation

Returns *AlarmType*

deactivate()
If active, register an end time and set as `active == False` Returns:

Alarm Rule

One *Alarm_Rule* is defined for each *AlarmType* in *ALARM_RULES*.

An alarm rule defines:

- The conditions for raising different severities of an alarm
- The dependencies between set values and alarm thresholds
- The behavior of the alarm, specifically whether it is `latch` ed.

Example

As an example, we'll define a `LOW_PRESSURE` alarm with escalating severity. A `LOW` severity alarm will be raised when measured `PIP` falls 10% below set `PIP`, which will escalate to a `MEDIUM` severity alarm if measured `PIP` falls 15% below set `PIP` **and** the `LOW` severity alarm has been active for at least two breath cycles.

First we define the name and behavior of the alarm:

```
Alarm_Rule(
    name = AlarmType.LOW_PRESSURE,
    latch = False,
```

In this case, `latch == False` means that the alarm will disappear (or be downgraded in severity) whenever the conditions for that alarm are no longer met. If `latch == True`, an alarm requires manual dismissal before it is downgraded or disappears.

Next we'll define a tuple of *Condition* objects for `LOW` and `MEDIUM` severity objects.

Starting with the `LOW` severity alarm:

```
conditions = (
    (
    AlarmSeverity.LOW,
    condition.ValueCondition(
        value_name=ValueName.PIP,
        limit=VALUES[ValueName.PIP]['safe_range'][0],
        mode='min',
        depends={
            'value_name': ValueName.PIP,
            'value_attr': 'value',
            'condition_attr': 'limit',
            'transform': lambda x : x-(x*0.10)
        })
    ),
    # ... continued in next block
```

Each condition is a tuple of an (*AlarmSeverity*, *Condition*). In this case, we use a *ValueCondition* which tests whether a value is above or below a set `'max'` or `'min'`, respectively. For the low severity `LOW_PRESSURE` alarm, we test if `ValueName.PIP` is below (`mode='min'`) some `limit`, which is initialized as the low-end of `PIP`'s safe range.

We also define a condition for updating the `'limit'` of the condition (`'condition_attr' : 'limit'`), from the `ControlSetting.value`` field whenever `PIP` is updated. Specifically, we set the `limit` to be 10% less than the set `PIP` value by 10% with a lambda function (`lambda x : x-(x*0.10)`).

Next, we define the `MEDIUM` severity alarm condition:

```
(
AlarmSeverity.MEDIUM,
condition.ValueCondition(
    value_name=ValueName.PIP,
    limit=VALUES[ValueName.PIP]['safe_range'][0],
    mode='min'
    depends={
        'value_name': ValueName.PIP,
        'value_attr': 'value',
        'condition_attr': 'limit',
        'transform': lambda x: x - (x * 0.15)
    },
) + \
condition.CycleAlarmSeverityCondition(
    alarm_type = AlarmType.LOW_PRESSURE,
    severity   = AlarmSeverity.LOW,
    n_cycles = 2
))
```

The first `ValueCondition` is the same as in the `LOW` alarm severity condition, except that it is set 15% below `PIP`.

A second *CycleAlarmSeverityCondition* has been added (with +) to the *ValueCondition* When conditions are added together, they will only return `True` (ie. trigger an alarm) if all of the conditions are met. This condition checks that the `LOW_PRESSURE` alarm has been active at a `LOW` severity for at least two cycles.

Full source for this example and all alarm rules can be found here

Module Documentation

Class to declare alarm rules

Classes:

Alarm_Rule(name, conditions[, latch, technical])	• name of rule

class pvp.alarm.rule.**Alarm_Rule**(*name:* pvp.alarm.AlarmType, *conditions*, *latch=True*, *technical=False*)

- name of rule
- conditions: ((alarm_type, (condition_1, condition_2)), ...)
- latch (bool): if True, alarm severity cannot be decremented until user manually dismisses
- silencing/overriding rules

Methods:

check(sensor_values)	Check all of our `conditions` .
reset()	

Attributes:

severity	Last Alarm Severity from `.check()` :returns: *AlarmSeverity*
depends	Get all ValueNames whose alarm limits depend on this alarm rule :returns: list[ValueName]
value_names	Get all ValueNames specified as value_names in alarm conditions

check(*sensor_values*)
Check all of our `conditions` .

Parameters **sensor_values** –

Returns:

property severity
Last Alarm Severity from `.check()` :returns: *AlarmSeverity*

reset()

property depends: List[*pvp.common.values.ValueName*]
Get all ValueNames whose alarm limits depend on this alarm rule :returns: list[ValueName]

property value_names: List[*pvp.common.values.ValueName*]
Get all ValueNames specified as value_names in alarm conditions

Returns list[ValueName]

Alarm Condition

Condition objects define conditions that can raise alarms. They are used by *Alarm_Rule* s.

Each has to define a *Condition.check()* method that accepts *SensorValues* . The method should return `True` if the alarm condition is met, and `False` otherwise.

Conditions can be added (+) together to make compound conditions, and a single call to `check` will only return true if both conditions return true. If any condition in the chain returns false, evaluation is stopped and the alarm is not raised.

Conditions can

Functions:

get_alarm_manager()	

Classes:

Condition([depends])	Base class for specifying alarm test conditions
ValueCondition(value_name, limit, mode, ...)	Value is greater or lesser than some max/min
CycleValueCondition(n_cycles, *args, **kwargs)	Value goes out of range for a specific number of breath cycles
TimeValueCondition(time, *args, **kwargs)	value goes out of range for specific amount of time
AlarmSeverityCondition(alarm_type, severity)	Alarm is above or below a certain severity.
CycleAlarmSeverityCondition(n_cycles, *args, ...)	alarm goes out of range for a specific number of breath cycles

pvp.alarm.condition.**get_alarm_manager**()

class pvp.alarm.condition.**Condition**(*depends: Optional[dict] = None*, **args*, ***kwargs*)

Bases: `object`

Base class for specifying alarm test conditions

Subclasses must define *Condition.check()* and `Conditino.reset()`

Condition objects can be added together to create compound conditions.

_child

if another condition is added to this one, store a reference to it

Type *Condition*

Parameters

- **depends** (*list*, *dict*) – a list of, or a single dict:

```
{'value_name':ValueName,
'value_attr': attr in ControlMessage,
 'condition_attr',
 optional: transformation: callable)
that declare what values are needed to update
```

- ***args** –
- ****kwargs** –

Methods:

__init__([depends])	**param depends**
check(sensor_values)	Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean
reset()	If a condition is stateful, need to provide some method of resetting the state

Attributes:

manager	The active alarm manager, used to get status of alarms

__init__(*depends: Optional[dict] = None*, **args*, ***kwargs*)

Parameters

- **depends** (*list*, *dict*) – a list of, or a single dict:

```
{'value_name':ValueName,
'value_attr': attr in ControlMessage,
 'condition_attr',
 optional: transformation: callable)
that declare what values are needed to update
```

- ***args** –
- ****kwargs** –

property manager
The active alarm manager, used to get status of alarms

Returns *pvp.alarm.alarm_manager.Alarm_Manager*

check(*sensor_values*) → bool
Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean

Parameters **sensor_values** (*SensorValues*) – SensorValues used to compute alarm status

Returns bool

reset()
If a condition is stateful, need to provide some method of resetting the state

class pvp.alarm.condition.**ValueCondition**(*value_name: pvp.common.values.ValueName*, *limit: (<class 'int'>, <class 'float'>)*, *mode: str*, **args*, ***kwargs*)

Bases: *pvp.alarm.condition.Condition*

Value is greater or lesser than some max/min

Parameters

- **value_name** (ValueName) – Which value to check
- **limit** (*int*, *float*) – value to check against
- **mode** (*'min'*, *'max'*) – whether the limit is a minimum or maximum
- ***args** –

- ****kwargs** –

operator
Either the less than or greater than operators, depending on whether mode is `'min'` or `'max'`

Type callable

Methods:

__init__(value_name, limit, mode, *args, ...)	**param value_name** Which value to check
check(sensor_values)	Check that the relevant value in SensorValues is either greater or lesser than the limit
reset()	not stateful, do nothing.

Attributes:

mode	One of 'min' or 'max', defines how the incoming sensor values are compared to the set value

__init__(*value_name: pvp.common.values.ValueName*, *limit: (<class 'int'>, <class 'float'>)*, *mode: str*, **args*, ***kwargs*)

Parameters

- **value_name** (*ValueName*) – Which value to check
- **limit** (*int, float*) – value to check against
- **mode** (*'min', 'max'*) – whether the limit is a minimum or maximum
- ***args** –
- ****kwargs** –

operator
Either the less than or greater than operators, depending on whether mode is `'min'` or `'max'`
Type callable

property mode
One of 'min' or 'max', defines how the incoming sensor values are compared to the set value

Returns:

check(*sensor_values*)
Check that the relevant value in SensorValues is either greater or lesser than the limit

Parameters **sensor_values** (*SensorValues*) –

Returns bool

reset()
not stateful, do nothing.

class pvp.alarm.condition.**CycleValueCondition**(*n_cycles: int*, **args*, ***kwargs*)
Bases: *pvp.alarm.condition.ValueCondition*

Value goes out of range for a specific number of breath cycles

Parameters **n_cycles** (*int*) – number of cycles required

_start_cycle
The breath cycle where the

Type int

_mid_check
whether a value has left the acceptable range and we are counting consecutive breath cycles

Type bool

Parameters

- **value_name** (ValueName) – Which value to check
- **limit** (*int*, *float*) – value to check against
- **mode** (*'min'*, *'max'*) – whether the limit is a minimum or maximum
- ***args** –
- ****kwargs** –

operator
Either the less than or greater than operators, depending on whether mode is `'min'` or `'max'`

Type callable

Attributes:

n_cycles	Number of cycles required

Methods:

check(sensor_values)	Check if outside of range, and then check if number of breath cycles have elapsed
reset()	Reset check status and start cycle

property n_cycles: int
Number of cycles required

check(*sensor_values*) → bool
Check if outside of range, and then check if number of breath cycles have elapsed

Parameters **()** (*sensor_values*) –

Returns bool

reset()
Reset check status and start cycle

class pvp.alarm.condition.**TimeValueCondition**(*time*, **args*, ***kwargs*)
Bases: *pvp.alarm.condition.ValueCondition*

value goes out of range for specific amount of time

Warning: Not implemented!

Parameters

- **time** (*float*) – number of seconds value must be out of range
- ***args** –
- ****kwargs** –

Methods:

__init__(time, *args, **kwargs)	**param time** number of seconds value must be out of range
check(sensor_values)	Check that the relevant value in SensorValues is either greater or lesser than the limit
reset()	not stateful, do nothing.

__init__(*time*, **args*, ***kwargs*)

Parameters

- **time** (*float*) – number of seconds value must be out of range
- ***args** –
- ****kwargs** –

check(*sensor_values*)

Check that the relevant value in SensorValues is either greater or lesser than the limit

Parameters **sensor_values** (*SensorValues*) –

Returns bool

reset()

not stateful, do nothing.

class pvp.alarm.condition.**AlarmSeverityCondition**(*alarm_type:* pvp.alarm.AlarmType, *severity:* pvp.alarm.AlarmSeverity, *mode:* *str* = *'min'*, **args*, ***kwargs*)

Bases: *pvp.alarm.condition.Condition*

Alarm is above or below a certain severity.

Get alarm severity status from *Alarm_Manager.get_alarm_severity()* .

Parameters

- **alarm_type** (*AlarmType*) – Alarm type to check
- **severity** (*AlarmSeverity*) – Alarm severity to check against
- **mode** (*str*) – one of 'min', 'equals', or 'max'. 'min' returns true if the alarm is at least this value (note the difference from ValueCondition which returns true if the alarm is less than..) and vice versa for 'max'.

  Note: 'min' and 'max' use >= and <= rather than > and <

- ***args** –
- ****kwargs** –

Methods:

__init__(alarm_type, severity[, mode])	Alarm is above or below a certain severity.
check([sensor_values])	Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean
reset()	If a condition is stateful, need to provide some method of resetting the state

Attributes:

mode	'min' returns true if the alarm is at least this value (note the difference from ValueCondition which returns true if the alarm is less than..) and vice versa for 'max'.

__init__(*alarm_type:* pvp.alarm.AlarmType, *severity:* pvp.alarm.AlarmSeverity, *mode:* *str* = *'min'*, **args*, ***kwargs*)

Alarm is above or below a certain severity.

Get alarm severity status from *Alarm_Manager.get_alarm_severity()* .

Parameters

- **alarm_type** (*AlarmType*) – Alarm type to check
- **severity** (*AlarmSeverity*) – Alarm severity to check against
- **mode** (*str*) – one of 'min', 'equals', or 'max'. 'min' returns true if the alarm is at least this value (note the difference from ValueCondition which returns true if the alarm is less than..) and vice versa for 'max'.

  Note: 'min' and 'max' use >= and <= rather than > and <

- ***args** –
- ****kwargs** –

property mode: str

'min' returns true if the alarm is at least this value (note the difference from ValueCondition which returns true if the alarm is less than..) and vice versa for 'max'.

Note: 'min' and 'max' use >= and <= rather than > and <

Returns one of 'min', 'equals', or 'max'.

Return type str

check(*sensor_values=None*)

Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean

Parameters **sensor_values** (*SensorValues*) – SensorValues used to compute alarm status

Returns bool

reset()

If a condition is stateful, need to provide some method of resetting the state

class pvp.alarm.condition.**CycleAlarmSeverityCondition**(*n_cycles*, **args*, ***kwargs*)

Bases: *pvp.alarm.condition.AlarmSeverityCondition*

alarm goes out of range for a specific number of breath cycles

Todo: note that this is exactly the same as CycleValueCondition. Need to do the multiple inheritance thing

_start_cycle
The breath cycle where the

Type int

_mid_check
whether a value has left the acceptable range and we are counting consecutive breath cycles

Type bool

Alarm is above or below a certain severity.

Get alarm severity status from *Alarm_Manager.get_alarm_severity()* .

Parameters

- **alarm_type** (*AlarmType*) – Alarm type to check
- **severity** (*AlarmSeverity*) – Alarm severity to check against
- **mode** (*str*) – one of 'min', 'equals', or 'max'. 'min' returns true if the alarm is at least this value (note the difference from ValueCondition which returns true if the alarm is less than..) and vice versa for 'max'.

  Note: 'min' and 'max' use >= and <= rather than > and <

- ***args** –
- ****kwargs** –

Attributes:

n_cycles

Methods:

check(sensor_values)	Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean
reset()	If a condition is stateful, need to provide some method of resetting the state

property n_cycles

check(*sensor_values*)
Every Condition subclass needs to define this method that accepts *SensorValues* and returns a boolean

Parameters **sensor_values** (*SensorValues*) – SensorValues used to compute alarm status

Returns bool

reset()
If a condition is stateful, need to provide some method of resetting the state

1.1.17.3 Main Alarm Module

Classes:

AlarmType(value)	An enumeration.
AlarmSeverity(value)	An enumeration.

Data:

ALARM_RULES	Definitions of all *Alarm_Rule* s used by the *Alarm_Manager*

class pvp.alarm.**AlarmType**(*value*)
An enumeration.

Attributes:

LOW_PRESSURE	
HIGH_PRESSURE	
LOW_VTE	
HIGH_VTE	
LOW_PEEP	
HIGH_PEEP	
LOW_O2	
HIGH_O2	
OBSTRUCTION	
LEAK	
SENSORS_STUCK	
BAD_SENSOR_READINGS	
MISSED_HEARTBEAT	
human_name	Replace .name underscores with spaces

LOW_PRESSURE = 1

HIGH_PRESSURE = 2

LOW_VTE = 3

HIGH_VTE = 4

LOW_PEEP = 5

HIGH_PEEP = 6

LOW_O2 = 7

HIGH_O2 = 8

OBSTRUCTION = 9

LEAK = 10

SENSORS_STUCK = 11

BAD_SENSOR_READINGS = 12

MISSED_HEARTBEAT = 13

property human_name: str
Replace .name underscores with spaces

class pvp.alarm.**AlarmSeverity**(*value*)
An enumeration.

Attributes:

HIGH
MEDIUM
LOW
OFF
TECHNICAL

HIGH = 3

MEDIUM = 2

LOW = 1

OFF = 0

TECHNICAL = -1

pvp.alarm.**ALARM_RULES = OrderedDict([(<AlarmType.LOW_PRESSURE: 1>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.HIGH_PRESSURE: 2>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.LOW_VTE: 3>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.HIGH_VTE: 4>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.LOW_PEEP: 5>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.HIGH_PEEP: 6>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.LOW_O2: 7>, <pvp.alarm.rule.Alarm_Rule object>), (<AlarmType.HIGH_O2: 8>, <pvp.alarm.rule.Alarm_Rule object>)])**
Definitions of all *Alarm_Rule* s used by the *Alarm_Manager*

See definitions here

1.1.18 coordinator module

The coordinator provides an interface between the process threads, and facilitates inter-process communication. It is a wrapper around xml-rpc, which allowed us to use defined data-structures such as *SensorValues*.

1.1.18.1 Submodules

1.1.18.2 coordinator

Classes:

CoordinatorBase([sim_mode])	
CoordinatorLocal([sim_mode])	**param sim_mode**
CoordinatorRemote([sim_mode])	

Functions:

get_coordinator([single_process, sim_mode])	

class pvp.coordinator.coordinator.**CoordinatorBase**(*sim_mode=False*)

Bases: object

Methods:

get_sensors()	
get_alarms()	
set_control(control_setting)	
get_control(control_setting_name)	
set_breath_detection(breath_detection)	
get_breath_detection()	
start()	
is_running()	
kill()	
stop()	

get_sensors() → *pvp.common.message.SensorValues*

get_alarms() → Union[None, Tuple[*pvp.alarm.alarm.Alarm*]]

set_control(*control_setting:* pvp.common.message.ControlSetting)

get_control(*control_setting_name:* pvp.common.values.ValueName) → *pvp.common.message.ControlSetting*

set_breath_detection(*breath_detection: bool*)

get_breath_detection() → bool

start()

is_running() → bool

kill()

stop()

class pvp.coordinator.coordinator.**CoordinatorLocal**(*sim_mode=False*)

Bases: *pvp.coordinator.coordinator.CoordinatorBase*

Parameters **sim_mode** –

_is_running

.set() when thread should stop

Type threading.Event

Methods:

__init__([sim_mode])	**param sim_mode**
get_sensors()	
get_alarms()	
set_control(control_setting)	
get_control(control_setting_name)	
set_breath_detection(breath_detection)	
get_breath_detection()	
start()	Start the coordinator.
is_running()	Test whether the whole system is running
stop()	Stop the coordinator.
kill()	

__init__(*sim_mode=False*)

Parameters **sim_mode** –

_is_running

.set() when thread should stop

Type threading.Event

get_sensors() → *pvp.common.message.SensorValues*

get_alarms() → Union[None, Tuple[*pvp.alarm.alarm.Alarm*]]

set_control(*control_setting:* pvp.common.message.ControlSetting)

get_control(*control_setting_name:* pvp.common.values.ValueName) → *pvp.common.message.ControlSetting*

set_breath_detection(*breath_detection: bool*)

get_breath_detection() → bool

start()
Start the coordinator. This does a soft start (not allocating a process).

is_running() → bool
Test whether the whole system is running

stop()
Stop the coordinator. This does a soft stop (not kill a process)

kill()

class pvp.coordinator.coordinator.**CoordinatorRemote**(*sim_mode=False*)
Bases: *pvp.coordinator.coordinator.CoordinatorBase*

Methods:

get_sensors()
get_alarms()
set_control(control_setting)
get_control(control_setting_name)
set_breath_detection(breath_detection)
get_breath_detection()

start()	Start the coordinator.
is_running()	Test whether the whole system is running
stop()	Stop the coordinator.
kill()	Stop the coordinator and end the whole program

get_sensors() → *pvp.common.message.SensorValues*

get_alarms() → Union[None, Tuple[*pvp.alarm.alarm.Alarm*]]

set_control(*control_setting:* pvp.common.message.ControlSetting)

get_control(*control_setting_name:* pvp.common.values.ValueName) → *pvp.common.message.ControlSetting*

set_breath_detection(*breath_detection: bool*)

get_breath_detection() → bool

start()
Start the coordinator. This does a soft start (not allocating a process).

is_running() → bool
Test whether the whole system is running

stop()
Stop the coordinator. This does a soft stop (not kill a process)

kill()
Stop the coordinator and end the whole program

pvp.coordinator.coordinator.**get_coordinator**(*single_process=False*, *sim_mode=False*) → *pvp.coordinator.coordinator.CoordinatorBase*

1.1.18.3 ipc

Functions:

get_sensors()	
get_alarms()	
set_control(control_setting)	
get_control(control_setting_name)	
set_breath_detection(breath_detection)	
get_breath_detection()	
rpc_server_main(sim_mode, serve_event[, ...])	
get_rpc_client()	

pvp.coordinator.rpc.**get_sensors**()

pvp.coordinator.rpc.**get_alarms**()

pvp.coordinator.rpc.**set_control**(*control_setting*)

pvp.coordinator.rpc.**get_control**(*control_setting_name*)

pvp.coordinator.rpc.**set_breath_detection**(*breath_detection*)

pvp.coordinator.rpc.**get_breath_detection**()

pvp.coordinator.rpc.**rpc_server_main**(*sim_mode*, *serve_event*, *addr='localhost'*, *port=9533*)

pvp.coordinator.rpc.**get_rpc_client**()

1.1.18.4 process_manager

Classes:

ProcessManager(sim_mode[, ...])	startCommandLine,

class pvp.coordinator.process_manager.**ProcessManager**(*sim_mode*, *startCommandLine=None*, *maxHeartbeatInterval=None*)

Bases: object

Methods:

start_process()	
try_stop_process()	
restart_process()	
heartbeat(timestamp)	

start_process()

try_stop_process()

restart_process()

heartbeat(*timestamp*)

1.1.19 Index

- genindex
- modindex

CHAPTER

TWO

MEDICAL DISCLAIMER

PVP1 is not a regulated or clinically validated medical device. We have not yet performed testing for safety or efficacy on living organisms. All material described herein should be used at your own risk and do not represent a medical recommendation. PVP1 is currently recommended only for research purposes.

This website is not connected to, endorsed by, or representative of the view of Princeton University. Neither the authors nor Princeton University assume any liability or responsibility for any consequences, damages, or loss caused or alleged to be caused directly or indirectly for any action or inaction taken based on or made in reliance on the information or material discussed herein or linked to from this website.

PVP1 is under continuous development and the information here may not be up to date, nor is any guarantee made as such. Neither the authors nor Princeton University are liable for any damage or loss related to the accuracy, completeness or timeliness of any information described or linked to from this website.

By continuing to watch or read this, you are acknowledging and accepting this disclaimer.

PYTHON MODULE INDEX

p

pvp.alarm, 188
pvp.alarm.alarm, 176
pvp.alarm.alarm_manager, 172
pvp.alarm.condition, 181
pvp.alarm.rule, 180
pvp.common.fashion, 169
pvp.common.loggers, 161
pvp.common.message, 158
pvp.common.prefs, 165
pvp.common.unit_conversion, 167
pvp.common.utils, 168
pvp.common.values, 152
pvp.controller.control_module, 144
pvp.coordinator.coordinator, 190
pvp.coordinator.process_manager, 194
pvp.coordinator.rpc, 193
pvp.gui.main, 110
pvp.gui.styles, 141
pvp.gui.widgets.alarm_bar, 123
pvp.gui.widgets.components, 136
pvp.gui.widgets.control_panel, 118
pvp.gui.widgets.dialog, 141
pvp.gui.widgets.display, 128
pvp.gui.widgets.plot, 133
pvp.io, 169
pvp.io.devices, 171
pvp.io.hal, 169

INDEX

Symbols

_DEFAULTS (*in module pvp.common.prefs*), 165
_DIRECTORIES (*in module pvp.common.prefs*), 165
_LOCK (*in module pvp.common.prefs*), 165
_LOGGER (*in module pvp.common.prefs*), 165
_LOGGERS (*in module pvp.common.loggers*), 161
_PID_update() (*pvp.controller.control_module.ControlModuleBase method*), 148
_PREFS (*in module pvp.common.prefs*), 165
_PREF_MANAGER (*in module pvp.common.prefs*), 165
__SimulatedPropValve() (*pvp.controller.control_module.ControlModuleSimulator method*), 152
__SimulatedSolenoid() (*pvp.controller.control_module.ControlModuleSimulator method*), 152
__analyze_last_waveform() (*pvp.controller.control_module.ControlModuleBase method*), 146
__calculate_control_signal_in() (*pvp.controller.control_module.ControlModuleBase method*), 147
__comptest() (*pvp.controller.control_module.ControlModuleBase method*), 146
__get_PID_error() (*pvp.controller.control_module.ControlModuleBase method*), 147
__get_hal() (*pvp.controller.control_module.ControlModuleDevice method*), 150
__init__() (*pvp.alarm.alarm.Alarm method*), 177
__init__() (*pvp.alarm.condition.AlarmSeverityCondition method*), 186
__init__() (*pvp.alarm.condition.Condition method*), 182
__init__() (*pvp.alarm.condition.TimeValueCondition method*), 185
__init__() (*pvp.alarm.condition.ValueCondition method*), 183
__init__() (*pvp.common.loggers.DataLogger method*), 163
__init__() (*pvp.common.message.ControlSetting method*), 160
__init__() (*pvp.common.message.SensorValues method*), 159
__init__() (*pvp.common.values.Value method*), 155
__init__() (*pvp.controller.control_module.ControlModuleBase method*), 146
__init__() (*pvp.controller.control_module.ControlModuleDevice method*), 149
__init__() (*pvp.controller.control_module.ControlModuleSimulator method*), 151
__init__() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 191
__init__() (*pvp.gui.widgets.components.OnOffButton method*), 140
__init__() (*pvp.gui.widgets.control_panel.StopWatch method*), 122
__init__() (*pvp.io.hal.Hal method*), 170
__save_values() (*pvp.controller.control_module.ControlModuleBase method*), 148
_start_new_breathcycle() (*pvp.controller.control_module.ControlModuleBase method*), 148
_test_for_alarms() (*pvp.controller.control_module.ControlModuleBase method*), 148
_changing_track (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
_child (*pvp.alarm.condition.Condition attribute*), 181
_control_reset() (*pvp.controller.control_module.ControlModuleBase method*), 148
_controls_from_COPY() (*pvp.controller.control_module.ControlModuleBase method*), 146
_dismiss() (*pvp.gui.widgets.alarm_bar.Alarm_Card method*), 125
_get_HAL() (*pvp.controller.control_module.ControlModuleDevice method*), 150
_get_control_signal_in() (*pvp.controller.control_module.ControlModuleBase method*), 147
_get_control_signal_out() (*pvp.controller.control_module.ControlModuleBase method*), 148
_heartbeat() (*pvp.gui.widgets.control_panel.HeartBeat*

method), 122
_increment_timer (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
_initialize_set_to_COPY() (*pvp.controller.control_module.ControlModuleBase method*), 146
_instance (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 174
_is_running (*pvp.coordinator.coordinator.CoordinatorLocal attribute*), 191
_last_heartbeat (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 120
_maximum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
_mid_check (*pvp.alarm.condition.CycleAlarmSeverityCondition attribute*), 187
_mid_check (*pvp.alarm.condition.CycleValueCondition attribute*), 184
_minimum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
_open_logfile() (*pvp.common.loggers.DataLogger method*), 163
_pressure_units_changed() (*pvp.gui.widgets.control_panel.Control_Panel method*), 119
_reset() (*pvp.controller.control_module.Balloon_Simulator method*), 151
_screenshot() (*pvp.gui.main.PVP_Gui method*), 117
_sensor_to_COPY() (*pvp.controller.control_module.ControlModuleBase method*), 146
_sensor_to_COPY() (*pvp.controller.control_module.ControlModuleDevice method*), 150
_sensor_to_COPY() (*pvp.controller.control_module.ControlModuleSimulator method*), 152
_set_HAL() (*pvp.controller.control_module.ControlModuleDevice method*), 150
_set_cycle_control() (*pvp.gui.main.PVP_Gui method*), 116
_singleStep() (*pvp.gui.widgets.components.DoubleSlider method*), 138
_start_cycle (*pvp.alarm.condition.CycleAlarmSeverityCondition attribute*), 187
_start_cycle (*pvp.alarm.condition.CycleValueCondition attribute*), 184
_start_mainloop() (*pvp.controller.control_module.ControlModuleBase method*), 149
_start_mainloop() (*pvp.controller.control_module.ControlModuleDevice method*), 150
_start_mainloop() (*pvp.controller.control_module.ControlModuleSimulator method*), 152
_state (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 120
_style (*pvp.gui.widgets.display.Display.self attribute*), 129
_update_time() (*pvp.gui.widgets.control_panel.StopWatch method*), 123
_value_changed() (*pvp.gui.widgets.display.Display method*), 131

A

abs_range (*pvp.common.values.Value property*), 156
abs_range (*pvp.gui.widgets.display.Display.self attribute*), 128
active (*pvp.alarm.alarm.Alarm attribute*), 177, 178
active_alarms (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
add_alarm() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
add_callback() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
add_dependency_callback() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
additional_values (*pvp.common.message.SensorValues attribute*), 159
Alarm (*class in pvp.alarm.alarm*), 176
alarm (*pvp.gui.widgets.alarm_bar.Alarm_Card attribute*), 125
Alarm_Bar (*class in pvp.gui.widgets.alarm_bar*), 123
Alarm_Card (*class in pvp.gui.widgets.alarm_bar*), 125
alarm_cards (*pvp.gui.widgets.alarm_bar.Alarm_Bar attribute*), 123
alarm_level (*pvp.gui.widgets.alarm_bar.Alarm_Bar property*), 125
Alarm_Manager (*class in pvp.alarm.alarm_manager*), 172
alarm_manager (*pvp.gui.main.PVP_Gui attribute*), 111
alarm_range (*pvp.gui.widgets.display.Display.self attribute*), 128
Alarm_Rule (*class in pvp.alarm.rule*), 180
ALARM_RULES (*in module pvp.alarm*), 189
Alarm_Sound_Player (*class in pvp.gui.widgets.alarm_bar*), 126
alarm_state (*pvp.gui.widgets.display.Display property*), 132
alarm_type (*pvp.alarm.alarm.Alarm property*), 178
alarms (*pvp.gui.widgets.alarm_bar.Alarm_Bar attribute*), 123
AlarmSeverity (*class in pvp.alarm*), 189
AlarmSeverityCondition (*class in pvp.alarm.condition*), 185
AlarmType (*class in pvp.alarm*), 188
aux_pressure (*pvp.io.hal.Hal property*), 171

B

BAD_SENSOR_READINGS (*pvp.alarm.AlarmType attribute*), 189

Balloon_Simulator (*class in pvp.controller.control_module*), 150
beatheart() (*pvp.gui.widgets.control_panel.HeartBeat method*), 122
BREATHS_PER_MINUTE (*pvp.common.values.ValueName attribute*), 153

C

callbacks (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
check() (*pvp.alarm.condition.AlarmSeverityCondition method*), 186
check() (*pvp.alarm.condition.Condition method*), 182
check() (*pvp.alarm.condition.CycleAlarmSeverityCondition method*), 187
check() (*pvp.alarm.condition.CycleValueCondition method*), 184
check() (*pvp.alarm.condition.TimeValueCondition method*), 185
check() (*pvp.alarm.condition.ValueCondition method*), 183
check() (*pvp.alarm.rule.Alarm_Rule method*), 180
check_files() (*pvp.common.loggers.DataLogger method*), 163
check_rule() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 174
clear_alarm() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
clear_all_alarms() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
cleared_alarms (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
close_button (*pvp.gui.widgets.alarm_bar.Alarm_Card attribute*), 125
close_logfile() (*pvp.common.loggers.DataLogger method*), 163
closeEvent() (*pvp.gui.main.PVP_Gui method*), 115
cmH2O_to_hPa() (*in module pvp.common.unit_conversion*), 168
Condition (*class in pvp.alarm.condition*), 181
CONTROL (*in module pvp.common.values*), 157
control (*pvp.common.values.Value property*), 156
CONTROL (*pvp.gui.main.PVP_Gui attribute*), 113
control (*pvp.gui.widgets.display.Display.self attribute*), 129
Control_Panel (*class in pvp.gui.widgets.control_panel*), 118
control_type (*pvp.common.values.Value property*), 156
control_width (*pvp.gui.main.PVP_Gui attribute*), 113
ControlModuleBase (*class in pvp.controller.control_module*), 144
ControlModuleDevice (*class in pvp.controller.control_module*), 149
ControlModuleSimulator (*class in pvp.controller.control_module*), 151
controls (*pvp.gui.main.PVP_Gui attribute*), 110
controls_set (*pvp.gui.main.PVP_Gui property*), 117
ControlSetting (*class in pvp.common.message*), 159
ControlValues (*class in pvp.common.message*), 160
coordinator (*pvp.gui.main.PVP_Gui attribute*), 111
CoordinatorBase (*class in pvp.coordinator.coordinator*), 190
CoordinatorLocal (*class in pvp.coordinator.coordinator*), 191
CoordinatorRemote (*class in pvp.coordinator.coordinator*), 192
create_signals() (*pvp.gui.widgets.components.EditableLabel method*), 139
cycle_autoset_changed (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 119
CycleAlarmSeverityCondition (*class in pvp.alarm.condition*), 186
cycles (*pvp.gui.widgets.plot.Plot attribute*), 134
CycleValueCondition (*class in pvp.alarm.condition*), 183

D

DataLogger (*class in pvp.common.loggers*), 162
deactivate() (*pvp.alarm.alarm.Alarm method*), 178
deactivate_alarm() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 175
decimals (*pvp.common.values.Value property*), 156
decimals (*pvp.gui.widgets.display.Display.self attribute*), 128
default (*pvp.common.values.Value property*), 156
dependencies (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
depends (*pvp.alarm.rule.Alarm_Rule property*), 180
depends_callbacks (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 173, 174
DerivedValues (*class in pvp.common.message*), 161
dismiss_alarm() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 175
Display (*class in pvp.gui.widgets.display*), 128
display (*pvp.common.values.Value property*), 156
DISPLAY_CONTROL (*in module pvp.common.values*), 158
DISPLAY_MONITOR (*in module pvp.common.values*), 157
DoubleSlider (*class in pvp.gui.widgets.components*), 137
doubleValueChanged (*pvp.gui.widgets.components.DoubleSlider attribute*), 137

E

EditableLabel (*class in pvp.gui.widgets.components*), 138

emit_alarm() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 175
emitDoubleValueChanged() (*pvp.gui.widgets.components.DoubleSlider method*), 137
end_time (*pvp.alarm.alarm.Alarm attribute*), 177, 178
enum_name (*pvp.gui.widgets.display.Display.self attribute*), 129
escapePressed (*pvp.gui.widgets.components.KeyPressHandler attribute*), 138
escapePressedAction() (*pvp.gui.widgets.components.EditableLabel method*), 139
eventFilter() (*pvp.gui.widgets.components.KeyPressHandler method*), 138

F

files (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
FIO2 (*pvp.common.values.ValueName attribute*), 154
flow_ex (*pvp.io.hal.Hal property*), 171
flow_in (*pvp.io.hal.Hal property*), 171
FLOWOUT (*pvp.common.values.ValueName attribute*), 154
flush_logfile() (*pvp.common.loggers.DataLogger method*), 163

G

get_alarm_manager() (*in module pvp.alarm.condition*), 181
get_alarm_severity() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 175
get_alarms() (*in module pvp.coordinator.rpc*), 193
get_alarms() (*pvp.controller.control_module.ControlModuleBase method*), 147
get_alarms() (*pvp.coordinator.coordinator.CoordinatorBase method*), 190
get_alarms() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
get_alarms() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
get_breath_detection() (*in module pvp.coordinator.rpc*), 193
get_breath_detection() (*pvp.controller.control_module.ControlModuleBase method*), 147
get_breath_detection() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
get_breath_detection() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
get_breath_detection() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
get_breath_detection() (*pvp.gui.main.PVP_Gui method*), 116
get_control() (*in module pvp.coordinator.rpc*), 193
get_control() (*pvp.controller.control_module.ControlModuleBase method*), 147
get_control() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
get_control() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
get_control() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
get_control_module() (*in module pvp.controller.control_module*), 152
get_coordinator() (*in module pvp.coordinator.coordinator*), 193
get_heartbeat() (*pvp.controller.control_module.ControlModuleBase method*), 149
get_past_waveforms() (*pvp.controller.control_module.ControlModuleBase method*), 148
get_pref() (*in module pvp.common.prefs*), 167
get_pressure() (*pvp.controller.control_module.Balloon_Simulator method*), 151
get_rpc_client() (*in module pvp.coordinator.rpc*), 193
get_sensors() (*in module pvp.coordinator.rpc*), 193
get_sensors() (*pvp.controller.control_module.ControlModuleBase method*), 147
get_sensors() (*pvp.coordinator.coordinator.CoordinatorBase method*), 190
get_sensors() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 191
get_sensors() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
get_version() (*in module pvp.common.utils*), 169
group (*pvp.common.values.Value property*), 157
gui_closing (*pvp.gui.main.PVP_Gui attribute*), 113

H

Hal (*class in pvp.io.hal*), 169
handle_alarm() (*pvp.gui.main.PVP_Gui method*), 115
HeartBeat (*class in pvp.gui.widgets.control_panel*), 120
heartbeat (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 118
heartbeat (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 121
heartbeat() (*pvp.coordinator.process_manager.ProcessManager method*), 194
HIGH (*pvp.alarm.AlarmSeverity attribute*), 189
HIGH_O2 (*pvp.alarm.AlarmType attribute*), 189
HIGH_PEEP (*pvp.alarm.AlarmType attribute*), 189
HIGH_PRESSURE (*pvp.alarm.AlarmType attribute*), 188
HIGH_VTE (*pvp.alarm.AlarmType attribute*), 189

history (*pvp.gui.widgets.plot.Plot attribute*), 134
hPa_to_cmH2O() (*in module pvp.common.unit_conversion*), 168
human_name (*pvp.alarm.AlarmType property*), 189

I

icons (*pvp.gui.widgets.alarm_bar.Alarm_Bar attribute*), 123
id (*pvp.alarm.alarm.Alarm attribute*), 177, 178
id_counter (*pvp.alarm.alarm.Alarm attribute*), 177
idx (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
IE_RATIO (*pvp.common.values.ValueName attribute*), 154
increment_delay (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
increment_level() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
init() (*in module pvp.common.prefs*), 167
init_audio() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
init_controls() (*pvp.gui.main.PVP_Gui method*), 117
init_logger() (*in module pvp.common.loggers*), 161
init_ui() (*pvp.gui.main.PVP_Gui method*), 114
init_ui() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
init_ui() (*pvp.gui.widgets.alarm_bar.Alarm_Card method*), 125
init_ui() (*pvp.gui.widgets.control_panel.Control_Panel method*), 119
init_ui() (*pvp.gui.widgets.control_panel.HeartBeat method*), 121
init_ui() (*pvp.gui.widgets.control_panel.StopWatch method*), 123
init_ui() (*pvp.gui.widgets.display.Display method*), 130
init_ui() (*pvp.gui.widgets.display.Limits_Plot method*), 132
init_ui() (*pvp.gui.widgets.plot.Plot_Container method*), 136
init_ui_controls() (*pvp.gui.main.PVP_Gui method*), 114
init_ui_labels() (*pvp.gui.widgets.display.Display method*), 130
init_ui_layout() (*pvp.gui.widgets.display.Display method*), 130
init_ui_limits() (*pvp.gui.widgets.display.Display method*), 130
init_ui_monitor() (*pvp.gui.main.PVP_Gui method*), 114
init_ui_plots() (*pvp.gui.main.PVP_Gui method*), 114
init_ui_record() (*pvp.gui.widgets.display.Display method*), 130
init_ui_signals() (*pvp.gui.main.PVP_Gui method*), 114
init_ui_signals() (*pvp.gui.widgets.display.Display method*), 130
init_ui_slider() (*pvp.gui.widgets.display.Display method*), 130
init_ui_status_bar() (*pvp.gui.main.PVP_Gui method*), 114
init_ui_toggle_button() (*pvp.gui.widgets.display.Display method*), 130
INSPIRATION_TIME_SEC (*pvp.common.values.ValueName attribute*), 153
is_running() (*pvp.controller.control_module.ControlModuleBase method*), 149
is_running() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
is_running() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
is_running() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
is_set (*pvp.gui.widgets.display.Display property*), 132

K

KeyPressHandler (*class in pvp.gui.widgets.components*), 138
kill() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
kill() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
kill() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 193

L

labelPressedEvent() (*pvp.gui.widgets.components.EditableLabel method*), 139
labelUpdatedAction() (*pvp.gui.widgets.components.EditableLabel method*), 139
launch_gui() (*in module pvp.gui.main*), 117
LEAK (*pvp.alarm.AlarmType attribute*), 189
limits_changed (*pvp.gui.widgets.plot.Plot attribute*), 134
Limits_Plot (*class in pvp.gui.widgets.display*), 132
limits_updated() (*pvp.gui.main.PVP_Gui method*), 115
load_file() (*pvp.common.loggers.DataLogger method*), 164
load_pixmaps() (*pvp.gui.widgets.control_panel.Lock_Button method*), 120

load_pixmaps() (*pvp.gui.widgets.control_panel.Start_Button method*), 119
load_prefs() (*in module pvp.common.prefs*), 167
load_rule() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 174
load_rules() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 174
load_state() (*pvp.gui.main.PVP_Gui method*), 116
LOADED (*in module pvp.common.prefs*), 165
Lock_Button (*class in pvp.gui.widgets.control_panel*), 120
lock_button (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 118
locked (*pvp.gui.main.PVP_Gui attribute*), 111
log2csv() (*pvp.common.loggers.DataLogger method*), 164
log2mat() (*pvp.common.loggers.DataLogger method*), 164
logged_alarms (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
logger (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 174
logger (*pvp.gui.main.PVP_Gui attribute*), 111
LOW (*pvp.alarm.AlarmSeverity attribute*), 189
LOW_O2 (*pvp.alarm.AlarmType attribute*), 189
LOW_PEEP (*pvp.alarm.AlarmType attribute*), 189
LOW_PRESSURE (*pvp.alarm.AlarmType attribute*), 188
LOW_VTE (*pvp.alarm.AlarmType attribute*), 188

M

make_dirs() (*in module pvp.common.prefs*), 167
make_icons() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
manager (*pvp.alarm.condition.Condition property*), 182
maximum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
MEDIUM (*pvp.alarm.AlarmSeverity attribute*), 189
minimum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
MISSED_HEARTBEAT (*pvp.alarm.AlarmType attribute*), 189
mode (*pvp.alarm.condition.AlarmSeverityCondition property*), 186
mode (*pvp.alarm.condition.ValueCondition property*), 183
module
    pvp.alarm, 188
    pvp.alarm.alarm, 176
    pvp.alarm.alarm_manager, 172
    pvp.alarm.condition, 181
    pvp.alarm.rule, 180
    pvp.common.fashion, 169
    pvp.common.loggers, 161
    pvp.common.message, 158
    pvp.common.prefs, 165
    pvp.common.unit_conversion, 167
    pvp.common.utils, 168
    pvp.common.values, 152
    pvp.controller.control_module, 144
    pvp.coordinator.coordinator, 190
    pvp.coordinator.process_manager, 194
    pvp.coordinator.rpc, 193
    pvp.gui.main, 110
    pvp.gui.styles, 141
    pvp.gui.widgets.alarm_bar, 123
    pvp.gui.widgets.components, 136
    pvp.gui.widgets.control_panel, 118
    pvp.gui.widgets.dialog, 141
    pvp.gui.widgets.display, 128
    pvp.gui.widgets.plot, 133
    pvp.io, 169
    pvp.io.devices, 171
    pvp.io.hal, 169
MONITOR (*pvp.gui.main.PVP_Gui attribute*), 113
monitor (*pvp.gui.main.PVP_Gui attribute*), 110
MONITOR_UPDATE_INTERVAL (*in module pvp.gui.styles*), 141
monitor_width (*pvp.gui.main.PVP_Gui attribute*), 113

N

n_cycles (*pvp.alarm.condition.CycleAlarmSeverityCondition property*), 187
n_cycles (*pvp.alarm.condition.CycleValueCondition property*), 184
name (*pvp.common.values.Value property*), 156
name (*pvp.gui.widgets.display.Display.self attribute*), 128

O

OBSTRUCTION (*pvp.alarm.AlarmType attribute*), 189
OFF (*pvp.alarm.AlarmSeverity attribute*), 189
OnOffButton (*class in pvp.gui.widgets.components*), 140
operator (*pvp.alarm.condition.CycleValueCondition attribute*), 184
operator (*pvp.alarm.condition.ValueCondition attribute*), 183
orientation (*pvp.gui.widgets.display.Display.self attribute*), 129
OUupdate() (*pvp.controller.control_module.Balloon_Simulator method*), 151
oxygen (*pvp.io.hal.Hal property*), 171

P

PEEP (*pvp.common.values.ValueName attribute*), 153
PEEP_TIME (*pvp.common.values.ValueName attribute*), 153
pending_clears (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174

pigpio_command() (*in module pvp.common.fashion*), 169
PIP (*pvp.common.values.ValueName attribute*), 153
PIP_TIME (*pvp.common.values.ValueName attribute*), 153
pixmaps (*pvp.gui.widgets.control_panel.Lock_Button attribute*), 120
pixmaps (*pvp.gui.widgets.control_panel.Start_Button attribute*), 119
play() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
playing (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 126
Plot (*class in pvp.gui.widgets.plot*), 133
plot (*pvp.common.values.Value property*), 157
plot_box (*pvp.gui.main.PVP_Gui attribute*), 111
Plot_Container (*class in pvp.gui.widgets.plot*), 135
PLOT_FREQ (*in module pvp.gui.widgets.plot*), 133
plot_limits (*pvp.common.values.Value property*), 157
PLOT_TIMER (*in module pvp.gui.widgets.plot*), 133
plot_width (*pvp.gui.main.PVP_Gui attribute*), 113
PLOTS (*in module pvp.common.values*), 158
PLOTS (*pvp.gui.main.PVP_Gui attribute*), 113
plots (*pvp.gui.widgets.plot.Plot_Container attribute*), 135
pop_dialog() (*in module pvp.gui.widgets.dialog*), 141
PRESSURE (*pvp.common.values.ValueName attribute*), 154
pressure (*pvp.io.hal.Hal property*), 171
pressure_units_changed (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 119
ProcessManager (*class in pvp.coordinator.process_manager*), 194
pvp.alarm
    module, 188
pvp.alarm.alarm
    module, 176
pvp.alarm.alarm_manager
    module, 172
pvp.alarm.condition
    module, 181
pvp.alarm.rule
    module, 180
pvp.common.fashion
    module, 169
pvp.common.loggers
    module, 161
pvp.common.message
    module, 158
pvp.common.prefs
    module, 165
pvp.common.unit_conversion
    module, 167
pvp.common.utils
    module, 168
pvp.common.values
    module, 152
pvp.controller.control_module
    module, 144
pvp.coordinator.coordinator
    module, 190
pvp.coordinator.process_manager
    module, 194
pvp.coordinator.rpc
    module, 193
pvp.gui.main
    module, 110
pvp.gui.styles
    module, 141
pvp.gui.widgets.alarm_bar
    module, 123
pvp.gui.widgets.components
    module, 136
pvp.gui.widgets.control_panel
    module, 118
pvp.gui.widgets.dialog
    module, 141
pvp.gui.widgets.display
    module, 128
pvp.gui.widgets.plot
    module, 133
pvp.io
    module, 169
pvp.io.devices
    module, 171
pvp.io.hal
    module, 169
PVP_Gui (*class in pvp.gui.main*), 110

Q

QHLine (*class in pvp.gui.widgets.components*), 139
QVLine (*class in pvp.gui.widgets.components*), 139

R

redraw() (*pvp.gui.widgets.display.Display method*), 131
register_alarm() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 175
register_dependency() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
reset() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
reset() (*pvp.alarm.condition.AlarmSeverityCondition method*), 186
reset() (*pvp.alarm.condition.Condition method*), 182
reset() (*pvp.alarm.condition.CycleAlarmSeverityCondition method*), 187

reset() (*pvp.alarm.condition.CycleValueCondition method*), 184
reset() (*pvp.alarm.condition.TimeValueCondition method*), 185
reset() (*pvp.alarm.condition.ValueCondition method*), 183
reset() (*pvp.alarm.rule.Alarm_Rule method*), 180
reset_start_time() (*pvp.gui.widgets.plot.Plot method*), 135
reset_start_time() (*pvp.gui.widgets.plot.Plot_Container method*), 136
restart_process() (*pvp.coordinator.process_manager.Process_Manager method*), 194
returnPressed (*pvp.gui.widgets.components.KeyPressHandler attribute*), 138
returnPressedAction() (*pvp.gui.widgets.components.EditableLabel method*), 139
rotation_newfile() (*pvp.common.loggers.DataLogger method*), 164
rounded_string() (*in module pvp.common.unit_conversion*), 168
rpc_server_main() (*in module pvp.coordinator.rpc*), 193
rules (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 173, 174
running (*pvp.gui.main.PVP_Gui attribute*), 111
runtime (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 118

S

safe_range (*pvp.common.values.Value property*), 156
safe_range (*pvp.gui.widgets.display.Display.self attribute*), 128
save_prefs() (*in module pvp.common.prefs*), 167
save_state() (*pvp.gui.main.PVP_Gui method*), 116
SENSOR (*in module pvp.common.values*), 157
sensor (*pvp.common.values.Value property*), 156
sensor_value (*pvp.gui.widgets.display.Display.self attribute*), 129
sensor_value (*pvp.gui.widgets.display.Limits_Plot attribute*), 132
SENSORS_STUCK (*pvp.alarm.AlarmType attribute*), 189
SensorValues (*class in pvp.common.message*), 158
set_breath_detection() (*in module pvp.coordinator.rpc*), 193
set_breath_detection() (*pvp.controller.control_module.ControlModuleBase method*), 147
set_breath_detection() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
set_breath_detection() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
set_breath_detection() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
set_breath_detection() (*pvp.gui.main.PVP_Gui method*), 116
set_control() (*in module pvp.coordinator.rpc*), 193
set_control() (*pvp.controller.control_module.ControlModuleBase method*), 147
set_control() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
set_control() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
set_control() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
set_control() (*pvp.gui.main.PVP_Gui method*), 115
set_dark_palette() (*in module pvp.gui.styles*), 141
set_duration() (*pvp.gui.widgets.plot.Plot method*), 134
set_duration() (*pvp.gui.widgets.plot.Plot_Container method*), 136
set_flow_in() (*pvp.controller.control_module.Balloon_Simulator method*), 151
set_flow_out() (*pvp.controller.control_module.Balloon_Simulator method*), 151
set_icon() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
set_indicator() (*pvp.gui.widgets.control_panel.HeartBeat method*), 122
set_locked() (*pvp.gui.widgets.display.Display method*), 131
set_mute() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
set_plot_mode() (*pvp.gui.widgets.plot.Plot_Container method*), 136
set_pref() (*in module pvp.common.prefs*), 167
set_pressure_units() (*pvp.gui.main.PVP_Gui method*), 116
set_safe_limits() (*pvp.gui.widgets.plot.Plot method*), 134
set_safe_limits() (*pvp.gui.widgets.plot.Plot_Container method*), 136
set_sound() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
set_state() (*pvp.gui.widgets.components.OnOffButton method*), 140
set_state() (*pvp.gui.widgets.control_panel.HeartBeat method*), 121
set_state() (*pvp.gui.widgets.control_panel.Lock_Button method*), 120
set_state() (*pvp.gui.widgets.control_panel.Start_Button method*), 119
set_units() (*pvp.gui.widgets.display.Display method*), 131

set_units() (*pvp.gui.widgets.plot.Plot method*), 135
set_units() (*pvp.gui.widgets.plot.Plot_Container method*), 136
set_value (*pvp.gui.widgets.display.Display.self attribute*), 129
set_value (*pvp.gui.widgets.display.Limits_Plot attribute*), 132
set_value() (*pvp.gui.main.PVP_Gui method*), 114
set_valves_standby() (*pvp.controller.control_module.ControlModuleDemo method*), 150
setColor() (*pvp.gui.widgets.components.QHLine method*), 139
setColor() (*pvp.gui.widgets.components.QVLine method*), 140
setDecimals() (*pvp.gui.widgets.components.DoubleSlider method*), 137
setEditable() (*pvp.gui.widgets.components.EditableLabel method*), 139
setLabelEditableAction() (*pvp.gui.widgets.components.EditableLabel method*), 139
setMaximum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
setMinimum() (*pvp.gui.widgets.components.DoubleSlider method*), 138
setpoint_ex (*pvp.io.hal.Hal property*), 171
setpoint_in (*pvp.io.hal.Hal property*), 171
setSingleStep() (*pvp.gui.widgets.components.DoubleSlider method*), 138
setText() (*pvp.gui.widgets.components.EditableLabel method*), 139
setValue() (*pvp.gui.widgets.components.DoubleSlider method*), 138
severity (*pvp.alarm.alarm.Alarm property*), 178
severity (*pvp.alarm.rule.Alarm_Rule property*), 180
severity (*pvp.gui.widgets.alarm_bar.Alarm_Card attribute*), 125
severity_map (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 127
singleStep() (*pvp.gui.widgets.components.DoubleSlider method*), 138
slider (*pvp.gui.widgets.plot.Plot_Container attribute*), 135
snoozed_alarms (*pvp.alarm.alarm_manager.Alarm_Manager attribute*), 172, 174
sound_player (*pvp.gui.widgets.alarm_bar.Alarm_Bar attribute*), 123
start() (*pvp.controller.control_module.ControlModuleBase method*), 149
start() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
start() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
start() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 192
start() (*pvp.gui.main.PVP_Gui method*), 115
Start_Button (*class in pvp.gui.widgets.control_panel*), 119
start_button (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 118
start_process() (*pvp.coordinator.process_manager.ProcessManager method*), 194
start_time (*pvp.gui.main.PVP_Gui attribute*), 111
start_time (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 121
start_timer() (*pvp.gui.widgets.control_panel.HeartBeat method*), 122
start_timer() (*pvp.gui.widgets.control_panel.StopWatch method*), 123
state_changed (*pvp.gui.main.PVP_Gui attribute*), 113
states (*pvp.gui.widgets.control_panel.Lock_Button attribute*), 120
states (*pvp.gui.widgets.control_panel.Start_Button attribute*), 119
staticMetaObject (*pvp.gui.main.PVP_Gui attribute*), 116
staticMetaObject (*pvp.gui.widgets.alarm_bar.Alarm_Bar attribute*), 125
staticMetaObject (*pvp.gui.widgets.alarm_bar.Alarm_Card attribute*), 126
staticMetaObject (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player attribute*), 127
staticMetaObject (*pvp.gui.widgets.components.DoubleSlider attribute*), 138
staticMetaObject (*pvp.gui.widgets.components.EditableLabel attribute*), 139
staticMetaObject (*pvp.gui.widgets.components.KeyPressHandler attribute*), 138
staticMetaObject (*pvp.gui.widgets.components.OnOffButton attribute*), 140
staticMetaObject (*pvp.gui.widgets.components.QHLine attribute*), 139
staticMetaObject (*pvp.gui.widgets.components.QVLine attribute*), 140
staticMetaObject (*pvp.gui.widgets.control_panel.Control_Panel attribute*), 119
staticMetaObject (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 122
staticMetaObject (*pvp.gui.widgets.control_panel.Lock_Button attribute*), 120
staticMetaObject (*pvp.gui.widgets.control_panel.Start_Button attribute*), 120
staticMetaObject (*pvp.gui.widgets.control_panel.StopWatch attribute*), 123
staticMetaObject (*pvp.gui.widgets.display.Display attribute*), 132
staticMetaObject (*pvp.gui.widgets.display.Limits_Plot*

attribute), 132
staticMetaObject (*pvp.gui.widgets.plot.Plot attribute*), 135
staticMetaObject (*pvp.gui.widgets.plot.Plot_Container attribute*), 136
stop() (*pvp.controller.control_module.ControlModuleBase method*), 149
stop() (*pvp.coordinator.coordinator.CoordinatorBase method*), 191
stop() (*pvp.coordinator.coordinator.CoordinatorLocal method*), 192
stop() (*pvp.coordinator.coordinator.CoordinatorRemote method*), 193
stop() (*pvp.gui.widgets.alarm_bar.Alarm_Sound_Player method*), 127
stop_timer() (*pvp.gui.widgets.control_panel.HeartBeat method*), 122
stop_timer() (*pvp.gui.widgets.control_panel.StopWatch method*), 123
StopWatch (*class in pvp.gui.widgets.control_panel*), 122
store_control_command() (*pvp.common.loggers.DataLogger method*), 163
store_derived_data() (*pvp.common.loggers.DataLogger method*), 163
store_program_data() (*pvp.common.loggers.DataLogger method*), 163
store_waveform_data() (*pvp.common.loggers.DataLogger method*), 163

T

TECHNICAL (*pvp.alarm.AlarmSeverity attribute*), 189
text() (*pvp.gui.widgets.components.EditableLabel method*), 139
textChanged (*pvp.gui.widgets.components.EditableLabel attribute*), 139
time_limit() (*in module pvp.common.utils*), 168
timed_update() (*pvp.gui.widgets.display.Display method*), 131
timeout (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 121
timeout() (*in module pvp.common.utils*), 168
TimeoutException, 168
timer (*pvp.gui.main.PVP_Gui attribute*), 111
timer (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 121
timestamps (*pvp.gui.widgets.plot.Plot attribute*), 134
TimeValueCondition (*class in pvp.alarm.condition*), 184
to_dict() (*pvp.common.message.SensorValues method*), 159
to_dict() (*pvp.common.values.Value method*), 157
toggle_control() (*pvp.gui.widgets.display.Display method*), 130
toggle_cycle_widget() (*pvp.gui.main.PVP_Gui method*), 116
toggle_lock() (*pvp.gui.main.PVP_Gui method*), 115
toggle_plot() (*pvp.gui.widgets.plot.Plot_Container method*), 136
toggle_record() (*pvp.gui.widgets.display.Display method*), 131
toggle_start() (*pvp.gui.main.PVP_Gui method*), 115
total_width (*pvp.gui.main.PVP_Gui attribute*), 113
try_stop_process() (*pvp.coordinator.process_manager.ProcessManage method*), 194

U

units (*pvp.gui.widgets.display.Display.self attribute*), 128
update() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 174
update() (*pvp.controller.control_module.Balloon_Simulator method*), 151
update_dependencies() (*pvp.alarm.alarm_manager.Alarm_Manager method*), 176
update_gui() (*pvp.gui.main.PVP_Gui method*), 113
update_icon() (*pvp.gui.widgets.alarm_bar.Alarm_Bar method*), 124
update_interval (*pvp.gui.widgets.control_panel.HeartBeat attribute*), 121
update_limits() (*pvp.gui.widgets.display.Display method*), 131
update_logger_sizes() (*in module pvp.common.loggers*), 162
update_period (*pvp.gui.main.PVP_Gui attribute*), 111
update_period (*pvp.gui.main.PVP_Gui property*), 117
update_period (*pvp.gui.widgets.display.Display.self attribute*), 128
update_sensor_value() (*pvp.gui.widgets.display.Display method*), 131
update_set_value() (*pvp.gui.widgets.display.Display method*), 131
update_state() (*pvp.gui.main.PVP_Gui method*), 116
update_value() (*pvp.gui.widgets.display.Limits_Plot method*), 133
update_value() (*pvp.gui.widgets.plot.Plot method*), 134
update_value() (*pvp.gui.widgets.plot.Plot_Container method*), 136
update_yrange() (*pvp.gui.widgets.display.Limits_Plot method*), 133

V

Value (*class in pvp.common.values*), 154
value() (*pvp.gui.widgets.components.DoubleSlider method*), 138
value_changed (*pvp.gui.widgets.display.Display attribute*), 130
value_names (*pvp.alarm.rule.Alarm_Rule property*), 180
ValueCondition (*class in pvp.alarm.condition*), 182
ValueName (*class in pvp.common.values*), 153
VALUES (*in module pvp.common.values*), 157
VTE (*pvp.common.values.ValueName attribute*), 154